JUN 01, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.2sygefww](https://doi.org/10.17504/protocols.io.2sygefww)

External link:
<https://doi.org/10.1371/journal.pone.0222449>

Protocol Citation: Tuyen Hoang, Harald Wiker, Tomas Mikal L Eagan, Christine Drengenes 2019. 16S Amplicon PCR for the V3-V4 region for the MicroCOPD samples. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.2sygefww>

MANUSCRIPT CITATION: Hoang T, Wiker H, Eagan TM, Drengenes C. 16S Amplicon PCR for the V3-V4 region for the MicroCOPD samples. Published protocol 2019; version 1, 1-16.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

16S Amplicon PCR for the V3-V4 region for the MicroCOPD samples

PLOS One

Tuyen Hoang¹, Harald Wiker¹, Tomas Mikal L Eagan², Christine Drengenes³

¹Department of Clinical Science, University of Bergen, Norway;

²Department of Thoracic Medicine, Haukeland University Hospital;

³Department of Thoracic Medicine, Haukeland University Hospital, Bergen, Norway

Tomas Mikal L Eagan

Department of Thoracic Medicine, Haukeland University Hospi...

ABSTRACT

This is the protocol for DNA extraction, PCR, and MiSeq sample preparation for the microbiome samples from sputum, and the bronchoscopies in the Bergen COPD Cohort study, Bergen COPD Exacerbation Study, the Bergen COPD Microbiome Study and the Bergen ILD Microbiome Study.

Slightly modified Illumina protocol for our low-biomass airways samples in the Bergen COPD & ILD studies.

The protocol that was the template for the PCR and sample preparation is:

16S Metagenomic Sequencing Library Preparation

Preparing 16S Ribosomal RNA Gene Amplicons for the Illumina MiSeq System Part # 15044223 Rev.B.

ATTACHMENTS

[16s-metagenomic-library-prep-guide-15044223-b.pdf](#)

Created: May 11, 2019

MATERIALS

Last Modified: Jun 01, 2019

MATERIALS

PROTOCOL integer ID:
23096

⊗ HotStart ReadyMix (KAPA HiFi PCR kit) **Kapa Biosystems Catalog #KK2601**

⊗ Tris HCl Buffer 1M Solution, Sterile pH 8.5 **Bio Basic Inc. Catalog #SD8141.SIZE.450ml**

⊗ PCR grade water **Contributed by users**

⊗ MicroAmp™ Clear Adhesive Film **Applied Biosystems**

⊗ 16S Amplicon forward primer 10uM **Illumina, Inc.**

⊗ 16S Amplicon reverse primer 10uM **Illumina, Inc.**

DNA extraction

1 Reagents:

Reagent	Supplier	Catalog Number
Sputa sol stock	Oxoid	SR0233
Molecular grade H2O	Merck	H2O MB1001
Lysozyme (10 mg/mL)	Sigma-Aldrich	L3790-10 x 1 mL
Mutanolysin (25,000 U/mL)	Sigma-Aldrich	M9901-50KU
Lysostaphin (4000 U/mL)	Sigma-Aldrich	L4402-5MG
FastDNA SPIN kit	MP Biochemicals	6540-600

2 Take the samples out from the -80 °C freezer and leave to thaw at room temperature.

Sample	Abbreviation	Volume for extraction (μl)
Negative control PBS	CTRL	450 μl
Oral wash	OW	1800 μl
Protected bronchoalveolar lavage	PBAL (PBAL 1 & PBAL 2)	1800 μl
Protected sterile brush	PSB (rPSB & IPSB)	450 μl
Small volume lavage	SVL	1800 μl

- 3 Vortex the samples for 10 seconds and spin down briefly.
- 4 Transfer 450 μL of the CTRL sample to a sterile 1.5 mL eppendorf tube.
- 5 Split the 1800 μL OW sample into three aliquots of 600 μL each and transfer to sterile 1.5 mL eppendorf tubes.
- 6 Split the 1800 μL PBAL sample into three aliquots of 600 μL each and transfer to sterile 1.5 mL eppendorf tubes.

- 7** Prepare a working solution of Sputasol according to the number of samples being processed:

# of samples:	7-8	9-12	13-16	17-20	21-24
Sputasol stock (ml)	0.750	1.125	1.500	1.875	2.250
Molecular grade H ₂ O (ml)	9.250	13.875	18.500	23.125	27.750
Total (ml)	10	15	20	25	30

- 8** Add 450 µL of the Sputasol working solution to the CTRL sample prepared in step 4. Mix by vortexing for 5 seconds.
- 9** Add 600 µL of the Sputasol working solution to each of the three aliquots of OW sample prepared in step 5. Mix by vortexing for 5 seconds.
- 10** Add 600 µL of the Sputasol working solution to each of the three aliquots of PBAL sample prepared in step 6. Mix by vortexing for 5 seconds.
- 11** Incubate the samples on a thermoshaker at 37°C with shaking at 1000 rpm for 15 minutes.
- 12** To pellet the bacterial cells, centrifuge the samples at 15700 x g at room temperature for 8 minutes.
- 13** Discard the supernatant from each sample, taking care not to disturb the pellet.

- 14 Resuspend the pellet in the CTRL sample with 250 μ L PBS.
- 15 Carefully pipette the entire volume up and down 10 times to mix.
- 16 Resuspend and combine the three pellets from the OW sample into one eppendorf tube in a total of 250 μ L PBS.
- 17 Carefully pipette the entire volume up and down 10 times to mix.
- 18 Resuspend and combine the three pellets from the PBAL sample into one eppendorf tube in a total of 250 μ L PBS.
- 19 Carefully pipette the entire volume up and down 10 times to mix.
- 20 To ensure proper mixing, vortex each sample for 15 seconds and spin down briefly.
- 21 Prepare an enzyme cocktail solution according to the number of samples being processed:

Component s	Volume/S ample (μ l)
Lysozyme (10 mg/ml)	25
Mutanolysin (25 000 U/ml)	3
Lysostaphin (4 000 U/ml in NaAC)	1.5
TE5 buffer (10 mM Tris HCl, 5mM EDTA, pH 8.0)	20.5
Total	50

- 22 Add 50 μ l of the enzyme-cocktail solution to each sample. Mix by pipetting the entire volume up and down 10 times. Vortex samples for 1 second.
- 23 Incubate the samples on a thermomixer at 37°C with shaking at 350 rpm for 1 hour.
- 24 Centrifuge the samples at 15700 x g at room temperature for 15 minutes to pellet any bacterial cells not lysed.
- 25 For each sample, transfer the supernatant containing bacterial DNA (about 300 μ L) to a separate

2 mL eppendorf tube and store at 4°C. Proceed with pellet.

- 26** Resuspend the pellet in 800 µL CLS-TC buffer (FastDNA SPIN kit) by pipetting up and down 10 times.
- 27** For each sample, transfer the suspension from step 26 to a Lysing Matrix A tube (FastDNA SPIN kit).
- 28** Homogenize the samples in the FastPrep-24 instrument at a speed setting of 6.0 m/s for 40 seconds.
- 29** Centrifuge the samples at 14000 x g at room temperature for 10 minutes.
- 30** Carefully pipette out 650 µL of the supernatant and transfer to the corresponding 2 mL eppendorf tube set aside in step 25. The total volume is now about 950 µL per sample (650 µL + 300 µL).
- 31** Add 950 µL Binding Matrix (FastDNA SPIN kit) to each sample and invert to mix.
- 32** Place the samples on a rotator and incubate for 5 minutes at room temperature at low speed.
- 33** Transfer 700 µl of each suspension to a SPIN filter (FastDNA SPIN kit).

- 34** Centrifuge the SPIN filter at 14000 x g for 1 minute.
- 35** Discard flow-through.
- 36** Repeat steps 33-35 until the whole sample has been spun through the SPIN filter.
- 37** Add 500 µl prepared SEWS-M (FastDNA SPIN kit) and resuspend each pellet by pipetting up and down several times.
- 38** Centrifuge at 14000 x g for 1 minute. Discard flow-through.
- 39** Without adding any liquid, centrifuge the SPIN filters a second time at 14000 x g for 2 minutes.
- 40** Replace the catch tube with a new catch tube.
- 41** Add 100 µL of DES (FastDNA SPIN kit) and gently resuspend the Binding Matrix by pipetting up and down several times.

- 42 Incubate at 55°C for 5 minutes on a heating block.
- 43 Centrifuge at 14000 x g for 1 minute to elute the DNA.
- 44 Check the purity of the DNA sample using Nanodrop (260/280).
- 45 Measure the DNA concentration using the Qubit fluorometer.

Amplicon PCR

- 46 Create an overview excel sheet and a plate- and pipetting-layout for the PCR run.
- 47 Select and thaw DNA samples for PCR.
- 48 Spin DNA samples at 13000 x g for 1 minute at room temperature.
- 49 Pipette 5 µl of each DNA sample into a well on a PCR plate, following the plate- and pipetting-layout.

- 50** Prepare a master mix containing the following components, in 2 Eppendorf tubes in a prePCR-hood/room. Vortex briefly to mix, and spin briefly to bring down the content.

Comp onent	Vol (μ l) per tube	Total vol (μ l)
16S Ampli con F prime r (10u M)	27.5	55
16S Ampli con R prime r (10u M)	27.5	55
2X KAPA HiFi HotSt art Ready Mix	687.5	1375
PCR grade water	357.5	715
Total	1100	2200

- 51** Combine the master mix from the two Eppendorf-tubes in a sterile trough. Using a multichannel pipette, transfer 20 μ l to each well of the PCR plate. Pipette up and down 10 times to mix.

- 52** Seal the PCR plate with MicroAmp film.

- 53** Centrifuge the PCR plate at 1000 x g for 1 minute at 20 °C.

- 54** Run PCR in a thermal cycler using the following program:

95°C	3 min	
------	-------	--

	95°C	30 sec	45 cycles, row 2->4
	55°C	30 sec	
	72°C	30 sec	
	72°C	5 min	
	4°C	indef	

- 55** Continue with Agarose Gel Electrophoresis for checking of 16S PCR amplicons or store PCR plate at -20°C until use.

Agarose electrophoresis

- 56** Thaw and spin down 16S amplicons in Amplicon PCR plate at 1000 x g at 20°C for 1 minute.
- 57** Cast two big 1.5% agarose gels containing GelRed (per gel use: 3g agarose + 200 ml 1xTAE + 10 µl GelRed). Use four 15-lane combs per gel.
- 58** On a new 96-well plate, add 1 µl 5x DNA loading buffer (use a multichannel pipette and pipette from buffer aliquots in 8 PCR tubes).
- 59** Using a multichannel pipette, transfer 5 µl of 16S amplicons from the Amplicon PCR plate to the 96-well plate containing 5x DNA loading buffer. Pipette up and down gently to mix.
- 60** Centrifuge plate briefly (10 seconds at 1000 x g) to bring down the content.

- 61 Load EZ 100bp ladder (3 µl) on the first lane of each lane-row.
- 62 Load samples using a multichannel pipette (remember to note down the sample order in loading).
- 63 Run gel in 1xTAE buffer at 80V for 15 minutes.
- 64 Increase to 100V and run until the bromophenol blue marker has moved ca. 2/3 of the gel (i.e. 1 hour).

Note

If 1% is used, shorten running time!

PCR Clean-Up to purify the 16S V3 & V4 amplicon away from.

- 65 Consumables:

Item	Quantity (per 96 samples)	Storage
10 mM Tris pH 8.5 (EB buffer)	6 ml	-15°C to -25°C

AMPure XP Beads	2.5 ml	2°C - 8°C
Freshly Prepared 80% Ethanol (EtOH)	50 ml	
96-well 0.2 ml PCR plate	2 plates	
Micro seal 'B' film		
Sterile trough		
10 µl pipette tips with filter	3 boxes	
200 µl pipette tips with filter	7 boxes	

Preparation: Bring the AMPure XP beads to room temperature.

- 66 Centrifuge the Amplicon PCR plate at $1,000 \times g$ at 20°C for 1 minute to collect condensation, carefully remove seal.
- 67 Vortex the AMPure XP beads for 30 seconds to make sure that the beads are evenly dispersed. Add an appropriate volume of beads to a trough depending on the number of samples processing (*ca 2.5 ml for 96 samples*).
- 68 Using a multichannel pipette, add 20 µl of AMPure XP beads to each well of the Amplicon PCR

plate, and gently pipette entire volume up and down 10 times. Change tips between columns.

- 69** Incubate at room temperature without shaking for 5 minutes.
- 70** Place the plate on a magnetic stand for 2 minutes or until the supernatant has cleared.
- 71** With the Amplicon PCR plate on the magnetic stand, use a 100 µl-multichannel pipette to remove first 30 µl and then a 10 µl-multichannel pipette to remove 10 µl and discard the supernatant. Change tips between samples.
- 72** With the Amplicon PCR plate on the magnetic stand, wash the beads with freshly prepared 80% ethanol as follows:
- Using a multichannel pipette, add 200 µl of freshly prepared 80% ethanol to each sample well.
 - Incubate the plate on the magnetic stand for 30 seconds.
 - Carefully remove and discard 195 µl the supernatant.
- 73** With the Amplicon PCR plate on the magnetic stand, perform a second ethanol wash as follows:
- Using a multichannel pipette, add 200 µl of freshly prepared 80% ethanol to each sample well.
 - Incubate the plate on the magnetic stand for 30 seconds.
 - Carefully remove and discard 195 µl the supernatant. Use a 10 µl-multichannel pipette to remove the rest of supernatant and excess ethanol.

Note

(Check for excess ethanol in each well by lifting up the plate from magnetic stand and looking from the bottom of each well. Try to remove all supernatant from wells.)

- 74** With the Amplicon PCR plate still on the magnetic stand, allow the beads to air-dry for 10 minutes.

- 75** Remove the Amplicon PCR plate from the magnetic stand. Using a multichannel pipette, add 52.5 µl of 10 mM Tris pH 8.5 to each well of the Amplicon PCR plate.
- 76** Leave stand for 1-2 minutes, and then gently pipette mix up and down 10 times, changing tips after each column. *Make sure that beads are fully resuspended.*
- 77** Incubate at room temperature for 2 minutes.
- 78** Place the plate on the magnetic stand for 2 minutes or until the supernatant has cleared.
- 79** Using a multichannel pipette, carefully transfer 45 µl of the supernatant from the Amplicon PCR plate to a new 96-well PCR plate. Change tips between samples to avoid cross-contamination. *Avoid taking AMPure beads along when pipetting the supernatant.*
- 80** Centrifuge the plate at 1000 x g at 20°C for 1 minute (to bring down the content and remove bubbles).
- 81** Transfer 5 µl of the cleaned-up PCR amplicons to a new 96-well PCR plate for Index PCR in next steps.
- 82** Seal the plate with the rest with Microseal “B” adhesive seal and store at -20°C.

Note

Safe stopping point.

If you do not immediately proceed to Index PCR, seal the plate with 5 µl amplicons with Microseal "B" adhesive seal and store it at 4°C for short-term (1 day) or at -15°C to -25°C until use (up to a week)

Index PCR - attaches dual indices and Illumina sequencing ...

83 Consumables:

Item	Quantity (for 96 samples)	Storage
2x KAPA HiFi HotStart Ready Mix	2 x 1325 µl	-15°C to -25°C
Nextera XT Index 1 Primers (N7XX) from the Nextera XT Index kit	96 x 5 µl	-15°C to -25°C
Nextera XT Index 2 Primers (S5XX) from the Nextera XT Index kit	96 x 5 µl	-15°C to -25°C
PCR grade water	2 x 530 µl	-15°C to -25°C

TruSeq Index Plate Fixture (FC-130-1005)	1	
Micro Amp Adhesive Clear Film	1	
Sterile trough	3	
10 µl pipette tips with filter	2 boxes +	
200 µl pipette tips with filter	1 box +	

84 Centrifuge the PCR plate with 5 µl of 16S PCR amplicon after clean-up, at 1000 x g at 20°C for 1 minute.

85 Place the plate on a TruSeq Index Plate Fixture (as shown in Fig. 4 below).

Figure 4 TruSeq Index Plate Fixture

- 86** Spin tubes containing Index primers briefly (1 second in a minicentrifuge) to bring down the content (use 2 ml-tubes without caps as “holding adapters”).
- 87** Arrange the Index 2 primer tubes (white caps, clear solution) vertically, aligned with rows A through H in the TruSeq Index Plate Fixture (Fig. 4). Using a multichannel pipette, pipette 5 µl of Index 2 primers to the wells on the PCR plate.
- 88** Arrange the Index 1 primer tubes (orange caps, yellow solution) horizontally, aligned with columns 1 through 12. Using a multichannel pipette, pipette 5 µl of Index 1 primers to the wells on the PCR plate.

Note

For pipetting convenience, begin first with the first 6 primer tubes aligned with columns 1-6 only. Then remove these, and continue with the next 6 primers tubes aligned with columns 7-12

- 89** Prepare a master mix in two 2 ml-tubes, containing:

Component	Volume per tube (µl)
2 x KAPA HiFi HotStart Ready Mix	1325 µl
PCR Grade water	530 µl
Total	1855 µl

Vortex to mix and spin down.

- 90** Transfer the master mix to a sterile trough, and pipette 35 µl to each well on the PCR plate using a multichannel pipette. Gently pipette up and down 10 times to mix.

91 Cover the plate with MicroAmp Adhesive Clear film.

92 Centrifuge the plate at 1,000 × g at 20°C for 1 minute.

93 Perform PCR on a thermal cycler using the following program:

95°C	3 min	
95°C	30 sec	8 cycles, row 2 -> 4
55°C	30 sec	
72°C	30 sec	
72°C	5 min	
4°C	indef	

94 Store Index PCR plate at -20°C until the next steps.

PCR Clean-Up for Index PCR Samples

95 Consumables:

Item	Quantity (per 96 samples)	Storage

10 mM Tris pH 8.5 (EB buffer)	3 ml	-15°C to -25°C
AMPure XP beads	6 ml	2°C to 8°C
Freshly Prepared 80% Ethanol (EtOH)	50 ml	
96-well 0.2 ml PCR plate	3 plates	
Micro seal 'B' film		
Sterile trough	4	
10 µl pipette tips with filter	3 boxes	
200 µl pipette tips with filter	7 boxes	

Preparation: Bring the AMPure XP beads to room temperature.

- 96 Centrifuge the Index PCR plate at $280 \times g$ at 20°C for 1 minute to collect condensation.
- 97 Vortex the AMPure XP beads for 30 seconds to make sure that the beads are evenly dispersed. Add an appropriate volume of beads to a trough (ca 6 ml for 96 samples).
- 98 Using a multichannel pipette, add 56 µl of AMPure XP beads to each well of the Index PCR plate

and gently pipette entire volume up and down 10 times. Change tips between columns.

- 99** Incubate at room temperature without shaking for 5 minutes.
- 100** Place the plate on a magnetic stand for 2 minutes or until the supernatant has cleared.
- 101** With the Amplicon PCR plate on the magnetic stand, use a 100 µl-multichannel pipette to remove first 90 µl and then a 10 µl-multichannel pipette to remove 10 µl and discard the supernatant. Change tips between samples.
- 102** With the Amplicon PCR plate on the magnetic stand, wash the beads with freshly prepared 80% ethanol as follows:
- Using a multichannel pipette, add 200 µl of freshly prepared 80% ethanol to each sample well.
 - Incubate the plate on the magnetic stand for 30 seconds.
 - Carefully remove and discard 195 µl the supernatant.
- 103** With the Amplicon PCR plate on the magnetic stand, perform a second ethanol wash as follows:
- Using a multichannel pipette, add 200 µl of freshly prepared 80% ethanol to each sample well.
 - Incubate the plate on the magnetic stand for 30 seconds.
 - Carefully remove and discard 195 µl the supernatant.
 - Use a 10 µl-multichannel pipette to remove the rest of supernatant and excess ethanol.
(Check for excess ethanol in each well by lifting up the plate from magnetic stand and looking from the bottom of each well. Try to remove all supernatant from wells.)
- 104** With the Index PCR plate still on the magnetic stand, allow the beads to air-dry for 10 minutes.
- 105** Remove the Index PCR plate from the magnetic stand. Using a multichannel pipette, add 27.5 µl of 10 mM Tris pH 8.5 to each well of the Index PCR plate.

- 106** Leave stand for 1-2 min, and then gently pipette mix up and down 10 times, changing tips after each column. *Make sure that beads are fully resuspended.*
- 107** Incubate at room temperature for 2 minutes.
- 108** Place the plate on the magnetic stand for 2 minutes or until the supernatant has cleared.
- 109** Using a multichannel pipette, carefully transfer 10 µl of the supernatant from the Index PCR plate to a new 96-well PCR plate. Label this plate with “first plate”. Change tips between samples to avoid cross-contamination.
- 110** Repeat step 109 (to obtain 2 PCR plates with 10 µl each Index PCR per plate). Label the plate with “second plate”. Change tips between samples to avoid cross-contamination.
- Note**

Avoid taking XP beads along when transferring the samples
- 111** Seal (with Microseal B film) and store the plate with minimal XP beads (usually “first plate”) at -20°C for library normalization (dilution) and pooling for use in sequencing with MiSeq.
- 112** Use the second plate for assessment of DNA concentration and analysis with BioAnalyzer (if not proceeding to these steps right after, seal and store the plate at -20°C until use).

Library Quantification, Normalization & Pooling

- 113** Calculate DNA concentration (C) in each Index PCR sample using the corresponding average concentration value (PicoGreenAve) ng/µl obtained from PicoGreen.

Note

We use Qubit!

$C \text{ (ng/}\mu\text{l)} = \text{PicoGreen Ave} \times 100$

- 114** Convert concentrations in ng/ μ l to nM, using the following formula:
 $[(\text{concentration in ng/}\mu\text{l}) / (660 \text{ mol} \times 630)] \times 10^6 = \text{concentration in nM}$
- 115** For normalization, dilute concentrated final library (i.e. index PCR) using 10 mM Tris pH 8.5 (EB buffer) to 4nM.
i. Transfer a desired volume (e.g. 5 μ l) of cleaned index PCR samples from the Index PCR plate with minimal contaminated XP beads after clean-up to a new PCR plate.
ii. Add an appropriate amount of EB buffer to each well to make a dilution to 4nM. Pipette up and down gently 5 times to mix.
iii. Label the plate with diluted index PCR samples with "IndexPCR" PLATE #, and "4nM".
- 116** Aliquot 5 μ l of each diluted index PCR sample and mix aliquots in an Eppendorf tube for pooling libraries.
- 117** Make 5 5 μ l-aliquots of the library pool, and store at -20°C until use in sample preparation for MiSeq sequencing.

Sample Preparation for MiSeq Sequencing - preparation

- 118** Create a txt/excel file containing the following info separated by tab/in column:

<i>Sample_ID</i>	<i>Sample_Name</i>	<i>Sample_Plate</i>	<i>Sample_Well</i>	<i>I7_Index_ID</i>	<i>index</i>	<i>I5_Index_ID</i>	<i>index 2</i>	<i>Sample_Project</i>	<i>Description</i>
------------------	--------------------	---------------------	--------------------	--------------------	--------------	--------------------	----------------	-----------------------	--------------------

* Tip: use a previous sample sheet as a template and copy over info for the following fields:

<i>I7_Index_ID</i>	<i>index</i>	<i>I5_Index_ID</i>	<i>index 2</i>
--------------------	--------------	--------------------	----------------

(index order is meant to be the same for all runs)

* If there is any notice about samples (DNA extraction, PCR product, etc), use the *Description* field

* Leave empty fields: *Sample_Plate*, *Sample_Well*, and *Sample_Project*

119 Thawing the MiSeq Reagent Cartridge

Note

An alternative to the thawing steps is to thaw reagents overnight in 2°C to 8°C storage. Reagents are stable up to one week when stored at this temperature.

119.1 Remove the reagent cartridge from -15°C to -25°C storage.

119.2 Place the reagent cartridge in a water bath containing enough room temperature deionized water to submerge the base of the reagent cartridge. Do not allow the water to exceed the maximum water line printed on the reagent cartridge.

119.3 Allow the reagent cartridge to thaw in the room temperature water bath for approx 60-90 minutes or until it is thawed completely.

119.4 Remove the cartridge from the water bath and gently tap it on the bench to dislodge water from the base of the cartridge. Dry the base of the cartridge.

119.5 Invert the reagent cartridge ten times to mix the thawed reagents, and then visually inspect

that all positions are thawed.

119.6 Visually inspect the reagents in positions 1,2, and 4 to make sure that they are fully mixed and free of precipitates.

119.7 Gently tap the cartridge on the bench to reduce air bubbles in the reagents.

Note

The MiSeq sipper tubes go to the bottom of each reservoir to aspirate the reagents, so it is important that the reservoirs are free of air bubbles.

119.8 Place the reagent cartridge on ice or set aside at 2°C to 8°C (up to 6 hours) until ready to set up the run. For best results, proceed directly to loading the sample and setting up the run.

Library Denaturing and MiSeq Sample Loading

120 Consumables

Item	Quantity	Storage
10 mM Tris pH 8.5 (EB buffer, Qiagen)	6 µl	-15°C to -25°C
HT1 (Hybridization buffer)	1540 µl	-15°C to -25°C (attached to the MiSeq reagent cartridge)

0.2 N NaOH (less than a week old)	10 μ l	-15°C to -25°C
PhiX Control Kit v3 (FC-110-3001)	4 μ l	-15°C to -25°C
MiSeq reagent cartridge	1 cartridge	-15°C to -25°C
1.7 ml microcentrifuge tubes (screw cap recommended)	3 tubes	
2.5 L ice bucket		

- 121** Set a heat block suitable for 1.7 ml microcentrifuge tubes to 96°C.
- 122** Prepare 0.2 N NaOH by diluting 2 μ l 10M NaOH in 98 μ l PCR-grade H₂O (total volume = 100 μ l).
- 123** In an ice bucket, prepare an ice-water bath by combining 3 parts ice and 1 part water.
- 124** Thaw a tube with 5 μ l pooled library on ice.

- 125** Thaw HT1 (Hybridization Buffer) at room temperature. Keep HT1 on ice (to chill) after thawing.
- 126** Combine the following volumes of pooled final DNA library and freshly diluted 0.2 N NaOH in a microcentrifuge tube:
- 4 nM pooled library (5 μ l)
 - 0.2 N NaOH (5 μ l)
- 127** Set aside the remaining dilution of 0.2 N NaOH to prepare a PhiX control within the next 12 hours.
- 128** Vortex briefly to mix the sample solution, and then centrifuge the sample solution at $280 \times g$ at 20°C for 1 minute.
- 129** Incubate for 5 minutes at room temperature to denature the DNA into single strands.
- 130** Add the following volume of pre-chilled HT1 to the tube containing denatured DNA:
- Denatured DNA (10 μ l)
 - Pre-chilled HT1 (990 μ l)
- Adding the HT1 results in a 20 pM denatured library in 1 mM NaOH.*
- 131** Place the denatured DNA on ice until you are ready to proceed to final dilution.
- 132** Dilute the denatured DNA to 10 pM by combining:
- 20 pM denatured library (300 μ l)
 - Pre-chilled HT1 (300 μ l)

- 133** Invert several times to mix and then pulse centrifuge the DNA solution.
- 134** Place the denatured and diluted DNA on ice.
- 135** Combine the following volumes to dilute the PhiX library to 4 nM:
- 10 nM PhiX library (2 µl)
 - 10 mM Tris pH 8.5 (3 µl)
- 136** Combine the following volumes of 4 nM PhiX and 0.2 N NaOH in a microcentrifuge tube:
- 4 nM PhiX library (5 µl)
 - 0.2 N NaOH (5 µl)
- 137** Vortex briefly to mix the 2 nM PhiX library solution.
- 138** Incubate for 5 minutes at room temperature to denature the PhiX library into single strands.
- 139** Add the following volumes of pre-chilled HT1 to the tube containing denatured PhiX library to result in a 20 pM PhiX library:
- Denatured PhiX library (10 µl)
 - Pre-chilled HT1 (990 µl)
- 140** Dilute the denatured 20 pM PhiX library to 10 pM as follows:
- 20 pM denatured PhiX library (300 µl)
 - Pre-chilled HT1 (300 µl)
- 141** Invert several times to mix and then pulse centrifuge the DNA solution.

142 Place the denatured and diluted PhiX on ice.

143 Combine the following volumes of denatured PhiX control library and the denatured amplicon library in a microcentrifuge tube:

- Denatured and diluted PhiX control (90 µl) – i.e 15% of a total vol of 600µl
- Denatured and diluted amplicon library (510 µl)

Note

15% PhiX control spike-in is used

144 Set the combined sample library and PhiX control aside on ice until you are ready to heat denature the mixture immediately before loading it onto the MiSeq v3 reagent cartridge.

145 Using a heat block, incubate the combined library and PhiX control tube at 96°C for 2 minutes.

Note

Perform the heat denaturation step immediately before loading the library into the MiSeq reagent cartridge to ensure efficient template loading on the MiSeq flow cell

146 After the incubation, invert the tube 1–2 times to mix and immediately place in the icewater bath.

147 Keep the tube in the ice-water bath for 5 minutes.