

Apr 11, 2023

Version 6

2023 GenomeTrakr Proficiency Testing exercise (PulseNet Harmonized) V.6

DOI

dx.doi.org/10.17504/protocols.io.e6nvw5m9dvmk/v6



Maria Balkey¹, Ruth Timme¹, Julie Haendiges¹, Tina Pfefer¹

¹U.S. Food and Drug Administration

GenomeTrakr

Tech. support email: genomeTrakr@fda.hhs.gov



Maria Balkey

US Food and Drug Administration

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.e6nvw5m9dvmk/v6>

Protocol Citation: Maria Balkey, Ruth Timme, Julie Haendiges, Tina Pfefer 2023. 2023 GenomeTrakr Proficiency Testing exercise (PulseNet Harmonized). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.e6nvw5m9dvmk/v6> Version created by **Tina Pfefer**



License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: March 15, 2023

Last Modified: April 11, 2023

Protocol Integer ID: 78837

Keywords: listeria on bhia plate, genometrakr proficiency testing exercise, applicable to all genometrakr lab, proficiency testing isolates in april, pt strain, listeria, fda genometrakr program, genometrakr lab, following proficiency testing isolate, salmonella, genometrakr, shigella esp23, cdc through fda, shigella sap23, gt proficiency testing exercise, cdc, sequencing data, completion of the entire proficiency test, entire proficiency test, participating laboratory

Abstract

Version Updates

v5: Added links to GalaxyTrakr_PT_exercise_report_2022.xlsx

v6: added 2023 PT strains; edited Steps 1, 2, 4, and 5 to include references to growth of *Listeria* on BHIA plates; Updated Tables in Steps 7.1 and 8.1 with sample information for 2023 PT strains.

This SOP outlines guidelines on how to process the isolates for the 2023 GenomeTrakr (GT) Proficiency Testing exercise.

This SOP is applicable to all GenomeTrakr labs participating in the 2023 GenomeTrakr Proficiency Testing exercise (PulseNet Harmonized).

The FDA GenomeTrakr program will ship the following proficiency testing isolates in April 2023.

ESP23-5286	<i>Escherichia/Shigella</i>
ESP23-9201	<i>Escherichia/Shigella</i>
SAP23-3307	<i>Salmonella</i>
LMP23-6714	<i>Listeria</i>

Completion of the entire proficiency test entails the following:

- GT participating laboratory generates sequencing data (fastq files) using the PT strains provided by CDC through FDA-CFSAN-WGS Program-GT in accordance to GT and/or PulseNet SOPs.
- Populate sample sheet according to 2023 GT Proficiency Testing exercise (PulseNet Harmonized) SOP
- Submission of sequencing records to the appropriate project on BaseSpace or Isilon according to GT SOPs
- By participating in the 2023 GT Proficiency Testing exercise (PulseNet Harmonized), GT labs provide consent to use the PT exercise data in subsequent analysis and manuscript publications. Participants will be acknowledged for their contribution on any publication that might require processing data from the 2023 GT PT exercise.



Materials

Materials Needed

Sterile sturdy forceps

- 1 ml pipetman
- 1 ml sterile pipet tips
- 1 µl and/or 10 µl sterile inoculating loop

Reagents Needed

- Trypticase Soy + 5% Sheep Blood Agar plates (BAP) or equivalent media
- BHIA plates
- Sterile reagent grade water or Phosphate Buffered Saline (0.01M PBS; pH 7.4)
- BHI broth
- 70% isopropyl alcohol

Troubleshooting

Safety warnings

! Biological Safety Warning: *Escherichia/Shigella*, *Salmonella* and *Listeria* strains are considered Level 2 biological agents by the U.S. Department of Health and Human Services. Use appropriate precautions when handling the vial or culture. Carry out laboratory work in a biological safety cabinet when applicable to ensure aseptic conditions and personal safety.

Before start

There are four sections in this protocol:

Section 1: Culture preparation of lyophilized isolates.

Section 2: Sequencing

Section 3: Data Transfer

Culture Preparation

1 ***Salmonella*, *Escherichia/Shigella*, and *Listeria* Lyophilized cultures:**

Day 1

Document the isolate number(s) and the lyophilized date(s) for your records. Wipe the aluminum cover and outside of the vial with isopropyl alcohol. Using sturdy forceps, aseptically remove the aluminum cover and rubber stopper from the vial containing the lyophilized culture. Wipe the outside of the rubber stopper and neck of the vial with isopropyl alcohol before removing the stopper.

2 Re-suspend the lyophilized cells with 1 ml of **sterile reagent grade water**. Allow to stand for a few minutes and/or mix gently to produce a uniform suspension. Plate **10 µl** of this suspension onto a blood agar plate (BAP) (or BHIA plate for *Listeria*) and incubate at 37°C overnight in aerobic conditions. It is recommended to plate in duplicate to ensure adequate growth.

3 Add the rest of the suspension to 5 mL of BHI broth and incubate the culture overnight at 37°C in aerobic conditions.

4 **Day 2 and 3**

Check the BAPs and BHIA plates; if the culture appears pure, pick an isolated colony, and streak it on a fresh plate; incubate at 37°C overnight in aerobic conditions. Use the growth from this plate to make DNA templates of the PT strains. Transfer culture to fresh medium and incubate at 37°C overnight; this will ensure that the same culture can be retested, if necessary.

5 **Optional**

If plates don't show bacterial growth, prepare a new plate by taking a loop from the BHI overnight culture (prepared at step 3), streak it on a BAP or BHIA plate as appropriate and incubate at 37°C overnight in aerobic conditions. On the next day check BAP or BHIA plate and proceed as step 4.

Sequencing

6 Perform DNA extraction, library preparation and sequencing according to lab's normal workflow described on SOPs posted at GenomeTrakr protocols.io.



<https://www.protocols.io/view/manual-dna-extraction-using-qiagen-dneasy-blood-an-81wgbp31qvpk/v1>

<https://www.protocols.io/view/procedure-for-operation-and-maintenance-of-the-ill-rm7vz8z52vx1/v1>

<https://www.protocols.io/view/illumina-dna-prep-m-tagmentation-library-preparati-x54v9m7ezg3e/v2>

<https://www.protocols.io/view/dna-quantification-using-the-qubit-fluorometer-81wgbp3x3vpk/v1>

PulseNet Labs will process isolates according to PulseNet Guidelines.

Isolates must be processed exactly as any routine isolate would be processed in the laboratory.

7 Sequencing sample sheets must be filled out according to Table 1.

7.1 **Sample_ID:** Include in this field the values from the **Sample_ID** column of Table 1, *do not modify these IDs*. You will also find this identifier in the vial of the lyophilized culture.

	A	B	C	D
	Sample_ID	Sample_Name	Project	Description
	ESP23-5286	ESP23-5286-M0XXXX-20230424	PR0403_2023_Proficiency_Testing_Exercise	<i>Escherichia/Shigella</i>
	ESP23-9201	ESP23-9201-M0XXXX-20230424	PR0403_2023_Proficiency_Testing_Exercise	<i>Escherichia/Shigella</i>
	SAP23-3307	SAP23-3307-M0XXXX-20230424	PR0403_2023_Proficiency_Testing_Exercise	<i>Salmonella</i>
	LMP23-6714	LMP23-6714-M0XXXX-20230424	PR0403_2023_Proficiency_Testing_Exercise	<i>Listeria</i>

Table 1: PT strain identifiers for Sequencing Sample Sheet (PT isolates processed on a sequencing run with historical and/or routine isolates)

Non PulseNet Labs: NextSeq 1000/2000 and MiSeq systems running Windows 10 (MCS v4 and up) have only Sample_ID available in the sequencing sample sheet, populate the Sample_ID column with the values in the **Sample_ID** column of Table 1.



PulseNet Labs: NextSeq 1000/2000 and MiSeq systems running Windows 10 (MCS v4 and up) have only Sample_ID available in the sequencing sample sheet, follow PulseNet guidelines to populate Sample_ID. Include the isolate identifier, lab ID, Instrument ID and run start date. (e.g. ESP23-5286-**GA**-M01001-20230424).

7.2 **Sample_name:**

Non PulseNet Labs: Fill out this field according to example provided in column Sample_Name of table 1. Include the isolate identifier, instrument ID (M0XXXX where "XXXX" corresponds to the instrument identifier) and run start date. (e.g. ESP23-5286_M01001_20230424).

PulseNet Labs: Include the isolate identifier, lab ID, Instrument ID (M0XXXX where "XXXX" corresponds to the instrument identifier) and run start date. (e.g. ESP23-5286-GA-M01001-20230424).

7.3 **Project:** Please fill out the project field with the project identifier PR0403_2023_Proficiency_Testing_Exercise.

7.4 **Description:** For your use only, we do not track this field. Organism names might be included in this field.

8 **Replicates**

Are you running more than one set of PT isolates in a run?

YES: proceed to Step 8.1

NO: proceed to Step 9.

If you don't have enough isolates for your sequencing run, you could run PT isolates multiple times. Because MiSeq does not allow duplicate sample_IDs, please fill out the sequencing sample sheet using the instructions included in **Step 8.1**. See Table 2 (Non-PulseNet) and 3 (PulseNet) for examples. **Only one set of files shall be shared with GenomeTrakr.**

8.1

Modify Isolate identifiers (IDs at vial of lyophilized culture) to create unique identifiers for PT replicated isolates by adding suffixes such as: "_2" or "_3" to isolate identifiers.

Non-PulseNet Laboratories: If you choose to run replicates, the sample sheet must contain unaltered isolate identifiers for each PT strain in the **Sample_name** field.

Identifiers in the sample_ID field must include the isolate identifier, replicate, instrument Id and start run date.

A	B	C	D
Sample_ID	Sample_Name	Project	Description
ESP23-5286-1-M0XXXX-20230424	ESP23-5286	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201-1-M0XXXX-20230424	ESP23-9201	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307-1-M0XXXX-20230424	SAP23-3307	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714-1-M0XXXX-20230424	LMP23-6714	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286-2-M0XXXX-20230424	ESP23-5286	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201-2-M0XXXX-20220424	ESP23-9201	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307-2-M0XXXX-20230424	SAP23-3307	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714-2-M0XXXX-20230424	LMP23-6714	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286-3-M0XXXX-20230424	ESP23-5286	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201-3-M0XXXX-20230424	ESP23-9201	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307-3-M0XXXX-20230424	SAP23-3307	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714-3-M0XXXX-20230424	LMP23-6714	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286-4-M0XXXX-20230424	ESP23-5286	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201-4-M0XXXX-20230424	ESP23-9201	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307-4-M0XXXX-20230424	SAP23-3307	PR0403_2023_Proficiency_Testing_Exercise	Salmonella

A	B	C	D
LMP23-6714-4-M0XXXX-20230424	LMP23-6714	PR0403_2023_Proficiency_Testing_Exercise	Listeria

Table 2: Strain identifiers for sequencing sample sheet (Case for runs that include replicates) (Non-PulseNet)

PulseNet Laboratories: Identifiers in the sample_ID and sample_Name fields must include the isolate identifier, replicate, lab ID, instrument Id and start run date.

A	B	C	D
Sample_ID	Sample_Name	Project	Description
ESP23-5286_1_GA_M0XXXX_20230424	ESP23-5286_1_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201_1_GA_M0XXXX_20230424	ESP23-9201_1_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307_1_GA_M0XXXX_20230424	SAP23-3307_1_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714_1_GA_M0XXXX_20230424	LMP23-6714_1_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286_2_GA_M0XXXX_20230424	ESP23-5286_2_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201_2_GA_M0XXXX_20230424	ESP23-9201_2_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307_2_GA_M0XXXX_20230424	SAP23-3307_2_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714_2_GA_M0XXXX_20230424	LMP23-6714_2_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286_3_GA_M0XXXX_20230424	ESP23-5286_3_GA_M0XXX_X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella

A	B	C	D
ESP23-9201_3_GA_M0XXXX_20230424	ESP23-9201_3_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307_3_GA_M0XXXX_20230424	SAP23-3307_3_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714_3_GA_M0XXXX_20230424	LMP23-6714_3_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Listeria
ESP23-5286_4_GA_M0XXXX_20230424	ESP23-5286_4_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
ESP23-9201_4_GA_M0XXXX_20230424	ESP23-9201_4_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Escherichia/Shigella
SAP23-3307_4_GA_M0XXXX_20230424	SAP23-3307_4_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Salmonella
LMP23-6714_4_GA_M0XXXX_20230424	LMP23-6714_4_GA_M0XXX X_20230424	PR0403_2023_Proficiency_Testing_Exercise	Listeria

Table 3: Strain identifiers for sequencing sample sheet (Case for runs that include replicates) (PulseNet)

8.2 **Project:** Please fill out the project field with the project identifier
PR0403_2023_Proficiency_Testing_Exercise

8.3 **Description:** For your use only, we do not track this field.

9 Quality of sequencing run

-Data analysis results must be reported to GenomeTrakr in the spreadsheet included in the attachment.

- Check the quality of your sequencing records by following the SOP **Assessing sequence quality in GalaxyTrakr**. Note that Shigella isolates will be predicted as *Escherichia coli* with PubMLST scanning of contigs. The result is acceptable. Report



your findings in the spreadsheet  GalaxyTrakr_PT_exercise_report_2... , sheet MicroRunQC.

- Run the tool **ShigaTyper** Shigella genosertotyping. Report your findings in the spreadsheet GalaxyTrakr_PT_exercise_report_2023, sheet ShigaTyper.
- Run the tool **SeqSero2** for *Salmonella* serotyping. Report your findings in the spreadsheet GalaxyTrakr_PT_exercise_report_2023, sheet SeqSero.

Data Transfer

10 Data transfer

After checking the quality of your records, transfer the data and associated QC results to GenomeTrakr.

10.1 Sharing a run in BaseSpace

- Click the Runs tab in the Illumina BaseSpace website .
- Select the run that you would like to share with the GenomeTrakr team.
- Go to the summary tab and click at the share button.
- Enter the email address for the FDA team (**gnometrakr@fda.hhs.gov**), and then click Add Collaborator.
- Click Save Settings. Your run will be automatically shared with the GenomeTrakr team.

10.2 Sharing a project in BaseSpace

- Click the Projects tab in the Illumina BaseSpace website .
- Select the project (**PR0403_2023_Proficiency_Testing_Exercise**) that you would like to share with the GenomeTrakr team.
- Click the share project button.
- Enter the email address for the FDA team (**gnometrakr@fda.hhs.gov**), and then click Add Collaborator.
- Click Save Settings. Your project will be automatically shared with the GenomeTrakr team.

10.3



Labs inside the FDA network must share the the sequencing files by transferring the sequencing run folder to the isilon storage drive.

11 **PT exercise completion notification**

Notify GenomeTrakr of your completion of the PT exercise by sending an email to: **genometrakr@fda.hhs.gov**. The subject line should include "2023 WGS Proficiency Testing_YourLabName", attach sequencing sample sheet and include the following information in the body of the email:

- a. Run name:
- b. Sequenced by:
- c. Results submitted by:
- d. MiSeq ID:
- e. Flow cell ID:

Select one of the options

- f. SOP Protocol:
 - PulseNet SOP
 - GenomeTrakr SOP

Non-PulseNet Laboratories, please include
in the attachments of the notification email.



GalaxyTrakr_PT_exercise_report_2...

PulseNet Laboratories must send the 2023 Proficiency Testing exercise data and results to PulseNet.

12



00:00:00