AUG 20, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.4fjgtkn

External link:
<https://www.ncbi.nlm.nih.gov/pubmed/31343979>

Protocol Citation: Norman Van Rhijn, Panagiotis Papastamoulis, Takanori Furukawa, Magnus Rattray, Elaine Bignell, Mike Bromley 2019. 96-well Plate Growth Curve Setup. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.4fjgtkn>

MANUSCRIPT CITATION: Papastamoulis P, Furukawa T, van Rhijn N, Bromley M, Bignell E, Rattray M (2019). Bayesian detection of piecewise linear trends in replicated time-series with application to growth data modelling. The International Journal Of Biostatistics. DOI: 10.1515/ijb-2018-0052

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

96-well Plate Growth Curve Setup

In 2 collections

Norman Van Rhijn¹, Panagiotis Papastamoulis², Takanori Furukawa¹, Magnus Rattray¹, Elaine Bignell¹, Mike Bromley¹

¹University of Manchester;

²Athens University of Economics and Business, Greece

Norman Van Rhijn
University of Manchester

ABSTRACT

Generally, growth assays for filamentous fungi have been performed on solid media, either as dilution series or spot tests. However, the solid media environment does not accurately mimic the environment encountered during infection (ie the mammalian lung). Previously, we have developed a methodology to perform liquid growth assays in time for *A. fumigatus* and other filamentous fungi including analysis via mathematical modelling.

This protocol is designed for *Aspergillus* species (and other filamentous fungi) to generate growth curves in liquid media in a 96-well plate. This can be done in high-throughput to generate 96 growth curves per run.

SAFETY WARNINGS

Spores need to be handled within a Class II cabinet.

Protocol status: Working
We use this protocol and it's working

Created: Jun 20, 2019

Last Modified: Aug 20, 2019

PROTOCOL integer ID:
24779

Plate preparation

1 Harvest and dilute spores in PBS+0.01% Tween to 4×10^5 spores/mL. 30m

1.1 Aliquot 5 uL of this spore stock per well of a CytoOne 96-well plate (Starlab) non-coated, this will be overlayed with 195 uL of liquid media to a total volume of 200 uL. 20m

Note

This equals to 2000 spores (absolute) per well. 1000 spores or 500 spores are possible to use, but result in more variance within replicates.

2 Fill the space between the wells with either media or H₂O. This will avoid an "edge effect", where the wells around the edge of the plate show increased growth compared to other wells. 5m

3 Cover the 96 well plate with a breathable cover to allow gas exchange. 2m

Machine setup

4 In the Manchester Fungal Infection Group we either use a Powerwave X-2 with KC software or a 5m

BioTek platereader. Set this up to run for 48:00:00 at 37 °C to measure OD₆₀₀ every 10 minutes. Take a blank reading before running the plate and preheat the machine. Run your experiment.

This will result in a "growth curve".

Extract or manipulate the data to look like attached (txt file format).

Analysis

- 5 See our next protocol for analysis.