

Sep 20, 2017

Analysis of choroid thickness on OCT image using MATLAB

DOI

dx.doi.org/10.17504/protocols.io.imqcc5w

Chao-Yuan Yeh¹

¹University of Southern California



Chao-Yuan Yeh

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

[Create free account](#)

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.imqcc5w>

Protocol Citation: Chao-Yuan Yeh 2017. Analysis of choroid thickness on OCT image using MATLAB. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.imqcc5w>

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Created: June 26, 2017

Last Modified: March 28, 2018

Protocol Integer ID: 6544

Keywords: analysis of choroid thickness, choroid thickness, thickness of choroid, oct images of both eye, choroid, using matlab, matlab script, written matlab script, oct image, matlab this protocol


Abstract

This protocol describes how to use a custom written MATLAB script to analyze thickness of choroid on OCT images of both eyes.


Troubleshooting




- 1 Make sure you have MATLAB 2016b (or newer) with image processing toolbox installed on your computer.
- 2 Download the MATLAB scripts from https://github.com/joe-of-all-trades/OCT_Analysis
- 3 On the Optovue OCT machine, export cross line OCT images. To allow interpolation, four cross line images are required for each eye. Single cross line image can still be processed by ***OCT_ImageAnalysis_for_CrossLine.m***.
- 4 Run ***OCT_file_processing.m*** to process OCT images. This script works on raw image generated by Optovue OCT machines. It will make the file name easier to handle and extract image content from cross line scans and line scans and save them into tiff format.
- 5 Run ***OCT_ImageAnalysis_for_CrossLine.m*** through all the tif files.
- 6 Follow the prompt and draw along the contour of the retina on the left side of the image.




Then follow the prompt and draw along the contour of the choroid on the left side of the image.



After you're done, choroid thickness will be calculated for you. Red lines represent auxiliary lines perpendicular to the curve. Green lines represent calculated thickness of the curve.



Repeat this process for the right side of the image. Another figure will appear and summarize the thickness measurement.



- 7 After running the script for all four cross lines for both eyes. Run ***OCT_Data_Comparison.m*** to generate interpolated data and compare results between two eyes. Choroid thickness data in eight corners will be plotted and an interpolated picture will be shown. An xls file summarizing statistics will also be generated.

