

Jul 25, 2024

Version 1

Automated Bar-Seq Library Preparation and Pooling V.1

DOI

<https://dx.doi.org/10.17504/protocols.io.3byl49qdjgo5/v1>

David Ross¹, Nina Alperovich¹

¹NIST



Open Datasets Initiative

Align to Innovate

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.3byl49qdjgo5/v1>

Protocol Citation: David Ross, Nina Alperovich 2024. Automated Bar-Seq Library Preparation and Pooling. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.3byl49qdjgo5/v1>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: May 24, 2024



Last Modified: July 25, 2024

Protocol Integer ID: 100565

Keywords: dna sample, sequencing, pooled sample, pooling protocol for automated bar, seq library preparation, automation protocol for plasmid dna extraction, dna extraction, assay protocol, dna concentration, based assay protocol, pool of all timepoint sample, plasmid dna extraction, first round of pcr attaches primer, multiplexed tube for subsequent illumina, pooling protocol, rounds of pcr, sample, fragment analyzer, pcr attaches primer, dna, samples from the same timepoint, procedure, seq, cleanup process, pcr, cleanup step, small quantity of each timepoint sample, timepoint sample

Disclaimer

The protocol outlined in this document was created as a part of the *Pooled, Growth-Based Assays for Protein Function Measurements* pipeline for Align to Innovate's Open Dataset Initiative. Align to Innovate is a non-profit research organization operating under open science principles with the goal of improving science research with programmable experiments. The Open Datasets Initiative is working to accelerate community-driven science with the use of automated labs to pioneer robust data collection methods and curated, high-fidelity, public biological datasets amenable to machine learning. This work was supported by Align to Innovate's Open Datasets Initiative which receives philanthropic funding in part from Griffin Catalyst.

Abstract

Protocol for automated Bar-Seq Library preparation

This protocol prepares 96 DNA samples, representing 24 samples from 4 different timepoints, for multiplexed Illumina sequencing. The process starts with two rounds of PCR, each followed by a bead-based cleanup. The first round of PCR attaches primers that serve as tags to identify the timepoint and sample (e.g., time-points 1-4 in the Pooled, Growth-Based Assay Protocol). The second round of PCR attaches flow-cell adapters required for Illumina sequencing. Following the PCR and cleanup steps, the protocol outlines a procedure for pooling all samples together to ensure balanced representation during sequencing. This process starts with quantifying each sample's DNA concentration and then diluting and pooling together samples from the same timepoint. The pooled samples' concentrations are then measured, and a small quantity of each timepoint sample is run on a gel to ensure the cleanup process was successful. The final step is to dilute and pool of all timepoint samples into one multiplexed tube for subsequent Illumina sequencing.

NOTES:

- Before implementing this protocol, prepare the magnetic bead suspension following the **Preparation of Sera-mag SpeedBeads protocol**.
- This protocol should be implemented after the [Automation Protocol for Plasmid DNA Extraction from E. coli](#) protocol
- A fragment analyzer can also be used instead of a gel to determine if the PCR clean-up process was a success

Materials

Starting samples:

- 96 DNA samples resulting from the Automation Protocol for Plasmid DNA Extraction from *E. coli* protocol

Reagents:

- Nuclease-free water (ThermoFisher Scientific 4387936)
- 80% Absolute Ethanol (Fisher Bioreagents BP2818500)
- Elution Buffer (Qiagen 19086)
- Phusion Flash PCR Mastermix (ThermoFisher Scientific F548L)
- Multiplexing Primers
- Universal Illumina Primers

Labware:

- Three 96-well DeepWell reagent plate (Abgene AB-0765) - one used as a reagent plate and two used as midi plates
- Two 96-well PCR plates (Bio-Rad HSP9635 or HSP9645)
- Three PCR plate lids (Agilent 202497-100)
- 96-well output plate (Eppendorf 30603303)

Primers:

	A	B	C
	Primer name	Description	Sequence
	BarSeq_1_F1	forward barSeq PCR 1 primer, with sample multiplex tag: CGTGTATCTT, row A	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNCGTGTATCTTGG CCTAGACGTGTGATAGACTCAGTCG
	BarSeq_1_F2	forward barSeq PCR 1 primer, with sample multiplex tag: ATTCATTGCA, row B	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNATTCATTGCAGG CCTAGACGTGTGATAGACTCAGTCG
	BarSeq_1_F3	forward barSeq PCR 1 primer, with sample multiplex tag: TCCTTCATAG, row C	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNTCCTTCATAGGG CCTAGACGTGTGATAGACTCAGTCG
	BarSeq_1_F4	forward barSeq PCR 1 primer, with sample multiplex tag: GAATGCACGA, row D	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNGAATGCACGAGG CCTAGACGTGTGATAGACTCAGTCG
	BarSeq_1_F5	forward barSeq PCR 1 primer, with sample multiplex tag: GGAATTGTTC, row E	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNGGAATTGTTCGG CCTAGACGTGTGATAGACTCAGTCG



A	B	C
BarSeq_1_F6	forward barSeq PCR 1 primer, with sample multiplex tag: CCGGACCACA, row F	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNCCGGACCACAG GCCTAGACGTGTGATAGACTCAGTCG
BarSeq_1_F7	forward barSeq PCR 1 primer, with sample multiplex tag: GACTTAGAAG, row G	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNGACTTAGAAGGG CCTAGACGTGTGATAGACTCAGTCG
BarSeq_1_F8	forward barSeq PCR 1 primer, with sample multiplex tag: TCTAGTCTTC, row H	ACACTCTTTCCCTACACGACGCTCTTC CGATCTNNNNNNNNNTCTAGTCTTCGG CCTAGACGTGTGATAGACTCAGTCG
BarSeq_1_R1	reverse barSeq PCR 1 primer, with sample multiplex tag: TAGAGTTGGA, column 1	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNTAGAGTTGGAT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R2	reverse barSeq PCR 1 primer, with sample multiplex tag: AGAGCACTAG, column 2	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNAGAGCACTAGT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R3	reverse barSeq PCR 1 primer, with sample multiplex tag: ACTCTACAGG, column 3	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNACTCTACAGGT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R4	reverse barSeq PCR 1 primer, with sample multiplex tag: CGGTGACACC, column 4	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNCGGTGACACCT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R5	reverse barSeq PCR 1 primer, with sample multiplex tag: GCGTTGGTAT, column 5	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNGCGTTGGTATT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R6	reverse barSeq PCR 1 primer, with sample multiplex tag: TGTGCTAACA, column 6	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNTGTGCTAACAT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R7	reverse barSeq PCR 1 primer, with sample multiplex tag: CCAGAAGTAA, column 7	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNCCAGAAGTAAT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R8	reverse barSeq PCR 1 primer, with sample multiplex tag: CTTATACCTG, column 8	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNCTTATACCTGT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R9	reverse barSeq PCR 1 primer,	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNACTAGAACTTT CGCAGACGTTTTGCAGACTCCTG

A	B	C
	with sample multiplex tag: ACTAGAACTT, column 9	
BarSeq_1_R1 0	reverse barSeq PCR 1 primer, with sample multiplex tag: TTAGGCTTAC, column 10	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNTTAGGCTTACT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R1 1	reverse barSeq PCR 1 primer, with sample multiplex tag: TATCATGAGA, column 11	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNTATCATGAGAT CGCAGACGTTTTGCAGACTCCTG
BarSeq_1_R1 2	reverse barSeq PCR 1 primer, with sample multiplex tag: CTCACACAAG, column 12	CTCGGCATTCTGCTGAACCGCTCTT CCGATCTNNNNNNNNNCTCACACAAGT CGCAGACGTTTTGCAGACTCCTG
BarSeq_2_F	forward barSeq PCR 2 primer	AATGATACGGCGACCACCGAGATCTAC ACTCTTCCCTACACGACGCTCTTCC GATCT
BarSeq_2_R	reverse barSeq PCR 2 primer	CAAGCAGAAGACGGCATACGAGATCG GTCTCGGCATTCTGCTGAACCGCTC TTCCGATCT

Transfer Plasmid DNA to PCR Plate and Dilute with DI water

- 1 Pre-heat the on deck thermocycler (ODTC) for 1st PCR step
- 2 Remove lids from PCR-Plate 1, Sample-Plate, and Reagent Plate
- 3 Pipette 10uL nuclease-free water to each well in PCR plate 1
 - Mix 6x after dispensing
 - dispense 1 mm above bottom of well
 - mix volume 25 uL
 - mixing speed 75 uL/s
 - mix position 0 mm (same as dispense position)
 - liquid following OFF
- 4 Transfer 35uL of extracted plasmid DNA to each well of PCR Plate 1
 - This step is done 24 samples (one time point) at a time, with user instructions to load each plasmid sample input plate (without lids)
 - aspirate 1 mm below liquid surface, with liquid following ON
 - dispense 0.5 mm below liquid surface, with liquid following ON
- 5 User instruction to remove plasmid sample input Sample-Plate

Run First PCR - using primers to identify samples from each timepoint

- 6 Add 28.125 uL Master Mix with reverse primer for the appropriate timepoint and sample to each corresponding well of PCR plate 1
 - dispense 0.5 mm below liquid surface, with liquid following On
 - no mixing
- 7 Add 28.125 uL Master Mix with forward primer for the appropriate timepoint and sample to each corresponding well of PCR plate 1
 - Mix 10x after dispensing
 - dispense 0.5 mm below liquid surface, with liquid following On
 - mix volume 70 uL
 - mixing speed 75 uL/s
 - mix position 1 mm (below dispense position)
- 8 Put a PCR plate lid 1 on the PCR plate 1.



- 9 Move PCR-Plate 1 to ODTc
- 10 Run First PCR Rx using conditions (101.25 uL volume) using the following cycles:
 1. 98 C for 60 s
 2. 3 cycles at:
 - 98 C for 10 s
 - 58 C for 20 s
 - 72 C for 20 s
 3. 72 C for 60 s
 4. 23 C for 10 s

First PCR Cleanup Part 1: Bind template plasmid DNA to the beads and keep the supernatant

- 11 Pipette 54 uL magnetic bead suspension into each well of Midi plate 1
 - Bead ratio: 0.6x
 - Mix the bead suspension 10x before pipetting
 - Get new tips for each well
 - Mix bead suspension again 2x before pipetting each well
 - Pipette bead suspension into Midi plate during the Step 1 PCR
- 12 Move PCR plate 1 from ODTc; take lid off of PCR plate 1
- 13 Close ODTc; run PCR pre-heat method on ODTc (to get ready for Step 2 PCR)
- 14 Shake Midi plate 1 with bead suspension for 10 second at 1800 RMP
- 15 Transfer 90 uL of each sample from PCR plate 1 to Midi plate 1
- 15.1 Mix 10x after dispensing
 - dispense 0.5 mm above bottom of well
 - mix volume 108 uL
 - mixing speed 100 uL/s
 - mix position 2 mm (below liquid surface)
 - liquid following ON
- 16 Incubate for 7 minutes at room temperature



- 17 Move Midi plate 1 to magnet base
- 18 Wait for 4 minutes
- 19 During 4 minute wait, move Midi Plate No. 2 to HHS.
- 20 Transfer 134 uL of supernatant from Midi plate 1 to Midi plate 2

First PCR Cleanup Part 2: bind PCR product to beads and discard the supernatant

- 21 Pipette 80.4 uL bead suspension to Midi plate 2 (on HHS)
 - Bead ratio: 0.6 for LacI
- 21.1 Mix the bead suspension 10x before pipetting
- 21.2 Get new tips for each well and mix bead suspension again 2x before aspiration for each well
- 21.3 Mix 10x after dispensing
 - dispense 0 mm below liquid level
 - 10 mix cycles
 - mix volume 193 uL
 - mixing speed: 100 uL/s
 - mix position 1 mm (below liquid surface)
 - liquid following ON
- 22 Incubate for 7 minutes at room temp, without shaking
- 23 During 7 minute wait, throw Midi Plate No. 1 to waste
- 24 Move Midi plate 2 to magnet base



- 25 Wait for 5 minutes
- 26 Remove and discard most of the supernatant (209.4 uL)
- 27 Pipette 200 uL 80% ethanol into Midi plate 2 (on magnet)
 - Jet dispense, from 20 mm height, using one tip for each row
- 28 Remove ethanol supernatant and discard
- 29 Using 50 uL tips, remove remaining supernatant and dispose
- 30 Move Midi plate 2 to HHS
- 31 Wait 30 s to allow ethanol to finish drying
- 32 Pipette 55 uL DI water into Midi plate 2
- 32.1 Mix 2x after dispensing
 - dispense 1 mm above bottom of well
 - mix volume 41.25 uL
 - mixing speed 100 uL/s
 - mix position 0 mm
 - liquid following OFF
- 33 Shake at 1800 RPM for 10 s
- 34 Shake at 400 RPM for 2 minutes
- 35 Incubate for 3 minutes at room temp without shaking
- 36 Move Midi plate 2 to magnet base



37 Start timer for 5 min incubation

37.1 While waiting, start next step by preparing the 2nd PCR reaction

Run 2nd PCR: using primers for Illumina multiplexing

38 Prepare PCR plate 2 by pipetting 54 μ L of 2x master mix (containing both forward and reverse universal primers for Illumina sequencing) into each well

38.1 Wait for 5 minute timer from the 1st PCR cleanup to finish.

39 Transfer 45 μ L of supernatant from Midi plate 2 to PCR plate 2

39.1 Mix 10x after dispensing

- dispense 0.5 mm below liquid surface
- mix volume 69.3 μ L
- mixing speed 100 μ L/s
- mix position 1 mm (lower than dispense position)
- liquid following ON

40 Put a new PCR plate lid on PCR plate 2.

41 Move PCR-Plate 2 to ODT

42 Run PCR plate 2 (99 μ L) using conditions:

1. 98 C for 60 s
2. 15 Cycles at:
 - 98 C for 10 s
 - 72 C for 30 s
3. 72 C for 120 s
4. 10 C for 20 s



Note: during the 2nd PCR run, throw Midi plate 2 to waste and prepare Midi plate 3 following the instructions below.

Second PCR cleanup: Bind PCR product to beads and discard supernatant

- 43 Move Midi plate 3 to HHS (while 2nd PCR is running)
- 44 Pipette 89 uL bead suspension to Midi plate 3
 - Bead ratio 1.0x
- 44.1 Mix the bead suspension 10x before pipetting
- 44.2 Get new tip for each well & mix beads suspension again 2x before pipetting each well
- 45 Wait for PCR to finish
- 46 Move PCR plate 2 from ODT; take lid off of PCR plate 2
- 47 Shake Midi plate 3 with bead suspension for 10 second at 1800 RMP
- 48 Transfer 89 uL from PCR plate 2 to Midi plate 3
- 48.1 Aspirate 0.5 mm above bottom of well with liquid following OFF
- 48.2 Mix 10x after dispensing
 - dispense 0.5 mm above bottom of well
 - mix volume 133.5 uL
 - mixing speed 100 uL/s
 - mix position 2 mm (below dispense position)
 - liquid following ON
- 49 Incubate for 7 minutes



- 50 Move Midi plate 3 to magnet base
- 51 Wait for 5 minutes
- 52 Remove and discard most of the supernatant (173 uL)
- 53 Pipette 200 uL 80% ethanol into Midi plate 4
 - jet dispense, from 20 mm height, using one tip for each row
- 54 Wait 30 s
- 55 Remove ethanol supernatant and discard
- 56 Using 50 uL tips, remove remaining supernatant and dispose
- 57 Move Midi plate 3 to HHS
- 58 Wait 30 s to allow ethanol to finish drying
- 59 Pipette 50 uL Elution buffer into Midi plate 3
- 59.1 Mix 2x after dispensing
 - dispense 1 mm above bottom of well
 - mix volume 37.5 uL
 - mix position 0 mm
 - liquid following OFF
- 60 Shake at 1800 RPM for 10 s
- 61 Shake at 400 RPM for 2 minutes

- 62 Incubate for 3 minutes without shaking
- 63 Move Midi plate 3 to magnet base
- 64 Wait for 5 minutes
 - Remove lid from Output plate during wait
- 65 Transfer 45 uL supernatant from Midi plate 3 to Output plate
- 66 Replace lid on Output plate

Pooling for Sequencing - Measure each sample's input plasmid DNA Concentration

- 67 Measure the DNA concentrations (C_n) for all input plasmid samples.
 - Note: The expected concentration ($C_{n's}$) is typically between 5 ng/uL and 40 ng/uL, though some samples for time point 4, the final growth plate, may end up lower than that.

Quantify the amount of amplified DNA

- 68 The amount of amplified DNA should also be measured for each sample at each time point using Qubit or an equivalent DNA quantitative measurement.
 - Expected yields are between 10 ng/uL and 30 ng/uL (in 45 uL), though some samples could have as little as 1 ng/uL (e.g., samples grown with the highest antibiotic concentration at the later time points).

Within each time point's samples, calculate the samples' final pooling volumes (V_f)

- 69 **Within each time point's samples (i.e., the 24 samples coming from each growth plate),** identify the the four lowest-concentration samples.
- 70 Calculate the mean of their concentrations in ng/ μ L, and define this concentration as the low concentration limit (C_{low}) for that time point.

- 71 For each sample, use the DNA concentration in ng/ μ L (C_n) and the following equation to calculate the pooling volume (V_n) for each sample:
- $V_n = C_{low}/(C_n + 0.1 \text{ ng/uL}) * 15 \mu\text{L}$
- 72 For each sample, set the final pooling volume (V_f) based on the following conditions:
- If $V_n < 4 \mu\text{L}$, set $V_f = 4 \mu\text{L}$
 - If $V_n > 15 \mu\text{L}$, set $V_f = 15 \mu\text{L}$.

Within each time point, pool all samples and measure the pooled-DNA concentration

- 73 **Within each time point's samples (i.e., the 24 samples coming each growth plate),** pool together all samples into one container according to their final pooling volumes (V_f 's) and store the remaining un-pooled samples at -20C.
- Note: This will give one pooled sample for each time-point (four pooled samples in total), each at a concentration of approximately 20 ng/uL and with a volume of 100 to 300 uL. This is much more DNA than is needed for sequencing.
- 74 For each of the four pooled samples, re-measure the DNA concentration and then store them at -20C until the next step can be accomplished.

Run a gel for each time point

- 75 Run a gel for each time point, and compared these results with those obtained during BarSeq method development.
- The gels should not have any visible bands that are shorter than the expected amplicon size (315 bp)
 - Shorter bands indicate incomplete PCR cleanup that could result in a very low sequence count from Illumina sequencing.
 - If the gels results look correct, proceed to the next step

Dilute each time point's sample and pool all samples for sequencing

- 76 To prepare samples for the sequencing provider, follow their instructions for **pre-made Illumina libraries**.
- Note: They usually give the sample requirements in molar concentration, typically 5 nmol/L, which is about 1 ng/uL for the BarSeq libraries. They usually require a minimum of 20 uL to 30 uL per lane of sequencing.
 - As an example, using the numbers above, a sample protocol might be to dilute 5 μ L of each pooled sample into nuclease-free water to a concentration of 1 ng/uL. This will give much more than the minimum required volume, but it avoids the extra variability from pipetting smaller volumes.



- 77 Combine the diluted time point samples to be run on the sequencer. In the following manner:
- For the first run: start with a single lane of NovaSeq, with all four time points combined in equal proportions. Then, after analyzing that data, if you need more data, order a second lane and send a second sample with only time points 2-4.
 - For consecutive runs: order a single lane of NovaSeq with the combine the time points with the following ratios: $\frac{1}{3}$:1:1:1 for time points 1-4.

Protocol references

This protocol is based on a similar protocol first described by Tack et al., Mol Syst Biol (2021), <https://doi.org/10.15252/msb.202010179>