APR 16, 2021

Automated Tissue Fixation with 10% NBF (Leica)

Donna Skinner¹, Teri Bowman¹, Jon Aster¹

¹Brigham and Women's Hospital

Laboratory of Systems Pharmacology

NCIHTAN

Madison Tyler

Laboratory of Systems Pharmacology

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.67uhhnw

Protocol Citation: Donna Skinner, Teri Bowman, Jon Aster 2021. Automated Tissue Fixation with 10% NBF (Leica). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.67uhhnw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Sep 09, 2019

Last Modified: Apr 16, 2021

PROTOCOL integer ID:
 27604

ABSTRACT

Human tissue specimens are frequently formalin-fixed and paraffin-embedded (FFPE) for histological staining, pathological review, and archival purposes. This is the routine fixation protocol used by the Specialized Histopathology (SHP) Core and the Department of Pathology at Brigham and Women's Hospital. The protocol uses the Leica ASP300 S Fully Enclosed Tissue Processor to automate fixation.

GUIDELINES

When fixatives other than 10% NBF are used, the case will be discussed with the investigator by management to ensure the best possible result.

MATERIALS

Materials & Equipment

- Leica ASP300 S Fully Enclosed Tissue Processor
- 10% Neutral buffered formalin
- Ethyl Alcohol
- Xylene
- Paraffin wax

BEFORE START INSTRUCTIONS

Fixation times can range from 8 to 72 hrs depending on size, thickness, and type of the sample. It is not recommended to fix the sample for longer time periods, as this can have deleterious effects for various assays.

Automated Tissue Fixation with 10% NBF (Leica)

- 1 Program the Leica ASP300 S Fully Enclosed Tissue Processor with either the short program or

long program depending on size, thickness, and type of sample.

2 Formalin Short Program

A	B	C	D
Station/Reagent	Time	Temp	P/V
10% NBF	00:12 min	40	Y
70% Ethyl Alcohol	00:15 min	40	Y
95% Ethyl Alcohol	00:15 min	40	Y
95% Ethyl Alcohol	00:20 min	40	Y
100% Ethyl Alcohol	00:20 min	40	Y
100% Ethyl Alcohol	00:20 min	40	Y
Xylene	00:20 min	40	Y
Xylene	00:30 min	40	Y
Paraffin Wax 1	00:05 min	60	Y
Paraffin Wax 2	00:10 min	60	Y
Paraffin Wax 3	00:15 min	60	Y

Table 1. Formalin short program.

3 Formalin Long Program

A	B	C	D
Station/Reagent	Time	Temp	P/V
10% NBF	02:00 hr	40	Y
70% Ethyl Alcohol	00:45 min	40	Y
95% Ethyl Alcohol	00:45 min	40	Y
95% Ethyl Alcohol	00:45 min	40	Y
100% Ethyl Alcohol	00:30 min	40	Y
100% Ethyl Alcohol	01:00 min	40	Y
100% Ethyl Alcohol	30 min	40	Y

A	B	C	D
Xylene	30 min	40	Y
Xylene	30 min	40	Y
Paraffin Wax 1	30 min	60	Y
Paraffin Wax 2	30 min	60	Y
Paraffin Wax 3	30 min	60	Y

Table 2. Formalin long program.