

Apr 27, 2023

Version 2

Automation Protocol for High-Efficiency and High-Quality Genomic DNA Extraction from *Saccharomyces Cerevisiae* V.2



DOI

<https://dx.doi.org/10.17504/protocols.io.8epv592p5g1b/v2>

Nina Alperovich¹, David Ross¹, Benjamin Scott²

¹NIST; ²Global Institute for Food Security



David Ross

NIST

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.8epv592p5g1b/v2>

Protocol Citation: Nina Alperovich, David Ross, Benjamin Scott 2023. Automation Protocol for High-Efficiency and High-Quality Genomic DNA Extraction from *Saccharomyces Cerevisiae*. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.8epv592p5g1b/v2> Version created by **David Ross**

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited



Protocol status: Working

We use this protocol and it's working

Created: April 27, 2023

Last Modified: April 27, 2023

Protocol Integer ID: 81101

Keywords: laboratory automation, synthetic biology, *Saccharomyces Cerevisiae*, DNA extraction, quality genomic dna extraction from *saccharomyces cerevisiae*, yeast liquid culture, quality genomic dna extraction, quality gDNA at high efficiency, genomic dna, automated liquid handler, quality gDNA, gDNA from larger volume, protocol for automated extraction, liquid culture, *saccharomyces cerevisiae*, dna, gDNA, microplate shaker, regulated microplate shaker, liquid handler, aliquots of liquid culture, extraction, automated extraction

Disclaimer

Certain commercial equipment, instruments, or materials are identified to adequately specify experimental procedures. Such identification neither implies recommendation nor endorsement by the National Institute of Standards and Technology nor that the equipment, instruments, or materials identified are necessarily the best for the purpose.

Abstract

Here, we describe a protocol for automated extraction of genomic DNA (gDNA) from yeast liquid culture and colonies. The protocol uses a Hamilton NGS-STAR automated liquid handler equipped with thermal-regulated microplate shakers. We have tested this automated protocol with 1.5 mL aliquots of liquid culture ($OD_{600} = 1$) to demonstrate that it can provide high-quality gDNA at high efficiency. The protocol can also be implemented with manual pipetting, and/or to extract gDNA from larger volumes of culture by proportionally increasing the volumes of reagents and using an appropriate magnet separation rack.

Materials

Reagents:

- PBS, pH 7.4 (Invitrogen, cat. no. AM9625)
- SDS, 20% (Quality Biological, cat. no. 351-066-101)
- EDTA, 0.5 mol/L, pH 8.0 (Ambion, cat. no. AM9262)
- TRIS-HCl, 1 mol/L, pH 8.0 (Invitrogen, cat. no. 15568)
- 1x TE Buffer, pH 8.0 (ThermoFisher Scientific, cat. no. 12090015)
- Polyethylene glycol (PEG-8000, Sigma, cat. no. 89510)
- Sodium Chloride, 5 mol/L (Invitrogen, cat. no. AM9760G)
- Zymolyase 20T (US Biological, Cas. No. 37340-57-1)
- 2-Mercaptoethanol (ThermoFisher Scientific, Cas. no. 60-24-2)
- RNase A/T1 Mix (ThermoFisher Scientific, cat. no. EN0551)
- Speed-Bead Magnetic Carboxylate Modified Particle (Cytiva, cat. no. 65152105050250)
- TWEEN-20 (Life technologies, cat. no. 003005)
- Ethanol, 200 proof (100%) (Fisher Scientific, cat. # 2716TR)
- Nuclease-Free Water (ThermoFisher Scientific, cat. no. AM9938)
- Elution Buffer: 10 mmol/L TRIS-HCl, pH 8.0 (Invitrogen, cat. no. 15568)

Additional labware:

- Spin Plate: 96-well deep-well plate (Eppendorf, cat. no. 951033405)
- Reaction Plate: 96-well midi plate (Abgene, AB-0765)
- Elution plate: Low-binding 96-well (Eppendorf, cat. no. 30603303)
- Universal Lid: (Agilent, cat. no. 202497-100)
- 12-Tube Magnetic Separation Rack (New England Biolabs, cat. no. S1509)
- Aluminum Foil Lids (Beckman Coulter, cat. no. 538619)
- Rubber Roller (Beckman Coulter, cat. no. 538619)
- 50 mL Conical Tubes (FALCON, cat. no. 352098)
- Microcentrifuge Tubes, 1.7 mL, RNase/DNase free (Costar, cat. no. 3207/3620)

Equipment:

- Automated liquid handler (Hamilton Robotics, NGS-STAR).
- Centrifuge for 96-well plates.



Preparation of Zymolyase Stock

- 1 Prepare stock of Zymolyase 20T by dissolving to a concentration of 1 U/ μ L in 1x PBS.
- 2 Divide Zymolyase solution into 800 μ L aliquots and store at -20 °C until use.

Preparation of 2-Mercaptoethanol Stock

- 3 Prepare 0.001 % stock of 2-Mercaptoethanol by diluting 1000-fold into 1x PBS.
- 4 Store at 4 °C until use.

Preparation of Digestion Master Mix

- 5 Add 22.72 mL 1x PBS
- 6 Add one aliquot (800 μ L) of Zymolyase Stock, 1 U/ μ L
- 7 Add 160 μ L 0.001 % 2-Mercaptoethanol
- 8 Add 320 μ L RNase A/T1 mix
- 9 Prepare fresh before each use

Preparation of 4x Lysate Buffer

- 10 Add 30 mL Nuclease-Free Water.



- 11 Add 60 mL 20% SDS
- 12 Add 4 mL 0.5 mol/L EDTA.
- 13 Add 6 mL 1 mol/L TRIS-HCL, pH 8.0
- 14 Store at 4 °C until use.

Preparation of Magnetic Beads Stock

- 15 Vortex Speed-Bead Magnetic beads to resuspend
- 16 Divide Speed-Bead Magnetic beads into 1 mL aliquots, each in in a 1.7 mL Microcentrifuge Tube.
- 17 Place micro-centrifuge tubes containing magnetic beads solutions on a 12-Tube Magnetic Separation Rack until beads are drawn to the magnet and solutions are clear.
- 18 Remove supernatant.
- 19 Add 1 mL 1x TE Buffer to each micro-centrifuge tube containing beads, remove from magnetic stand, and mix by vortex.
- 20 Repeated steps 17-19 two additional times.
- 21 Store bead stock in the dark at 4 °C. Bring bead stock to room temperature and mix well by vortexing before use.

Preparation of Magnetic Bead/PEG-NaCl Binding Buffer



- 22 Add 9 g PEG-8000 to a new 50 mL Conical Tube.
- 23 Add 10 mL 5 mol/L Sodium Chloride.
- 24 Add 500 μ L 1 mol/L Tris-HCl, pH8.
- 25 Add 500 μ L 0.5 mol/L EDTA.
- 26 Fill Conical Tube to the 40 mL mark with Nuclease-Free Water, then vortex to mix every 1-2 minutes until PEG goes into solution.
- 27 Add 27.5 μ L Tween-20 and mix gently by inverting 4-6 times.
- 28 Vortex Magnetic Beads Stock (from previous section) to mix thoroughly.
- 29 Add Magnetic Bead Stock:
 - 29.1 For extraction of gDNA from liquid culture, add 3 mL Magnetic Beads Stock to the 50 mL Conical Tube containing the PEG/NaCl solution.
 - 29.2 For extraction of gDNA from yeast colonies, add 1 mL Magnetic Beads Stock to the 50 mL Conical Tube containing the PEG/NaCl solution.
- 30 Fill Conical Tube to the 50 mL mark with Nuclease-Free Water, and gently mix until solution is a uniform brown color.
- 31 Wrap in foil (or place in dark container) and store at 4 °C, mix well by vortexing before use.

Preparation of yeast liquid culture for extraction

- 32 Aliquot 1.5 mL (OD₆₀₀=1) in 1-3 columns of 96-well Spin Plate using automated multichannel pipettes.

- 33 Centrifuge at 3700 g for 10 min at 10 °C.
- 34 Remove supernatant from each well using automated liquid handler.
- 34.1 Use modified labware definition or other constraints on automated pipetting to set a minimum distance of 2.5 mm between the bottom of each well and the tip of each pipette.
- 34.2 The resulting volume of the remaining cell pellet and residual liquid should be approximately 50 µL.
- 35 Seal Spin Plate using Aluminum Foil Lid and Rubber Roller.
- 36 Store at -20 °C.
- 37 Defrost at 4 °C for 30-60 minutes before continuing with extraction.

Preparation of yeast colonies for extraction

- 38 Pick up colonies from agar plate and resuspend each in 50 µL of PBS in 1-3 columns of 96-well Spin Plate.
- 39 Proceed with DNA extraction from resuspended colonies as described below.

Yeast genomic DNA extraction. The following steps are implemented with an automated liquid handler, but can also be implemented with manual pipetting

- 40 Pipette 315 µL of Digestion Master Mix into each well containing a cell pellet or resuspended colony; mix 30 times by repeated aspiration and dispense.
- 41 Transfer 315 µL from each well containing resuspended cell sample to a well in a new Reaction Plate located on a thermal-regulated microplate shaker.

- 42 Incubate at 37 °C for 25 minutes while shaking at 400 rpm.
- 43 Pipette 100 µL of 4x Lysate Buffer to each well and mix 3 times by repeated aspiration and dispense.
- 44 Incubate at 75°C for 10 minutes.
- 45 Move the Reaction Plate to a microplate shaker at room temperature.
- 46 Pipette 200 µL Magnetic Bead/PEG-NaCl Binding Buffer into each well.
- 46.1 Mix Magnetic Bead/PEG-NaCl Binding Buffer well by repeated aspiration and dispense before each transfer.
- 46.2 After dispensing beads into each well, mix beads with cell lysate with one aspiration and dispense.
- 46.3 During this step, the ratio of Magnetic Bead/PEG-NaCl Binding Buffer to the lysed sample volume is 0.48 (200 µL : 415 µL). Consequently, because of the relatively low PEG concentration in the mixture, only high molecular weight chromosomal DNA will bind to the beads.
- 47 Incubate at room temperature for 7 minutes while shaking at 300 rpm.
- 48 Move the Reaction Plate to a 96-well magnet base to pull the magnetic beads out of suspension
- 49 Wait for 5 minutes.
- 50 Remove and discard supernatant from each sample (approximately 615 µL).
- 51 Add 600 µL 80 % Ethanol to each well.



- 52 Wait for 1 minute.
- 53 Remove and discard the ethanol supernatant.
- 54 Add 300 μ L 80 % Ethanol to each well.
- 55 Wait for 1 minute.
- 56 Remove and discard the ethanol supernatant.
- 56.1 After removing most of the supernatant with large-volume tips, use smaller-volume tips (e.g., 50 μ L) to remove any residual supernatant from the bottom of each well.
- 57 Allow magnetic bead pellets to dry for 5 minutes at room temperature.
- 58 Move the Reaction Plate from magnet base to an automated thermal-regulated microplate shaker with temperature set to 60 °C.
- 59 Add 100 μ L Nuclease-Free Water to each well (note: can also use EB)
- 60 Resuspend beads by repeated aspiration and dispense (12 times).
- 61 Incubate at 60 °C for 10 s while shaking at 1000 rpm.
- 62 Incubate at 60 °C for 7 minutes while shaking at 400 rpm.
- 63 Move the Reaction Plate back to the magnet base.
- 64 Wait for 5 minutes.



- 65 Transfer the supernatant from each well to a new, clean well in the same plate (still on the magnet base).
- 66 Wait for 5 minutes.
- 67 Transfer the supernatant with extracted DNA from each well to a well in the Elution plate.
- 68 Cover the Elution plate containing the extracted DNA samples with a Universal Lid.