AUG 14, 2019

Biochemical Measures of Neuropathy - Genotyping

Eva Feldman¹

¹University of Michigan - Ann Arbor

Diabetic Complications Consortium

Tech. support email: rmcindoe@augusta.edu

Lili Liang

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.3qdgms6

External link:

<https://www.diacomp.org/shared/document.aspx?id=54&docType=Protocol>

Protocol Citation: Eva Feldman 2019. Biochemical Measures of Neuropathy - Genotyping. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.3qdgms6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Jun 04, 2019

Last Modified: Aug 14, 2019

PROTOCOL integer ID:
24037

ABSTRACT

Summary:

Oxidative stress is highly correlated with the metabolic changes caused by hyperglycemia. Increased levels of glucose overload mitochondria and result in the production of reactive oxygen species (ROS). In addition, the flow of excess glucose through cellular pathways decreases the cell's normal ability to detoxify ROS. As a result, the neurons and axons of the peripheral nervous system contain increased levels of ROS and decreased antioxidant capacity. The following assays are used to measure these changes in rodent models of diabetic neuropathy.

Diabetic Complication:

Neuropathy

Keywords: Biochemical Measures of Neuropathy, diabetic neuropathy, genotyping

MATERIALS

MATERIALS

✂ 10X PCR Buffer **Life Technologies Catalog #10966-034**

✂ 50 mM MgCl₂ **Life Technologies Catalog #10966-034**

✂ 10 mM dNTPs **Life Technologies Catalog #10297-018**

✂ Taq DNA Polymerase (5 U/μL) **Life Technologies**

✂ 0.5 mL thin-walled tubes **Life Science Products Catalog #LS-9350-X**

✂ Forward primer (10 μM) **Contributed by users**

✂ Reverse primer (10 μM) **Contributed by users**

✂ DNA Template (200 ng/μL) **Contributed by users**

✂ Sterile deionized H₂O **Contributed by users**

Note:

Life Technologies ([RRID:SCR_008817](#))

PCR Amplification:

- 1 Prepare a Master Mix containing the following volumes of reagents:

Reagent	μL
10X PCR Buffer	2.5
50 mM MgCl ₂	1.25
dNTPs	2.5
M13 Forward Primer	1.0
M13 Reverse Primer	1.0
Taq DNA Polymerase	0.2
DNA Template	2.0
Sterile dd-H ₂ O	14.55
Total Volume =	25.00

- 2 Into each 0.5mL thin-walled tube, pipette 23.0-μL of Master Mix.
- 3 Pipet 2.0 μL of DNA template (200 ng/μL) into thin-walled tube. Mix gently, and spin down.

4 Amplify using the following program:

94°C, 5 min. → 94°C, 1 min. → 64°C, 1 min. → 72°C, 1.5 min. → 72°C, 10 min. → 4°C, soak
Repeat 35 times

5 Verify PCR product by separation on 2% agarose.

6 Store tubes at –20°C until needed. If plates will be used within a few days they may be stored at 4°C.