AUG 07, 2019

Catalase

Eva Feldman¹, Frank Brosius²

¹University of Michigan; ²University of Arizona

Diabetic Complications Consortium

Tech. support email: rmcindoe@augusta.edu

Lili Liang

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.3e6gjhe

External link:
<https://www.diacomp.org/shared/document.aspx?id=38&docType=Protocol>

Protocol Citation: Eva Feldman, Frank Brosius 2019. Catalase. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.3e6gjhe>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: May 28, 2019

Last Modified: Aug 07, 2019

PROTOCOL integer ID:
23742

ABSTRACT

Summary:

Describes the protocol used by the DiaComp to detect and quantify catalase activity in a tissue.

Diabetic Complications:

Cardiovascular

Nephropathy

Neuropathy

Retinopathy

Uropathy

Wound-Healing

Pediatric Endocrinology

Keywords: catalase activity, diabete, Cardiovascular, Nephropathy, Neuropathy, Retinopathy, Nephropathy, Uropathy, Wound-Healing, Pediatric Endocrinology

MATERIALS

MATERIALS

Amplex Red Catalase Assay Kit **Molecular Probes Catalog #A-22180**

Reagent Preparation:

Amplex Red reagent: Prepare a 10mM stock solution. (Enough for 2 plates) Bring DMSO and Amplex Red reagent to room temp. Just prior to use dissolve 1 vial (.26mg) of 20mM Amplex Red reagent in 100µL of DMSO. Store stock solution an –20°C, protected from light.

Reaction Buffer (5X) (0.25M sodium phosphate, pH 7.4): Dilute 4mL of Reaction buffer in 16mL of de-ionized water.

HRP (Horseradish peroxidase) 100U/mL: Combine 15µL of 200U/mL HRP stock solution with 15µL of 1X Reaction Buffer. Store frozen at –20°C. **Reagent supplied with kit is 20U. Dissolve content with 200µL 1X Reaction buffer and divide into 22µL aliquots.**

20mM H₂O₂: (Make fresh each time.) Dilute (check bottle for %) 17.9µL H₂O₂ (3.8%) in 982.1µL dH₂O. (Check label for exact concentration) (23µL 3% H₂O₂ into 977µL dH₂O) Use promptly.

Catalase: Prepare a 1000U/mL stock. Reagent supplied with kit is 100U. Dissolve vial in 100µL dH₂O. Aliquot and store at –20°C. Make **10U/mL** with 1µL 1000U/mL stock into 99µL dH₂O. Make **1U/ml** with 10µL 10U/ml into 90µL dH₂O.

1 Sample Preparation:

Prepare Stock solution of Catalase then prepare standard curve as follows:

Volume of Catalase stock	Volume of 1X Buffer	Final Catalase Concentration
0 µL	75 µL	0 mU/mL
18.75 µL of 1 U/mL	56.25 µL	62.5 mU/mL
37.50 µL of 1 U/mL	37.5 µL	125 mU/mL
7.5 µL of 10 U/mL	67.5 µL	250 mU/mL
15 µL of 10 U/mL	60 µL	500 mU/mL
30 µL of 10 U/mL	45 µL	1000 mU/mL

(Final concentration will be fourfold lower, 0 to 10 µM)

TISSUE:

2. 1. Homogenate tissue in 1X Reaction Buffer **on ice**.
2. Using a black plate, pipette 25µL of diluted standards, controls (if any) and samples into wells. *(Final concentration will be fourfold lower, 0 to 10µM)*
3. Prepare stock solution of 20mM H₂O₂ then prepare a 40µM H₂O₂ dilution by adding 10µl of 20mM H₂O₂ to 4.99mL 1X Reaction Buffer.
4. Pipet 25µL of 40µM H₂O₂ solution into each well.
5. Incubate for 30 minutes at room temp.
6. Prepare stock solution of 10mM Amplex Red reagent and divide into 50µL aliquots and freeze immediately.
7. Prepare stock solution of 100U/ml HRP and divide into 20mL aliquots.
8. Prepare 100µM Amplex Red reagent containing 0.4U/mL HRP by adding 50µL of 10mM Amplex Red stock solution and 20µL of 100U/ml HRP stock solution to 4.93mL 1X Reaction Buffer.
9. Begin 2nd phase of reaction by adding 50µL of the above to each well.
10. Place plate into Fluroskan holder and click "**START**".
11. Take 4 readings @ 15 minute intervals using 544/590 filter pairs. (Generally take 3 reading which would be after 30 min. incubation as recommended.)
12. Save raw data as an Excel file into the CTx data folder. Use the naming convention CTXXXX.xls, where XXXX is the date in mmdd format.
13. Select Process>Organize. Choose the appropriate data to organize (usually Measure1), then click "**OK**". This re-arranges the data into columns.
14. Save organized data as an Excel file into the Catalase data folder. Use the naming convention ctXXXXor.xls, where XXXX is the date in mmdd format.