JUN 09, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.wgkfbuw

Protocol Citation: Steven W Wilhelm, Gary Lecleir, Ashley A Humphrey 2019. Cleaning up a biohazardous spill inside a centrifuge. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.wgkfbuw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Dec 17, 2018

Last Modified: Jun 09, 2019

PROTOCOL integer ID:
 18668

Keywords: biohazard, spill, centrifuge

Cleaning up a biohazardous spill inside a centrifuge

In 2 collections

Steven W Wilhelm¹, Gary Lecleir¹, Ashley A Humphrey²

¹The University of Tennessee, Knoxville;

²University of Tennessee, Knoxville

The Aquatic Microbial Ecology Research Group - AMERG (The Buchan, Zinser and Wilhelm labs)

Ashley A Humphrey

University of Tennessee, Knoxville

ABSTRACT

BIOLOGICAL SPILL RESPONSE method to be used in the event of a biosafety incident.

ATTACHMENTS

[SpillCleanupProcedure_2017.07.27 \(1\).pdf](#)

SAFETY WARNINGS

For chemical safety, please refer to the SDS (Safety Data Sheet).

1 Close off the area and allow aerosols to settle.

- 2 Notify others including supervisor.
- 3 Wait 00:30:00 to allow the aerosols to settle.
- 4 Don appropriate PPE: laboratory coat, safety glasses and Nitrile gloves.
- 5 Remove rotors and bucket and place in Biosafety Cabinet.
- 6 Thoroughly disinfect the inside and outside of the centrifuge rotor, cups and accessories and allow proper contact time. After disinfection, move to sink for a thorough rinse, dry thoroughly.
- 7 For the centrifuge interior, follow the steps above for cleaning up a spill outside a Biosafety Cabinet.
- 8 Segregate contaminated cleanup materials into the appropriate biohazardous waste containers.