SEP 02, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.6wehfbe

External link:
<https://doi.org/10.1371/journal.pone.0224822>

Protocol Citation: Xiaohua Du, Xia Liu, Mawolo James Blackar, Yingjie Zhou, Haifeng Wang, Fayang Liu, Zhiqing He 2019. Coulter Method for the detection of Physiological parameters. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.6wehfbe>

MANUSCRIPT CITATION:
 Liu X, Mawolo JB, Du X, Zhou Y, Wang H, Liu F, He Z, Marela HA (2019) Investigation of biochemical and physiological parameters of the newborn Saiga antelope (*Saiga tatarica*) in Gansu Province, China. PLoS ONE 14(11): e0224822. doi: [10.1371/journal.pone.0224822](https://doi.org/10.1371/journal.pone.0224822)

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Coulter Method for the detection of Physiological parameters

PLOS One

Xiaohua Du¹, Xia Liu², Mawolo James Blackar², Yingjie Zhou³, Haifeng Wang², Fayang Liu², Zhiqing He⁴

¹Faculty of Veterinary Medicine, Gansu Agricultural University, Lanzhou City, Gansu Province, People's Republic of China;

²College of Life Science and Technology, Gansu Agricultural University, Lanzhou City, Gansu Province, People's Republic of China;

³Gansu Endangered Animal Protection Center of State Forestry Administration, Wuwei Town, Gansu Province, People's Republic of China;

⁴Gansu Endangered Animal Protection Center of State Forestry Administration, Wuwei City, Gansu Province, People's Republic of China

Du Xiaohua

ABSTRACT

Instrument

A **Coulter counter** ^[1] ^[2] is an apparatus for counting and sizing particles suspended in electrolytes. It is used for cells, bacteria, prokaryotic cells and virus particles. ^[3] A typical Coulter counter has one or more microchannels that separate two chambers containing electrolyte solutions. As fluid containing particles or cells is drawn through each microchannel, each particle causes a brief change to the electrical resistance of the liquid. The counter detects these changes in electrical resistance.

Coulter principle

The *Coulter principle* states that particles pulled through an orifice, concurrent with an electric current, produce a change in impedance that is proportional to the volume of the particle traversing the orifice. This pulse in impedance originates from the displacement of electrolyte caused by the particle. The Coulter principle was named for its inventor, Wallace H. Coulter. The principle has found commercial success in the medical industry, particularly in hematology, where it can be applied to count and size the various cells that make up whole blood.

Cells, being poorly conductive particles, alter the effective cross-section of the conductive microchannel. If these particles are less conductive than the surrounding

Protocol status: Working
We use this protocol and it's working

Created: Aug 28, 2019

Last Modified: Sep 02, 2019

PROTOCOL integer ID:
27302

liquid medium, the electrical resistance across the channel increases, causing the electric current passing across the channel to briefly decrease. By monitoring such pulses in electric current, the number of particles for a given volume of fluid can be counted. The size of the electric current change is related to the size of the particle, enabling a particle size distribution to be measured, which can be correlated to mobility, surface charge, and concentration of the particles.

The Coulter Counter is a vital constituent of today's hospital laboratory. Its primary function being the quick and accurate analysis of complete blood counts (often referred to as CBC). The CBC is used to determine the number or proportion of white and red blood cells in the body. Previously, this procedure involved preparing a blood cell stain and manually counting each type of cell under a microscope, a process that typically took a half-hour.

Coulter Counters have a wide variety of applications including paint, ceramics, glass, molten metals and food manufacture. They are also routinely employed for quality control.

A Coulter counter played an important role in the development of the first ever cell sorter, and was involved in the early days of the development of flow cytometry. Even today, some flow cytometers utilize the Coulter Principle to provide highly accurate information about cell size and count.

Many investigators have designed a variety of devices based on the Coulter Principle, and generated peer-reviewed publications featuring data generated by these devices. A few of these devices have also been commercialized. All implementations of the Coulter Principle feature trade offs between sensitivity, noise shielding, solvent compatibility, speed of measurement, sample volume, dynamic range, and reliability of device manufacture.

ATTACHMENTS

[Coulter Method for the detection of Physiological parameters.docx](#)

GUIDELINES

Development

Wallace H Coulter discovered the Coulter Principle in the late 1940s (though a patent was not awarded until October 20, 1953). Coulter was influenced by the atomic bombs dropped on Hiroshima and Nagasaki. These events motivated Coulter to simplify and improve blood cell analysis so that large populations could be screened rapidly, as would be necessary in the event of a nuclear war. Partial funding of the project came from a grant award from the Office of Naval Research.^{[4][5]}

"Coulter Principle" refers to the use of an electric field for counting and sizing dilute suspensions of particles in conducting liquids. Wallace H. Coulter was awarded US Patent *#2,656,508, Means for Counting Particles Suspended in a Fluid*. The Coulter Principle is most commonly employed in a Coulter counter, which is an analytical instrument designed for a specific task such as counting cells. However, there are numerous other ways to implement the Coulter Principle. Several of these have been attempted, some with commercial success, and some purely for academic research. To date, the most commercially successful application of the Coulter Principle is in hematology, where it is used to obtain information about patients' blood cells.

The Coulter Principle relies on the fact that particles moving in an electric field cause measurable disturbances in that field. The magnitudes of these disturbances are proportional to the size of the particles in the field. Coulter identified several requirements necessary for practical application of this phenomenon. First, the particles should be suspended in a conducting liquid. Second, the electrical field should be physically constricted so that the movement of particles in the field causes detectable changes in the current. Finally, the particles should be dilute enough so that only one at a time passes through the physical constriction, preventing an artifact known as coincidence.

While the Coulter Principle can be implemented in a variety of designs, there are two that have become the most commercially relevant. These include an aperture format and a flow cell format.

Experimental Procedures

Coincidence

Anomalous electrical pulses were generated because the concentration of samples were so high that multiple particles entered the aperture simultaneously. This situation is known as coincidence. This condition occurred because there was almost no way to ensure that a single large pulse is the result of a single large particle or multiple small particles entering the aperture at once. To prevent this situation, samples must be fairly dilute, but we didn't dilute our samples because we needed accurate results.

Particle path

The shape of the generated electrical pulse varied with the particle path through the aperture. Shoulders and other artifacts occurred because the electric field density varied across the diameter of the aperture. This variance is a result of both the physical constriction of the electric field and also the fact that the liquid velocity

varied as a function of radial location in the aperture. In the flow cell format, this effect was minimized since sheath flow ensured each particle traveled an almost identical path through the flow cell. In the aperture format, we used the signal processing algorithms to correct the artifacts resulting from particle path.

Conductive particles

Conductive particles are often common concern for considering the Coulter Principle. Although, there are interesting scientific questions, on this procedure, it rarely affects the results of an experiment. This is because the conductivity difference between most conductive materials and ions in liquid (referred to as the discharge potential) is so great that most conductive materials act as insulators in a Coulter counter. The voltage we used to break down this potential barrier is referred to as the breakdown voltage. For those highly conductive materials that presented problems, the voltage we used during our Coulter experiment reduced below the breakdown potential (which was determined empirically).

Porous particles

We use the Coulter principle to measured the volume of our samples, since the disturbance in the electric field is proportional to the volume of electrolyte displaced from the aperture. This medium has led to confusion amongst researchers who are use to optical measurements from microscopes or other systems that only view two dimensions and also show the boundaries of a sample. The Coulter Principle, on the other hand measured three dimensions and the volume displaced by a sample.

Direct current vs alternating current

Direct current has been used in the Coulter counters found in most research and cell laboratories. Direct current measurements are useful for an array of particles and allow for simplified data acquisition and processing. Base on this, we used direct current to simplified and processed data acquisition. Alternating current measurements are sometimes used in clinical hematology instruments, due to the special nature of cell membranes. At low frequencies (below 500 kHz), alternating and direct current measurements behave essentially the same way. At intermediate frequencies (500 kHz - 6 MHz), the plasma membrane of cells can become polarized, leading to a decreased capacitance of the measurement systems. However, at high frequencies (6-20 MHz), the cell membrane loses its polarization, and the electrical pulses provide information about the cell cytoplasm.

Major applications

Hematology/Physiology

The most successful and important application of the Coulter Principle is in the characterization of blood cells. The technique has been used to diagnose a variety of diseases, and is the standard method for obtaining red blood cell counts (RBCs) and white blood cell counts (WBCs) as well as several other common parameters. When combined with other technologies such as fluorescence tagging and light scattering, the Coulter Principle can help produce a detailed profile of patients' blood cells.

Cell count and size

In addition to clinical counting of blood cells (cell diameters of ~6-10 micrometres, typically), the Coulter principle has established itself as the most reliable laboratory method for counting a wide variety of cells, ranging from bacteria (< 1 micrometre in size), fat cells (~400 micrometre), plant cell aggregates (>~1200 micrometre), and stem cell embryoid bodies (~900 micrometre). The technique has become so standardized that ASTM International has published a procedure on the topic: *ASTMF2149-01(2007) Standard Test Method for Automated Analyses of Cells-the Electrical Sensing Zone Method of Enumerating and Sizing Single Cell Suspensions*.

Particle characterization

The Coulter Method has proved useful for applications well beyond cellular studies. The fact that it individually measures particles, is independent of any optical properties, is extremely sensitive, and is very reproducible has appeal to a wide variety of fields. Consequently, the Coulter Principle has been adapted to the nanoscale to produce a novel nanoparticle characterization technique called Tunable Resistive Pulse Sensing, or TRPS. TRPS enables high-fidelity analysis of a diverse set of nanoparticles, including (but not limited to): functionalized drug delivery nanoparticles, Virus-like particles (VLPs), liposomes, exosomes, polymeric nanoparticles, microbubbles.

Benefits of Coulter Method

- Increase productivity with consistently reliable results.
- With its versatile closed tube sampling system, the coulter save time and enhance safety for laboratorians.
- Flexible specially-formulated reagents, fully automated QC and calibration platforms provide consistently reliable results.

References

1. W.R. Hogg, W. Coulter; Apparatus and method for measuring a dividing particle size of a particulate system; United States Patent 3557352

2. U.S. Patent 7,397,232 Coulter counter
3. R.W. DeBlois, C.P. Bean (1970). "Counting and sizing of submicron particles by the resistive pulse technique". Review of Scientific Instruments. 41 (7): 909–916.
4. Bibcode:1970RSci...41..909D. doi:10.1063/1.1684724. Marshall Don. Graham (2003). "The Coulter Principle: Foundation of an Industry". Journal of Laboratory Automation. 8 (6): 72–81.
5. Cytometry volume 10, a DVD series produced by the Purdue University Cytometry Labs
<http://www.cyto.purdue.edu/cdroms/cyto10a/seminalcontributions/coulter.html>

SAFETY WARNINGS

Safety methods to follow during our experiment at the laboratory:

1. Wear lab coat and gloves while in the lab. When you enter the lab, switch on exhausted fans and make sure that all the chemicals and reagents required for the experiment are available.
2. Clean all working apparatus with chromic acid and distilled water and ensure that all the apparatus are free from water droplets while performing the experiment.
3. Calibrated the electronic weigh balance before taking the measurements.
4. Ensure that the spectrophotometer working properly was done.
5. Make sure the cuvette is handled with tissue paper and didn't touch it with your hands.
6. Wipe the cuvette with tissue paper before placing in the spectrophotometer.
7. Clean all apparatus with soap and distilled water. Upon completion of the experiment, recap the reagent bottles. Switch off the light and exhaust fan before leaving the lab.
8. Discarded the used gloves in a waste bin.