CoVIDScandx V.1

Muhammad Munir¹

¹The Lancaster University

muhammad.munir

VERSION 1

SEP 07, 2020

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bkzikx4e

Protocol Citation: Muhammad Munir 2020. CoVIDScandx. protocols.io

<https://dx.doi.org/10.17504/protocols.io.bkzikx4e>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: Sep 07, 2020

Last Modified: Sep 08, 2020

PROTOCOL integer ID:

41738

1 Add smart lysis buffer with 1:1 ratio in the VTM sample

- 2 Heat for 68C for 8 minutes and then 95C for 10 minutes
- 3 Cool and take 3 ul to colorimetric LAMP master mix (NB: M1800)
- 4 Vortex reaction mix and centrifuge to collect material
- 5 Pipet 24 µl per reaction into desired reaction vessels and add 1 µl of sample. Mix by vortexing or by pipetting if using a plate or similar vessel, centrifuge to collect if necessary. Check that reaction solutions have a bright pink colour, which indicates initial high pH required for successful pH-LAMP reaction.
- 6 Incubate at 65°C for 30 minutes.
- 7 Cool on ice for 2 minutes
- 8 Add 5 ul of the LAMP reaction with the sample buffer
- 9 Load to the LFD strip and let it absorb

10 Visualize the bands for internal control and SARS-CoV-2

11 Scan with mReader to record results