JAN 06, 2020

Deep Cleaning Bacteria Incubator

Kenneth Schackart¹

¹University of Arizona

Yoon Lab

Kenneth Schackart
University of Arizona

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.ba2tigen

External link:
https://www.marshallscientific.com/v/vspfiles/files/manuals/Fisher_Scientific_Isotemp600_Oven-manual.pdf

Protocol Citation: Kenneth Schackart 2020. Deep Cleaning Bacteria Incubator. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.ba2tigen>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Jan 06, 2020

Last Modified: Jan 06, 2020

PROTOCOL integer ID:
31539

ABSTRACT

This protocol details how to deep-clean the bacteria incubator (Fisher Scientific Isotemp 637F). This cleaning should be performed at least every 4 months during regular usage, more frequently under heavy usage. Full Owner's Manual is attached to this protocol as a link (I do not claim any credit in the development of that document; it is for reference only)

GUIDELINES

Gloves must be worn at all times.
Do not use bleach on stainless steel.

MATERIALS

- VersaClean Laboratory Cleaner solution (Concentration of 30:1 in warm water)
- 70% Ethanol Solution (EtOH)
- Sponge
- Screwdriver
- Waste beaker

SAFETY WARNINGS

Do not turn incubator back on or close door until the interior is completely dry – 24 hours.

Disassemble

- 1 Before beginning any work on the incubator, turn off the power using the switch at the bottom.

- 2 Remove incubator door:
 - Loosen the screws on the inside of the hinges (hold the door so it doesn't fall when released).
 - Slide door out of the hinges by pulling outward.
 - The hinge on the top will fall out when the door is removed, or come off with the door.
 - Each hinge has a small plastic bushing the goes between the hinge and the incubator; do not lose these bushings.
 - Set aside for now on lab bench.

You can see the white bushing directly on top of the right of the hinge.

- 3 Remove rack shelving by pulling straight outward. Set aside for now on lab bench.

- 4 Remove the shelf holders:
 - Make note of the placement of each shelf at this time.

- Each horizontal shelf holder can be removed by lifting up, rotating the bottom up/out, then sliding the top downward and out.
- Set aside for now on lab bench.

5 Remove vertical mounts:

- Make note of the orientation of the mounts at this time.
- Remove the screws holding the mount in place; 2 screws per mount.
- Set aside for now on lab bench.

6 Remove the bottom tray:

- Undo the 2 screws holding the tray in place.
- Lift the tray straight up to remove (this may be assisted by inserting a small curved metal pick into the holes to help lift the back end up).
- Set aside for now on lab bench.

Wash

- 7 Using a sponge, wash all interior surfaces of the incubator with VersaClean solution (or other laboratory cleaner - do not use bleach on stainless steel).
 - My recommendation is to dip the sponge in cleaner, and scrub one surface (e.g. the top) with VersaClean, then wring sponge into a waste container and rewipe surface. This should leave the surfaces only slightly damp, not wet.
 - Work from the top-down: top, sides, then bottom last.
 - Avoid the mechanical moving parts on the bottom, or be cautious in washing them.
- 8 Spray all interior surfaces with EtOH and wipe with paper towels, but avoid directly spraying the mechanical moving parts on the bottom.
- 9 Using a sponge with VersaClean, wash the door and all interior parts including the rack mounts, racks, and shelf holders. Spray with EtOH to disinfect just prior to reinstalling.

Reassemble

- 10 Reassemble the incubator by reversing the disassembly instructions.

-
- 11** Allow the incubator to completely dry before closing door or powering on, especially since ethanol vapor is flammable!