OCT 23, 2020

Detection and Sorting of Extracellular Vesicles and Viruses using nanoFACS

Aizea Morales-Kastresana¹,

Joshua A Welsh¹, Jennifer Jones¹

¹Translational Nanobiology Section, Laboratory of Pathology, Center for Cancer Research, National Cancer Institute, National Institutes of Health

Translational Nanobiology Section

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bj6xkrfn

Protocol Citation: Aizea Morales-Kastresana, Joshua A Welsh, Jennifer Jones 2020. Detection and Sorting of Extracellular Vesicles and Viruses using nanoFACS. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bj6xkrfn>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Aug 22, 2020

Last Modified: Oct 23, 2020

PROTOCOL integer ID:
 40887

Joshua A Welsh

DISCLAIMER

This protocol summarizes key steps for a specific type of assay, which is one of a collection of assays used for EV analysis in the NCI Translational Nanobiology Section at the time of submission of this protocol. Appropriate use of this protocol requires careful, cohesive integration with other methods for EV production, isolation, and characterization.

Keywords: Astrios EQ, jet-in-air, small particle, flow cytometry, extracellular vesicles, flow virometry, nanoFACS

ABSTRACT

Recent advances in high resolution flow cytometry (HRFC), which show improvements in both light scatter and fluorescence sensitivity have resulted in the development of techniques that better isolate, stain and analyze single EVs (Boing et al., 2014; Groot Kormelink et al., 2016; Morales-Kastresana, Musich, Welsh, Telford, Demberg, Wood, Bigos, Ross, Kachynski, Dean, Fetton, et al., 2019; Morales-Kastresana et al., 2017; Stoner et al., 2016; van der Vlist, Nolte-'t Hoen, Stoorvogel, Arkesteijn, & Wauben, 2012). Below, we describe protocols to fluorescently label EVs using CFDA-SE (hereinafter called CFSE), as well as antibodies targeted at specific EV surface proteins. We also provide guidelines for residual dye and antibody removal, appropriate data acquisition by HRFC and EV counting by HRFC. **Figure 1** summarizes these protocols.

The EVs used in this protocol are derived from the DC2.4 cell line, and bone marrow derived dendritic cells (BMDCs). The DC2.4 cell line are immature dendritic cells (DCs) with very low expression of typical DC markers on their surface (unpublished observation and (Hargadon, Forrest, & Reddy, 2012)) and that release a morphologically homogeneous population of EVs (~130 nm in diameter). DC2.4 EVs will be used to demonstrate a CFSE staining in **Basic Protocol 1**, as well as being used as a negative control for antibody-based staining methods (**Basic Protocol 2**). Bone marrow dendritic cell (BMDC)-derived EVs are more heterogeneous in diameter (100-200 nm) and composition (Morales-Kastresana, Musich, Welsh, Telford, Demberg, Wood, Bigos, Ross, Kachynski, Dean, Felton, et al., 2019), and express DC markers such as MHC-II. BMDC EVs will be used to demonstrate antigen-specific staining with fluorochrome-conjugated antibodies in **Basic Protocol 2**. DC2.4 and BMDC-derived EVs were isolated by serial ultracentrifugation, with concentration and diameter distribution characterized by NTA, as described before (Morales-Kastresana, Musich, Welsh, Telford, Demberg, Wood, Bigos, Ross, Kachynski, Dean, Fetton, et al., 2019; Morales-Kastresana et al., 2017).

Basic Protocol 1

1

Protocol

NAME

Bulk EV staining with CFSE protein binding dye

CREATED BY

Joshua A Welsh

PREVIEW

- 1.1** Prepare 15 μL of DPBS containing between $\sim 1 \times 10^8$ to $\sim 2.5 \times 10^9$ DC2.4-derived EVs in a 1.7 ml microfuge tube.

Note

Note: This number of EVs correspond to ~ 0.1 - $2.5 \mu\text{L}$ of EV stock, if prepared as described previously (Morales-Kastresana et al., 2017), starting with 60 ml of supernatant from cells cultured for 48h in EV depleted medium. The reaction volume and CFSE concentrations are optimized for staining approximately 1×10^9 EVs. The use of different EV numbers may therefore result in suboptimal staining. When larger amounts of labeled EVs are required, this protocol can be scaled up with similar results.

- 1.2** In a separate 1.7 ml microfuge tube, prepare 15 μL of DPBS containing 80 μM of CFDA-SE, from a 10 mM stock solution of CFDA-SE in DMSO. For this, add 0.12 μL of 10 mM CFDA-SE in 15 μL of DPBS. Intermediate dilutions in DPBS may be prepared to facilitate the dilution process.

Note

Note: It is recommended to store the CFDA-SE stock reconstituted in DMSO in small aliquots at -80°C . Avoid using DPBS or other aqueous solutions for storing purposes, since aqueous solutions favor the hydrolysis of diacetic groups in CFDA-SE and therefore decrease the permeability of the dye and incorporation into EVs (Banks et al., 2013; Bergsdorf et al., 2003; Hoefel et al., 2003). Also due to this risk of hydrolysis, it is critical for the aliquots to be stored in an anhydrous manner. If the EV number and reaction volume are scaled up in step 1, scale up the CFDA-SE quantity, so as to achieve an 80 μM CFDA-SE solution.

- 1.3** Pipette the CFDA-SE solution on top of the EV solution. Mix the solution by pipetting and incubate for two hours at 37°C in the dark. The incubated CFSE concentration is now 40 μM .

Note

Note: The incubation time may be extended to increase CFDA-SE's incorporation into EVs. However, the authors have observed a decrease in the EV number after long incubation periods (Morales-Kastresana et al., 2017).

- 1.4** 15 minutes before the incubation completion time, wash a NAP-5 size exclusion chromatography (SEC) column with 10 ml of DPBS. Never allow the column to dry.

Note

Note: To automate the washing process, a pump can be setup to help add DPBS onto the column.

- 1.5** Prepare collection tubes for the collection of two fractions. To facilitate the visualization of eluted sample, use a marker pen to draw a line indicating the 500 μ L mark on each collection tube.

Note

Note: The first fraction is the “dead volume,” of buffer alone, that elutes before fractions containing material from the loaded sample. The majority of DC2.4 EVs appear in fraction 2. Some EVs may however elute in fraction 3. Free dye elutes in fractions ~7-8.

- 1.6** When the two-hour incubation period is complete, increase the CFSE-stained EV preparation volume to 100 μ L by adding 70 μ L of DPBS and mix by pipetting.

Note

Note: If the staining is scaled up in step 1 to increase the number of EVs, then add DPBS to a final volume of 100 μ L. If the total volume is higher than 100 μ L, use multiple columns to wash the sample.

- 1.7** Pipette the 100 μ L of CFSE-stained EVs on to the SEC column and immediately start collecting 500 μ L fractions.

- 1.8** When the 100 μ L of sample has completely entered the column bed, add 500 μ L of DPBS and continue collecting fractions. ~80% of the eluted EVs will be collected in fraction 2.

Representative contour plots of unstained and CFSE-stained DC2.4 EV (before and after SEC). The number of fluorescein molecules incorporated in the EVs is shown in the plot (protocol for calculating the MESF is described in Basic Protocol 6). Red box indicates system reference noise.

Basic Protocol 2

2

Protocol

NAME

Antigen-specific staining of EV markers with fluorochrome-conjugated antibodies

CREATED BY

Joshua A Welsh

PREVIEW

- 2.1 Wash one qEV column per EV preparation with 20 mL of DPBS. Never allow the columns to become dry.
- 2.2 Pipette 1×10^9 EVs in a 10 μ L volume of DPBS and add 2 μ g of Fc Block reagent to block Fc receptors. Incubate with no agitation for 10 minutes at room temperature.

Note

Note: The presence of Fc receptors on EVs is not well documented. However, adding Fc Block will not only block putative Fc receptors, but also serves as a source of protein to block other non-specific binding sites of fluorescent antibodies.

- 2.3** In a 1.7 mL microfuge tube, pipet 1.5 µg of fluorochrome-conjugated antibody and add DPBS to a final volume of 120 µL per sample. Mix by pipetting up and down. Prepare a master mix if multiple samples are to be stained with the same antibody.

Note

Note: This antibody quantity is a reference starting point when testing a new antibody. Antibody titration is recommended to achieve optimal staining and avoid the use of unnecessary material. Many anti-human antibodies are provided in a test volume format (µL per test) rather than in concentration (µg mL⁻¹).

- 2.4** Transfer the 120 µL of the antibody solution to an airfuge tube and mark one side of the tube with a waterproof marker. Place the tube with a corresponding balance into an A100/18 rotor, with the mark facing up. Place a lid on the rotor.

Note

Note: The mark is a reference for the location of the antibody aggregates after airfuging. Using the rotor cover can reduce sample evaporation during centrifugation.

- 2.5** Place the rotor into the airfuge and close the airfuge lid tightly. Open the air source to centrifuge until the gauge reads 22 psi (~130,000 RCF) and leave for 5 minutes.

Note

Note: In order to avoid extreme heat during centrifugation, airfuge step can be performed in a cold room. Alternatively, the authors cool down the rotor before using it.

- 2.6** When the airfuge step is complete, use sharp tweezers to remove the tubes from the rotor and place them on the corresponding rack.

2.7 Gently pipet off the top 70 μ L of solution and add it on top of the EV solution.

Note

Note: Antibody aggregate pellets cannot always be observed. For that reason, leaving a reasonable volume in the bottom of the tube and using the top part of the solution is recommended.

2.8 Incubate the EVs with antibody 15-30 minutes in the dark at room temperature whilst being gently agitated.

Note

Note: As with CFSE, time is a parameter that can be modulated to increase the labeling with antibodies. The authors have observed slight improvements of staining with certain epitopes when increasing the staining period up to 1 hour.

2.9 Prepare collection tubes for 12 fractions. To facilitate the visualization of eluted sample, use a marker to draw a line indicating the volume of each fraction (500 μ L) on the side of the collection tubes.

2.10 Add DPBS to the EV prep to a final volume of 500 μ L and proceed to remove unbound antibody using SEC with qEV columns. Samples that are not going to be immediately loaded on the columns can be stored at in the dark at 4°C.

2.11 Wait until all of the DPBS used for pre-washing the column has entered the column bed. Immediately load 500 μ L of the sample and simultaneously start collecting 500 μ L fractions.

2.12 Keep adding DPBS (500 μ L each time) and whilst collecting fractions. Stained EVs will start eluting in fraction 7, with the majority in fractions 8-9. For maximum recovery, harvest fraction 10 too.

2.13 Store EVs at 4°C and in the dark until performing flow cytometric analysis. Alternatively, some antibody-fluorochrome conjugates can resist one freeze/thaw cycle and therefore,

labeled EVs can be stored at -80°C if being analyzed at a later date.

2.14 qEV columns can be stored at 4°C and reused with extensive washing. Authors recommend washing them with a minimum 50 ml of DPBS, to elute as much remaining antibody as possible, followed by 10 ml of 20% ethanol diluted in DPBS to keep the columns aseptic during storage. When reusing a column, wash 40 ml of DPBS to make sure that any traces of ethanol are removed.

2.15 Proceed to analysis.

Representative contour plots show PBS, unstained, and MHCII-stained BMDC EV (before and after SEC) and control DC2.4 EVs that lack of MHCII on their surface. Red box indicates system reference noise.

Basic Protocol 3

3

Protocol

NAME

Astrios EQ instrument setup and sample acquisition

CREATED BY

Joshua A Welsh

PREVIEW

3.1 Turn on the instrument at least an hour before running the samples. Let the pressure and stream to stabilize and prime the fluidics system to remove bubbles in the circuit. Turn lasers on and allow them to warm up whilst shuttered.

3.2 Wash the sample line with FACS Rinse solution for about 20 minutes at a high differential pressure (1 psi over the sheath pressure). Repeat same procedure with clean DPBS for 20 more minutes.

3.3 Adjust the vertical alignment.

Note

Note: To test whether the stream is vertical or not, raise the nozzle while looking at the relative position of the stream to the pinholes. If the position doesn't change, the stream is vertically well aligned; if it does change, then the verticality must be tweaked. A good vertical alignment maximizes the detection of parameters in all laser paths.

3.4 Set the triggering threshold to the 561-SSC channel. Adjust the triggering threshold channel and voltage to allow the visualization of the noise population, to approximately 10,000 events per second.

Note

Note: The noise is a random representation of the diffusely scattered photons from the laser beam and stream intercept. Because of the high event rate of these low-level signals, inclusion of these events on some instruments is only feasible in a very limited way due to limitations in their baseline restoration algorithms and other signal processing attributes. Background noise is informative because: i) it serves as a window into the population of EVs that fall under the triggering-threshold; ii) it allows the determination of too much free dye in the interrogation point; and iii) it helps to identify when EV samples are being analyzed at a concentration that is too high and is therefore at risk of coincident detection (Morales-Kastresana et al., 2017; van der Pol, van Gemert, Sturk, Nieuwland, & van Leeuwen, 2012). For these reasons, the authors refer to the noise as "background reference noise".

3.5 Load a sample containing a mix of 200 nm yellow and red beads at 1×10^7 beads mL^{-1} (about a 1×10^6 -fold dilution of original stock) to fine-tune the stream:laser alignment.

Note

Note: Any combination of beads that are excited by different lasers can be used. The goal is to have two populations, whose fluorescence will be collected in different pinholes, in a way that the stream is aligned according to two pinholes. This ensures the correct vertical alignment.

Note

Note: The dimmer the fluorescent beads used, the better the alignment will be for EVs.

- 3.6** Open a dot plot depicting yellow (~515 nm) and red (~605 nm) fluorescent channels in each axis (or correspondent fluorescent axis for the chosen beads). Tweak the alignment until the fluorescence signal is optimized for both bead sets. While doing this, try to keep the total event rate, including the noise, around 10,000-20,000.

Note

Note: Alignment will be optimal when the distance between the noise and bead populations is the biggest in terms of fluorescence, while the bead population remains as tight as possible. Also, the event rate should not increase significantly with respect to DPBS alone, since the contribution of the beads to the overall rate is insignificant.

Note

Note: It is very useful to monitor Time versus any parameter (fluorescence and scatter) to determine how the distance between noise and beads varies with alignment adjustments.

a)

During alignment

b)

Before alignment

c)

After alignment

Plots with the time parameter on the X-axis and scatter or fluorescent on the Y-axis during the alignment process to monitor the relative separation between the beads and noise (a). b and c show scatter and fluorescent parameters before (b) and after (c) optimizing the alignment. YB indicates 200 nm yellow fluorescent beads, RB indicates 200 nm red fluorescent beads.

Note

Note: If the bead populations and the reference noise population appear as “split” populations, there is probably drop-drive noise (especially with the 70 µm nozzle), which can be eliminated by dropping the break off point by lowering the amplitude or frequency, to the extent possible while maintaining stable Intelli-Sort settings and a stable breakoff point.

- 3.7** Once aligned, acquire and save a representative sample of the beads used for the alignment.

Note

This file will serve as a reference for future alignment, allowing comparisons between the alignment between experiments.

- 3.8** Wash the sample line with DPBS for 5 minutes at 1 psi.

Note

Note: As the sample lines clear, monitor the disappearance of the 200 nm beads from the alignment steps over time by referring to the time versus scatter or fluorescence plots. If alignment beads do not disappear completely, increase the pressure up to a differential pressure of 1.4. If beads are still observed, there may be contaminating beads in the sample path (either in the lines, junction points between fittings, or in the nozzle reservoir), and you will need to clear these residual beads before continuing. It can be useful to wash the line with rinse solution and followed by DPBS for 5 minutes each. Once washing is done, change rinse and DPBS tubes for future uses, as these solutions may contain contaminating beads from the sample injection tube.

- 3.9** Save an acquisition of 30 seconds of clean DPBS.

- 3.10** Load the EV sample, increase pressure up to 1 psi to accelerate the sample loading and then decrease the differential pressure to 0.3 psi.

Note

Note: A 0.3 psi differential pressure is offered here as a reference for what commonly works on AstriosEQ sorters that we have worked with. A differential pressure that permits stable detection of single EVs, avoiding coincidence, should be selected. Authors recommend keeping the differential pressure the same for all samples to be analyzed during an experiment, since event rate can be a valuable source of information.

- 3.11** Wait until the event rate is stable (use the time versus scatter plot). Then stop the acquisition in the software (not the machine) and start acquiring again for 30 seconds. Save the acquisition.

Note

Note: Authors recommend saving acquisitions of the same duration, in order to compare event numbers among samples.

- 3.12** Wash the sample line with DPBS for 5 minutes at 1 psi.

- 3.13** Run samples, save acquisitions and wash with DPBS, as described above. *Keep samples at 4°C and in the dark.*

- 3.14** Wash the sample line with FACSClean, FACS Rinse and DPBS consecutively, for 10 minutes each at 2 psi. Turn off the lasers and fluidics. Rinse the nozzle and place a FACS tube cap with some clean water under the nozzle tip.

3.15

ARTIFACT	SOURCE	WHAT TO DO
False positive signal	Drop drive noise	Reduce amplitude and/or frequency

Nonspecific binding of antibodies	Add control EVs that do NOT express the target antigen	
Antibody/dye aggregates	Add control EVs that do NOT express the target antigen	
Add control antibody/dye alone		
Background antibody/dye fluorescence	Add control antibody/dye alone	
Noise fluorescence shifting		
Low resolution of positive signal over the noise/negative population	Background antibody/dye fluorescence	Wash the excess antibody/dye
No positive signal detection	Low antigen density	Use a more sensitive instrument, brighter fluorochrome and/or brighter staining method
Low antibody affinity	Increase concentration and/or incubation time	
Antibody/fluorochrome conjugate not working properly	Change antibody/clone/fluorochrome/lot	
Unexpected increase in event rate	Change in differential pressure of sheath tank and sample line	Time vs scatter parameter to identify these peak
Stuck material in sample line/nozzle	Flush sample line with detergent and PBS with increased sample pressure differential (boost).	
Contaminating material from previous sample	Run filtered PBS before and after samples, to minimize the presence of contaminating particles.	
Coincident detection of particles	Sample is too concentrated	Dilute the sample

Use spike in beads to ensure an operational concentration range		
Nanobubbles at the interrogation point	Are created due to the high pressure in the nozzle	Run PBS control between samples to have that source of noise identified and controlled

A list of the most frequent artifacts observed by the authors is listed, along with a brief explanation and the suggested controls to consider, either to identify or to avoid such artifacts.

Basic Protocol 4

4

Protocol

NAME

Counting particles and EVs on Astrios EQ with spike in reference beads

CREATED BY

Joshua A Welsh

PREVIEW

- 4.1 Prepare a stock of 200 nm polystyrene reference beads by diluting 50 μ L of Fluosphere beads in 50 ml of DPBS (1000-fold dilution) and keep this solution as a '*Big Stock*'. Determine the particle concentration of the beads in the '*Big Stock*' using NTA or any other method.

Note

Companies provide an estimated concentration of beads in solution, based on the size and weight. This tends to be a rough estimation, we therefore recommend preparing a 10,000x dilution of 'Big Stock' beads in DPBS to perform quantification of particles per ml. Polystyrene beads are stable in DPBS, with no significant loss of particles and fluorescence over time. Polystyrene particles can therefore be quantified once and used in the next couple of years. As a reference, the authors' 'Big Stock' of 200 nm yellow-green beads listed in the Material section is at $\sim 1.8 \times 10^{10}$ particles/ml, measured by NTA. It is worth noting that every time an aliquot is taken from the 'Big Stock' its concentration will likely decrease. It is therefore recommended that several smaller 'Big Stock' aliquots to ensure better consistency over long periods of time.

- 4.2 Vortex the bead stock solution for 5 seconds.
- 4.3 Prepare a working stock of beads by diluting them 100-fold in DPBS ($\sim 1.8 \times 10^8$ particles/ml). Keep this working stock for the whole experiment.
- 4.4 Prepare EVs in a known volume of DPBS and annotate the dilution, if any. Note that if coming from Basic Protocols 1 and 2, expect some EV loss during the protocol and dilute accordingly.
- 4.5 Vortex the working stock of beads for 5 seconds.
- 4.6 Spike the beads by diluting them 100-fold in the EV prep. The final concentration of reference beads, if following previous example, will be 1.8×10^6 particles/ml. Mix the sample.
- 4.7 Run the EV sample containing the spiked reference beads as described in Basic Protocol 4.

Note

200 nm polystyrene beads may overlap in scatter with some EVs. The authors therefore use 200 nm yellow-green beads if EVs are stained PE or APC, or 200 nm red beads if EVs are stained with CFSE or FITC. This allows differentiation of the EV and bead populations more easily and will provide a more reliable count.

4.8 Record the EV and bead event count as shown.

Note

The detectable concentration is dependent upon the limit of detection of the instrument and how the gating strategy was defined. In accordance with the MIFlowCyt-EV framework, parameters being used to gate EVs and the threshold of the instrument should be calibrated into standard units to allow the detectable concentration to be defined in standard units that can be validated (Welsh et al., 2020). See dx.doi.org/10.17504/protocols.io.bjcqkivw for information on how to calibrate fluorescence and light scatter parameters into standard units.

EV concentration of a solution can be calculated using beads as spike in references.a) nanoFACS plots depicting FSC and SSC signals in the 488 nm laser channel for control PBS, 200 nm polystyrene spike in beads and DC2.4 EVs + spike in beads. b) Determined concentration of DC2.4 EVs from nanoparticle tracking analysis and Astrios EV using calculated based on reference bead counts. Experiments were performed in triplicate (median $\pm$ 25th, 75th percentile plotted for NTA and Astrios EQ data).

4.9 Multiply the EV count by the bead concentration before dividing by the bead count, as shown in the following **equation**:

$$\text{Detectable EV Concentration} = \frac{(\text{EV Gate Count} \times \text{Bead concentration})}{(\text{Bead Gate Count})}$$

Note

*Note: If there is any noise or background source in the EV gate, these background counts must be subtracted to approximate EV counts. A clear example of this background would be any events that we get when running clean PBS. In that case use **equation**:*

Detectable\ EV\ Concentration = $\frac{(\text{EV\ Gate\ Count} - \text{Background\ Count}) \times \text{Bead\ concentration}}{(\text{Bead\ Gate\ Count})}$

Further information

- 5 Background information, critical parameters, and troubleshooting steps can be found within the manuscript.