VERSION 2

JUL 16, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bhqkj5uw

Protocol Citation: Neilier Junior 2020. Dialysis using D-Tubes. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bhqkj5uw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Jun 19, 2020

Last Modified: Jul 16, 2020

PROTOCOL integer ID:
 38380

Keywords: Dialysis, Protein Purification, Proteomics, Sample Preparation, Desalination, Purification

Dialysis using D-Tubes V.2

Neilier Junior¹

¹Universidade Federal de Viçosa

Neilier Junior

Universidade Federal de Viçosa, University of Manitoba

ABSTRACT

Sample preparation is crucial for successful biomolecule analysis. For this, the method must be individually chosen, based on the characteristics of the sample and the analyte. How some strategies can compromise the integrity of the molecule of interest. The dialysis D-tubes allow the concentration of samples with no precipitation by organic solvents or salts.

MATERIALS

D-Tubes Dialyzers

Milli-Q Water

1000 mL Beaker

Automatic micropipette

Floating rack

Plastic film

SAFETY WARNINGS

Wear personal protective equipment: gloves, lab coat and mask.

BEFORE START INSTRUCTIONS

Organize your workspace

Make sure all solutions and equipment are available. Plan the experiment!

Material preparation

- 1 Choose the cutting mass (3.5 to 14 kDa) of the dialysis D-Tube based on the biomolecule to be purified.

Procedure

- 2 Complete the dialysis D-Tube with Milli-Q water and let it equilibrate for 15 min
- 3 Remove the water and weigh the dialysis tube
- 4 Add the sample with the aid of a pipette and weigh the tube again
- 5 Prepare a beaker with 1000 mL of Milli-Q water (or exchange buffer) and a magnetic stir bar
- 6 Place the dialysis tube on a floating rack and immerse the tube in the water in the beaker
- 7 Cover the beaker with plastic film, start stirring and monitor the conductivity of the solution in the beaker. When the conductivity stops increasing, the concentrations between the sample and the solution are in equilibrium

If conductivity is not controlled: 1 h of dialysis suffices to desalinate 3 M of ammonium sulfate in these conditions
- 8 Remove the dialysis tubes from the beaker and weigh the contents