AUG 21, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bj4zkqx6

Protocol Citation: Angel A Justiz-Vaillant 2020. Direct ELISA for investigating the binding of chemically-made Protein-LAG to immunoglobulins..
protocols.io
<https://dx.doi.org/10.17504/protocols.io.bj4zkqx6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Aug 21, 2020

Last Modified: Aug 21, 2020

PROTOCOL integer ID:
 40825

Direct ELISA for investigating the binding of chemically-made Protein-LAG to immunoglobulins.

Angel A Justiz-Vaillant¹

¹University of the West Indies St. Augustine

University of the West Indies

angel.vaillant@sta.uwi.edu

Angel A Justiz-Vaillant

University of the West Indies St. Augustine

ABSTRACT

Protein LAG (SpLAG) is an immunoglobulin-binding protein that interacts with the Fc and Fab regions of many mammalian immunoglobulins. It is produced by a chemical coupling of individual proteins and then mixing it up to the appropriate protein ratio. SpLAG binds well to some avian immunoglobulin [1].

References

1. Vaillant AJ, McFarlane-Anderson N, Wisdom B, Mohammed W, Vuma S, et al. (2013) Immunoglobulin-binding Bacterial Proteins (IBP) Conjugates and their Reactivity with Immunoglobulin in Enzyme-Linked Immunosorbent Assays (ELISA). J Anal Bioanal Tech 4: 175. doi:10.4172/2155-9872.1000175

MATERIALS

MATERIALS

Horseradish peroxidase (HRP) Gold Biotechnology Catalog #P-100

Nunc™ 96-Well Polystyrene Round Bottom Microwell Plates, V 96 well plate, Non-Treated, clear, without lid, Sterile Thermo Fisher Catalog #260210

Staphylococcal Protein-A Sigma Aldrich

Protein-L from P. Magnus Contributed by users

Streptococcal protein G by Sigma Aldrich Contributed by users

- 1 This ELISA is used to study the interaction of protein-LAG (SpLAG) with diverse immunoglobulins.

- 2 The 96 well microtitre plate is coated overnight at 4°C with 1 µg/µl per well of purified immunoglobulins or 50 µl of any animal sera in carbonate-bicarbonate buffer pH 9.6.
- 3 Then plate is treated with bovine serum albumin solution and washed 4X with PBS-Tween.
- 4 Then 50 µl of peroxidase-labeled-protein-LAG conjugate diluted 1:5000 in PBS-non-fat milk is added to each well and incubated for 1.30h at RT. After that the plate is washed 4X with PBS-Tween.
- 5 Pipette 50 µl of 3,3',5,5' - tetramethylbenzidine (TMB; Sigma-Aldrich) to each well.
- 6 The reaction is stopped with 50 µl of 3M H₂SO₄ solution.
- 7 The plate is visually assessed for the development of colour and read in a microplate reader at 450 nm.
- 8 A cut-off point should be calculated as the mean of the optical density of negative controls x 2 SD.