JAN 01, 2020

DNA Quantification

In 2 collectionsAddgene The Nonprofit Plasmid Repository¹¹Addgene

Addgene The Nonprofit Plasmid Repository

Addgene

ABSTRACT

This protocol is for DNA quantification. To see the full abstract and additional resources, please visit <https://www.addgene.org/protocols/dna-quantification/>.

OPEN ACCESS**DOI:**dx.doi.org/10.17504/protocols.io.4g8gtzw**External link:**<https://www.addgene.org/protocols/dna-quantification/>**Protocol Citation:** Addgene

The Nonprofit Plasmid Repository 2020. DNA Quantification. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.4g8gtzw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: Jun 20, 2019**Last Modified:** Aug 17, 2020**PROTOCOL integer ID:**

24832

1 Before measuring any samples, be sure to 'blank' the spectrophotometer using the solution the

DNA is resuspended in, but with no DNA added. 'Blanking' measures the background inherent to the machine and your solvent.

- 2 If using a NanoDrop to measure your samples, place 1 µL 2 µL of mini-prepped DNA onto the pedestal.

- 3 Close the lid and click measure, be sure to record the concentration and purity.

Note

Note, purity is measured under the 260/280 column (A good purity ranges from 1.80-2.00).

- 4 Repeat for each sample.

Note

Notes:

- Keep in mind that despite the accuracy of the NanoDrop, if two consecutive samples have significantly different concentrations, it is possible that the difference between them has affected the accuracy of the NanoDrop. It is a good idea to re-zero any spectrophotometer between samples if they are expected to vary significantly in concentration.
- DNA dissolved in water is going to have a greater variance in concentration readings than a DNA sample dissolved in buffer (such as TE). You will get much more accurate and consistent readings from DNA in a buffered solution.