JAN 11, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.ww7ffhn

External link:
<http://iopscience.iop.org/article/10.1088/1741-2552/aad78e/pdf>

Protocol Citation: Kirill Aristovich, Matteo Donega, David Holder 2019. EIT data acquisition in rat sciatic nerve using stimulation of tibial and peroneal branches.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.ww7ffhn>

MANUSCRIPT CITATION:
 Kirill Aristovich et al 2018 J. Neural Eng. 15 056025

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development
 We are still developing and optimizing this protocol

Created: Jan 09, 2019

EIT data acquisition in rat sciatic nerve using stimulation of tibial and peroneal branches

In 1 collection

Kirill Aristovich¹, Matteo Donega², David Holder¹

¹University College London, University of London; ²Galvani Bioelectronics

SPARC

Tech. support email: info@neuinfo.org

Kirill Aristovich

ABSTRACT

High-level protocol to obtain the EIT imaging dataset and reconstruct images for the paper.

GUIDELINES

Please consult UCL EIT team before attempting the protocol, the complexity of the required hardware and software is such, that the collaboration will be much more effective. It is possible to create customised easy-to-operate systems and software packages which are tailored for the specific paradigm

SAFETY WARNINGS

Chemicals used during the electrode interface construction are toxic.

BEFORE START INSTRUCTIONS

consult UCL EIT team before attempting the protocol

PROTOCOL integer ID:
19135

Keywords: imaging, neural
traffic, EIT, electroseuticals,
cuff electrode

Hardware

1

Prepare nessesary hardware:

Equipment	
UCL EIT system	NAME
EIT system	TYPE
Custom	BRAND
n/a	SKU
https://github.com/EIT-team/ScouseTom	LINK
Number of channels: up to 256 Current frequency range: DC - 20kHz Sampling frequency: 100kHz Onboard stimulator	SPECIFICATIONS

Neural Interface

2

Prepare neural interface:
Follow the instructions in teh paper <http://iopscience.iop.org/article/10.1088/1741-2552/aae868>

Safety information

Some of the chemical components are toxic
Chemical hood cabinet is required

Expected result

Software

3

Software

UCL EIT system

NAME

UCL EIT group

DEVELOPER

<https://github.com/EIT-team/ScouseTom/tree/master/src>

SOURCE LINK

Animal Preparation

4

Procedures were performed on a vibration isolated table (Thorlabs Inc., USA) and, throughout experiments, the core body temperature of the rat was controlled with a homeothermic heating unit (Harvard Apparatus, Kent, UK) and maintained at 37°C

- 5 Sprague-Dawley adult male rats weighing 300 to 450g.
- 6 Anaesthesia is induced with a 5% isoflurane in 100% O₂.
- 7 ntubated using a small animal laryngoscope

Equipment	
Small animal laryngoscope	NAME
Laryngoscope	TYPE
Custom 3D printed	BRAND
n/a	SKU
https://www.thingiverse.com/thing:148315	LINK

- 8 18G cannula and mechanical ventilation is provided, using a Harvard Apparatus Inspira Ventilator (Harvard Apparatus, Ltd, UK), with a 50/50% gas mixture of O₂ and air.
- 9 Arterial and venous access was established through cannulation (BD Insite/Vialon, Becton, Dickinson U.K. Ltd.) of the right femoral vessels.
- 10 The arterial blood pressure should is monitored (Cardiacap 5, Datex Ohmeda) and the mean arterial pressure (MAP) kept between 90 and 110mmHg using labetalol and adrenaline as necessary.

- 11 Once intravenous access had been established, a constant infusion of propofol is initiated. Monitor the depth with pinch reflex.

Surgery

- 12 Access sciatic trunc, common peroneal nerve and tibial nerve following the guide:
<https://onlinelibrary.wiley.com/doi/full/10.1002/mus.21652>

Note

Access to the common peroneal nerve was established through a 2mm lateral incision in biceps femoris near the knee joint.

Electrode placement

- 13 Implant the stimulation electrodes on common peroneal and tibial branches (bipolar Ag hook electrodes)
- 14 Implant the EIT array on main sciatic trunc

Expected result

Experimental design and setup. (a) The EIT array was placed on the main sciatic nerve running in the posterior compartment of the thigh approximately 20 mm distal to the greater sciatic foramen. (b-c) Two stimulating electrode pairs were placed on the tibial and common peroneal nerves. EIT images were recorded during the repeated activity in the sciatic nerve evoked with 5Hz stimulation of each branch at a time.

Data acquisition

- 15 For all studies, a square biphasic (positive first) constant current temporal waveform was delivered using a balanced current source (Keithley, UK model no 6221) with:
 - 1-3 mA amplitude,
 - 50 μ s pulse width,
 - 5Hz frequency (200 ms inter-stimulus time)
- 16 Program the EIT system with the following parameters:

☐ EIT_parameters.mat
- 17 Connect stimulator to the tibial stimulation electrode. Collect the data
- 18 Connect stimulator to the common peroneal stimulation electrode. Collect the data

Data Processing and Image reconstruction

19 Load and process the data using the following software:

Software		
UCL LoadData		NAME
UCL EIT team		DEVELOPER
https://github.com/EIT-team/Load_data		SOURCE LINK

20 Forward solution

Software		
PEITS		NAME
UCL EIT team		DEVELOPER
https://github.com/EIT-team/PEITS		SOURCE LINK

21 Inverse solution and images generation

Software

Image Reconstruction

NAME

UCL EIT team

DEVELOPER

<https://github.com/EIT-team/Reconstruction>

SOURCE LINK