VERSION 2

DEC 11, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.v68e9hw](https://doi.org/10.17504/protocols.io.v68e9hw)

External link:
<https://doi.org/10.7717/peerj.8157>

Protocol Citation: Kiichi Hirota, Yoshiyuki Matsuo, Mikio Hayashi 2019.
 Electrophysiological studies.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.v68e9hw>

MANUSCRIPT CITATION:
 Kusunoki M, Hayashi M, Shoji T, Uba T, Tanaka H, Sumi C, Matsuo Y, Hirota K, Propofol inhibits stomatocytin-1-sensitive voltage-dependent K channels in pancreatic β -cells and enhances insulin secretion. PeerJ doi: [10.7717/peerj.8157](https://doi.org/10.7717/peerj.8157)

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Electrophysiological studies V.2

PeerJ

Kiichi Hirota¹, Yoshiyuki Matsuo¹, Mikio Hayashi¹

¹Kansai Medical University

Yoshiyuki Matsuo

BEFORE START INSTRUCTIONS

■ Extracellular bath solution

For patch-clamp experiments, extracellular solutions consisted of Krebs-Ringer bicarbonate HEPES (KRBH) buffer (140 mM NaCl, 3.6 mM KCl, 0.5 mM NaH₂PO₄, 0.5 mM MgSO₄, 1.5 mM CaCl₂, 2 mM NaHCO₃, 10 mM HEPES, and 0.1% bovine serum albumin) containing 2 mM glucose; pH was adjusted to 7.4 with NaOH.

■ Gramicidin D

In gramicidin-perforated patch experiments, gramicidin D (Sigma-Aldrich) was dissolved in DMSO at 20 mg/ml and then diluted to a final concentration of 0.1 mg/ml in a standard KCl-rich pipette solution containing 150 mM KCl and 10 mM HEPES; pH was adjusted to 7.4 with KOH.

■ Pipette solution

In whole-cell current recording, the standard K⁺ pipette solution used contained 60 mM potassium aspartate, 65 mM KCl, 1 mM KH₂PO₄, 5 mM ethylenediaminetetraacetic acid, 3 mM K₂ATP, and 5 mM HEPES; pH was adjusted to 7.4 with KOH.

■ Experimental condition

Experiments were conducted at 23–30 °C.

■ References

Hayashi M. et al. 2016, Pflugers Arch 468:1171-81. doi: 10.1007/s00424-016-1806-9
 Zhang Y. et al. 2016, Neurosci Lett 616:93-7. doi: 10.1016/j.neulet.2016.01.058
 MacDonald PE. et al. 2002, J Biol Chem 277:44938-45. doi: 10.1074/jbc.M205532200

Created: Dec 06, 2018

Last Modified: Dec 11, 2019

PROTOCOL integer ID:
18368

- 1 MIN6 cells were incubated in an extracellular bath solution containing 2 mM glucose for 30 min at 37 °C before patch-clamp experiments.
- 2 Membrane potential and whole-cell current were recorded using the EPC 800 patch-clamp amplifier (HEKA Elektronik Inc. Holliston, MA, USA).
 - The amplifier was driven by Clampex 9 (Axon Instruments, Union City, CA, USA) to allow the delivery of a voltage-step protocol with concomitant digitization of the current. The capacitance transient current was compensated by the amplifier.
 - Patch pipettes (G-1.5; Narishige, Tokyo, Japan) had a resistance of 5–7 M Ω when filled with the pipette solutions.
 - The membrane potential was corrected for the liquid junction potential at the tip of the patch pipette in the KRBH buffer and for that at the tip of the indifferent reference electrode filled with KRBH buffer and placed in the bath.
 - Whole-cell capacitance and series resistance (R_s) were 7.4 ± 1.5 pF and 13.3 ± 1.3 M Ω ($n = 6$), respectively.
 - The R_s was not electronically compensated during the experiments, and the potentials reported here have not been corrected for the R_s .
- 3 The whole-cell current was filtered at 1 kHz with an internal four-pole Bessel filter, sampled at 2 kHz, and transferred to digital signals through Digidata 1322A (Axon Instruments).
- 4 A subsequent current analysis was performed using Clampfit 9 (Axon Instruments).