AUG 13, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bjnrkmd6

Protocol Citation: Angel A Justiz-Vaillant 2020. ELISA for anti-SpA antibodies in hyper-immune egg whites..
protocols.io
<https://dx.doi.org/10.17504/protocols.io.bjnrkmd6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Aug 13, 2020

Last Modified: Aug 13, 2020

PROTOCOL integer ID:
 40369

ELISA for anti-SpA antibodies in hyper-immune egg whites.

Angel A Justiz-Vaillant¹

¹University of the West Indies St. Augustine

University of the West Indies

angel.vaillant@sta.uwi.edu

Angel A Justiz-Vaillant

University of the West Indies St. Augustine

MATERIALS

MATERIALS

96-Well Microtiter™ Microplates, Polystyrene, 330µL, Nonsterile, Flat-bottom **Thermo Fisher Catalog #9205**

3 3'5 5'-Tetramethylbenzidine Liquid Substrate Supersensitive for ELISA **Sigma Aldrich Catalog #T4444-100ML**

Staphylococcal Protein-A **Sigma Aldrich**

Anti-Protein-A **Contributed by users**

- 1 The 96 well polystyrene microplate (U-shaped bottom) is coated with 500 ng of SpA (Sigma-Aldrich) in coating buffer for 4 h at 37oC.

- 2 The microplate is washed four times with PBS-Tween-20 and blocked with 3% non-fat milk in PBS, 25 µl/well, 1h, RT.
- 3 The microplate is washed four times again.
- 4 Samples of egg whites are added in 1:10 dilutions.
- 5 After incubation for 1h at RT the microplate is washed four times.
- 6 Then, 50 µl of peroxidase-labelled protein-A diluted 1:3000 (Sigma-Aldrich) is added.
- 7 The microplate is incubated for 1h at RT.
- 8 The microplate is washed four times.
- 9 Tetramethylbenzidine (TMB) solution (50 µl) is added to each well.
- 10 Perform a further incubation of 15 min in the dark

- 11** The reaction is stopped with 3M H₂SO₄.
- 12** Read the microplate in a microplate reader at 450 nm.
- 13** Calculate the cut-off point as appropriate.