AUG 23, 2020

ELISA for quantification of human immunoglobulin G (IgG) in serum or plasma.

Angel A Justiz-Vaillant¹, Belkis Ferrer-Cosme²

¹University of the West Indies St. Augustine;

²"Saturnino Lora Torres' Provincial Teaching Clinical Surgical Hospital. Cuba

University of the West Indies

angel.vaillant@sta.uwi.edu

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bj7skrne

Protocol Citation: Angel A Justiz-Vaillant, Belkis Ferrer-Cosme 2020. ELISA for quantification of human immunoglobulin G (IgG) in serum or plasma..
protocols.io
<https://dx.doi.org/10.17504/protocols.io.bj7skrne>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Aug 23, 2020

Last Modified: Aug 23, 2020

PROTOCOL integer ID:
 40914

Angel A Justiz-Vaillant

University of the West Indies St. Augustine

ABSTRACT

IgG is a monomer with an approximate molecular weight of 146 Kd and a serum concentration of 9.0 mg/mL. It is synthesized mostly in the secondary immune response to pathogens. IgG can activate the classical pathway of the complement system, and it also is highly protective. The four subclasses of IgG include IgG1, IgG2, IgG3, and IgG4. IgG crosses the placenta, protecting the neonate from infectious diseases. [1]

- 1 An anti-human IgG coating antibody is adsorbed onto the microwells by incubation overnight at 4°C with carbonate-bicarbonate buffer.

- 2 Add 50 µl of human serum or plasma. Human IgG present in the serum or plasma binds to antibodies adsorbed into the microwells.
- 3 The microplate is blocked with 3% non-fat milk-PBS buffer and later wash to remove unbound proteins.
- 4 Fifty (50) µl of biotin-conjugated anti-IgG antibody is added. The optimal dilution must be investigated.
- 5 The microplate is rewashed with PBS-Tween 20 buffer, pH 7.4.
- 6 One hundred µl of streptavidin-HRP conjugate is added and it binds to the biotin-conjugated anti-IgG antibody. The optimal dilution of this conjugate must be investigated.
- 7 The plate is washed following incubation to remove the unbound Streptavidin-HRP.
- 8 Add 100 µl of 3,3',5,5'- tetramethylbenzidine (TMB; Sigma-Aldrich) into each well.
- 9 Incubate the microwells in the dark for 20 min.

- 10** A colored product is formed in proportion to the quantity of IgG present in the sample or standard.
- 11** The reaction is terminated by addition of 100 μ l 3M H₂SO₄ and the absorbance is measured at 450 nm.
- 12** A standard curve is made from 7 human IgG standard dilutions and the human IgG sample concentration is determined.
- 13** For better results place the microplate on a microplate shaker in every incubation.