AUG 28, 2020

Enzyme linked immunosorbent assay (ELISA) for determining the serum concentration of IL-33 in humans.

Angel A Justiz-Vaillant¹

¹University of the West Indies St. Augustine

University of the West Indies

angel.vaillant@sta.uwi.edu

Angel A Justiz-Vaillant

University of the West Indies St. Augustine

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bjy9kpz6

Protocol Citation: Angel A Justiz-Vaillant 2020. Enzyme linked immunosorbent assay (ELISA) for determining the serum concentration of IL-33 in humans.. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bjy9kpz6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Aug 19, 2020

Last Modified: Aug 28, 2020

PROTOCOL integer ID:
40705

- 1 Ninety-six well ELISA plates are coated with monoclonal anti-human antibodies to interleukin-33 (IL-33).

- 2 Patient serum samples are added to the plates.
- 3 The plate is incubate x 1.30 hour at RT.
- 4 The plate is washed 4 times with PBS-tween buffer.
- 5 The wells are incubated with a biotin conjugated anti-human IL-33 for 1.30 hour at RT.
- 6 The plate is washed again as above.
- 7 Add to the plate a peroxidase-labeled streptavidin conjugate and incubate it for 1 hour at RT.
- 8 After a further washing procedure a substrate solution reactive is added and allowed to produced a colored reaction in positive controls.
- 9 The level of IL-33 in the sample is proportional to the colored product developed.

- 10 The addition of 3M H₂SO₄ stops the reaction.
- 11 The absorbance is measured at 450 nm.
- 12 The IL-33 concentration can be calculated by generating an standard curve using lineal regression.