AUG 22, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bj54kq8w

Protocol Citation: Angel Justiz-Vaillant 2020. Enzyme-linked immunosorbent assay (ELISA) for studying the presence of anti-Salmonella antibody in layer hen's egg yolks.. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bj54kq8w>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Aug 22, 2020

Last Modified: Aug 22, 2020

PROTOCOL integer ID:
 40860

🌐 Enzyme-linked immunosorbent assay (ELISA) for studying the presence of anti-Salmonella antibody in layer hen's egg yolks.

Angel Justiz-Vaillant¹

¹University of The West Indies. St. Augustine. Trinidad and Tobago.

University of the West Indies

angel.vaillant@sta.uwi.edu

Angel A Justiz-Vaillant

University of the West Indies St. Augustine

ABSTRACT

Enzyme-linked immunosorbent assay (ELISA) for studying the presence of anti-*Salmonella* antibody in layer hens was a reproducible and feseable test used to meassure IgY development after vaccination.

GUIDELINES

After adding the substrate, keep the ELISA plate in a very dark place to get reproducible results.

MATERIALS

MATERIALS

⊗ Anti-Chicken IgY, HRP Conjugate, 300ul Promega Catalog #G1351

⊗ LPS Sigma-aldrich Catalog #L3129

⊗ eBioscience™ TMB Solution (1X) Thermo Fisher Catalog #00-4201-56

⊗ Nunc™ 96-Well Polystyrene Round Bottom Microwell Plates, U 96 well plate, Non Treated, clear, with lid, Sterile Thermo Fisher Catalog #268200

⊗ ELISA Coating Buffer (5X) BioLegend Catalog #421701

- 1 U-shaped bottom's ninety-six well polystyrene microplate purchased at Sigma-Aldrich, St. Louis USA was incubated with (2 µg/well) of the LPS (Sigma –Aldrich) from Salmonella Typhimurium

in coating buffer (overnight at 4 °C.)

- 2 The microtiter plates was washed four times, with 10 % PBS-Tween-20.
- 3 The microplate was blocked with 3% non-fat milk in PBS (25 µl/well).
- 4 The microplate was incubated 1 hr at RT .
- 5 The microplate was washed four times.
- 6 Then a 50 µl aliquot of the egg yolk (Ig)Y solutions in a concentration of 1.25 mg/ml was added in triplicate. The IgY concentration was assessed by ELISA and sample were titrated with sample buffer until it got the expected IgY concentration.
- 7 After incubating for one hour at RT, the microplate was washed four times.
- 8 Fifty (50 µl) of the anti-IgY-HRP conjugate (Sigma-Aldrich) diluted to 1:30000 with conjugate diluent was added into each well.
- 9 The microplate was incubated for 1 hr at RT.

- 10** Then, the microplate was washed four times.
- 11** Fifty (50 µl) of tetramethylbenzidine (TMB, Sigma-Aldrich) was added into each well.
- 12** The microplate was further incubated for 15 minutes in the dark.
- 13** Fifty (50 µl) 3M HCl was added to the microplate for stopping the reaction.
- 14** After that, reaction color development was measured with a microplate reader (Synergy™ Neo Hybrid Multi-Mode Microplate Reader).
- 15** The cut-off point was an OD of 0.51, and it was calculated from the XOD of the negative control times 3. This ELISA tested triplicates of a total of 90 IgY preparations.