JUL 09, 2021

DOI:
dx.doi.org/10.17504/protocols.io.bigdkbs6

Protocol Citation: Andreas Glenthoej, Jesper Petersen 2021. Eosin-5'-maleimide (EMA) binding test with fluorescent beads.

protocols.io
<https://dx.doi.org/10.17504/protocols.io.bigdkbs6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Jul 10, 2020

Last Modified: Jul 09, 2021

PROTOCOL integer ID:
 39141

Keywords: membranopathy, hereditary spherocytosis, EMA, eosin-5-maleimide, hemolysis, anemia

Eosin-5'-maleimide (EMA) binding test with fluorescent beads

Andreas Glenthoej¹, Jesper Petersen¹

¹Danish Center for Hemoglobinopathies

Danish Center for Hemoglobinopathies
 Tech. support email: andreas.glenthoej@regionh.dk

Andreas Glenthoej
 Danish Center for Hemoglobinopathies

DISCLAIMER

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to [protocols.io](#) is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with [protocols.io](#), can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

The flow cytometry-based eosin-5'-maleimide (EMA) binding test is used for reliable diagnostics of hereditary spherocytosis, a relatively common hereditary anemia. In this modified version of the EMA-binding test, we utilize commercially available rainbow beads to overcome the need for three to six healthy controls. However, two healthy controls are still used in this protocol as an extra safety measure.

GUIDELINES

The eosin-5'-maleimide (EMA) binding test was performed on blood within 48 hours of sampling. We recommend using EDTA-stabilized blood and store the samples at 4°C until analysis.

All EMA binding tests were performed on a FACS Canto II (BD Biosciences, Franklin Lakes, NJ, United States) using standard filter options (530/30). Other flow cytometers should do as well.

This protocol deviates from other EMA-binding tests by utilizing commercially available fluorescent beads instead of three to six healthy control samples.

As rainbow bead MFI is slightly higher than the average fluorescence level of healthy controls, one can utilize a constant calibration factor (CF) to make the results comparable to the traditional EMA-binding test.

The reason for this CF is solely to ensure that the EMA value obtained can be compared with the EMA values stated in the literature.

New lots of rainbow beads should be calibrated towards the previous to ensure consistency of the EMA values calculated. Otherwise, a new CF should be calculated.

We routinely use two healthy controls with each sample as an extra safety measure. We do not age-match these to the patient. If possible, a travel control should be analyzed along with the patient sample.

MATERIALS

MATERIALS

Sodium chloride Contributed by users

5-Maleimido-eosin for fluorescence $\geq 93\%$ (HPLC) Merck MilliporeSigma (Sigma Aldrich) Catalog #63184

Rainbow Fluorescent Particles 3.0-3.4 μm (mid-range FL1 fluorescence) BD Biosciences Catalog #556298

Dulbecco's Phosphate Buffered Saline Merck Catalog #D8537

FluoroSpheres Calibration beads Agilent Technologies Catalog #K011011-2

Bovine Serum Albumin Sigma Aldrich Catalog #A4503

SAFETY WARNINGS

Treat all human blood as potentially infectious.
Laboratory coats or appropriate gowns must be worn in the laboratory and fastened properly.
Gloves must be worn when handling human blood.
All open cuts and abrasions must be covered.
Use disposable equipment wherever possible.

BEFORE START INSTRUCTIONS

If normalization to normal controls is desired, the constant calibration factor (CF) should be calculated in advance.

Preparation of EMA Reagent batches

13m

1 Preparation of stock of EMA-reagent:

5-Maleimido-eosin for fluorescence $\geq 93\%$ (HPLC) Merck MilliporeSigma (Sigma-Aldrich) Catalog #63184

- 1.1 1. Dissolve 10 mg EMA in 20 mL PBS.
2. Mix well.

2m

- 1.2 1. Prepare 80 new 500 μ L PCR tubes by labeling them (MM.YY)
2. Transfer aliquots of 250 μ L EMA solution to each tube.

10m

- 1.3 Store in freezer.
Aliquots of frozen dissolved EMA dye should be stable for at least six months.
Aliquots should be discarded after use (i.e. *not* frozen again for later use).

1m

Prepare samples

2h 16m

- 2 Remove a tube of stock EMA (0.5 mg/mL) aliquot from the freezer to thaw in a dark place
@ Room temperature .

1m

3 Prepare samples

30m

3.1

1. Label a 14 ml tube **for each sample**
2. 10 mL 0.9% NaCl is added to each tube
3. 100 μ L of mixed blood is added to its corresponding tube
4. Mix/Vortex tubes
5. 1500 x g, Room temperature, 00:05:00
6. Discard supernatant

7m

3.2

Wash twice:

1. Add 10 mL 0.9% NaCl
2. Mix/Vortex
3. 1500 x g, Room temperature, 00:05:00
4. Discard supernatant

12m

3.3

Transfer 5 μ L of washed packed red cells to a new (labelled) 1.5 mL eppendorf tube

2m

3.4

- Add 25 μ L of EMA stock solution
- Incubate for 01:00:00 @ Room temperature in a dark place
- Mix every 00:15:00

1h

3.5

1. 1500 x g, Room temperature, 00:00:30
2. Carefully discard supernatant

2m

3.6

1. Add 500 μ L PBS
2. Mix/Vortex

2m

3.7

1. Prepare FACS tubes by adding labels

1m

2. Transfer 25 µL of labelled red cells to the new tube

3.8 Wash the labelled RBC's 3 times:

18m

1. Add 3 mL PBS with 0.5% BSA
2. Mix/Vortex
3. 1500 x g, Room temperature, 00:05:00
4. Discard supernatant

3.9 1. Add 500 µL PBS with 0.5% BSA

1m

2. Mix/Vortex

3.10 Samples are now ready to be run on the flow cytometer.

Run samples on the Flow Cytometer

22m

4 Prepare FACS-tubes for beads:

5m

1. Label 3 tubes for beads and add 500 µl PBS to each.
2. Add 1 drop of beads to each tube (Calibration beads, Blank beads, and rainbow beads).

Calibration beads and blank beads are part of a calibration kit from Agilent (#K011011-2)
Rainbow beads are purchased separately (BD Biosciences, #556298)

5 Setup flow cytometer (Canto II):

2m

1. Set it to LOG-mode for FSC, SSC, and FL1
2. Apply a threshold of 5000 for the FSC.

6 For each patient sample:

15m

1. Start running the Calibration beads and adjust the voltage of the PMT in such a way that the fluorescence FL1 is approximately 50,000 (or as close hereto as possible). Acquire 15,000 events.
2. Run Blank beads and acquire 15,000 events.
3. Run Rainbow beads and acquire 15,000 events.
4. Run Control samples and acquire 15,000 events.
5. Run Patient sample and acquire 15,000 events.

Analyze results

5m

7 The analysis of the results can be done differently.

Step 7 includes a Step case.

Beads + CF

Beads only

Controls only

5m