SEP 08, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bk2hkyb6

Protocol Citation: eesha.sharma.phd 2020. FLASH amp.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.bk2hkyb6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Sep 08, 2020

Last Modified: Sep 08, 2020

PROTOCOL integer ID:
 41769

Keywords: Xprize

FLASH amp

eesha.sharma.phd¹

¹Quantification by Design, Stanford University

XPRIZE Rapid Covid Testing

eesha.sharma.phd

DISCLAIMER

This method is currently used for research and development, not for diagnostic purposes.

ABSTRACT

FLASH amp is a method that allows for room temperature detection and amplification of a viral RNA sequence of interest submitted as part of the Xprize competition. The main components of the reaction are not included here and are given generic names such as enzyme X as they are currently not protected under IP. The steps outlined with tubes labeled with the generic names would allow any lab to reproduce the method.

Version 1 of this method outlines the protocol for a fluorescence-based readout that can be used with a qPCR machine or TECAN style plate reader.

Version 2 (to be released shortly) will be a colorimetric readout that can be assessed by eye.

The outlined protocol assumes the user performing this in a laboratory setting will setup reactions in a clean room and analyze results in a separate post-amplification room. In addition, since our amplification is rapid and at room temperature, care but be taken to add ligand to the master mix quickly and the reactions be sealed.

MATERIALS

STEP MATERIALS

Pyrophosphatase, Inorganic (E.coli) - 50 units **New England Biolabs Catalog #M0361L**

Corning® Low Volume 384-well Black Flat Bottom Polystyrene **Contributed by users Catalog #3821BC**

Sample collection

- 1 Sample should be collected 1 hour after any food is consumed.
10 minutes before collection rinse mouth well with water.
Collect passive drool or pooled saliva below tongue into collection tube.

Set-up

- 2 Resuspend freeze dried master mix 1 with 7.04 µL of water
- 3 Resuspend freeze dried master mix 2 in 1 µL water and add to above
- 4 Add 0.3 µL
Pyrophosphatase, Inorganic (E.coli) - 50 units **New England Biolabs Catalog #M0361L**
- 5 Add 0.5 µL
Undisclosed enzyme Z
- 6 Add 0.5 µL
Undisclosed Enzyme Y

7 Add 0.13 µL
Undisclosed Enzyme X

8 Put mix in

Corning® Low Volume 384-well Black Flat Bottom Polystyrene **Contributed by users Catalog #3821BC**

Add 0.5 µL of sample and seal plate.

Note

A large master mix of the above components can be mixed and aliquoted into wells.

Data collection

9 Place plate in TECAN plate reader.

Equipment

SPARK	NAME
Microwell plate reader	TYPE
TECAN	BRAND
SPARK	SKU
https://www.tecan.com/blog/spark-multimode-microplate-reader-for-high-performance-cell-based-fluorescence-assays	LINK

Set gain to 120, excitation 480 and emission 550.

A qPCR machine can also be used.

Collect data at T= 30 min.

Data can also be collected every 1s for 30 min to monitor kinetics of reaction in the particular plate reader being used if not TECAN SPARK reader.

Fluorescence values of >4000 are positive and <4000 are negative.

Expected result