SEP 02, 2020

Fluorescent mutants screening in 96 well plates - *Chlamydomonas reinhardtii*

Joao Vitor Molino¹

¹Ronin Institute

Joao Vitor Molino

Ronin Institute

DISCLAIMER

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.big9kbz6

Protocol Citation: Joao Vitor Molino 2020. Fluorescent mutants screening in 96 well plates - *Chlamydomonas reinhardtii*. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.big9kbz6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Jul 12, 2020

Last Modified: Sep 02, 2020

PROTOCOL integer ID:
39169

Keywords: 96 well plate, *Chlamydomonas reinhardtii*, growth, High-throughput screening (HTS)

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to [protocols.io](#) is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with [protocols.io](#), can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

This protocol is intended for screening experiments with algae cells. Nevertheless, with modifications, it can be used for other microorganisms. It is possible to use this protocol with deep-well plates, with modifications.

GUIDELINES

All steps described in this protocol are intended to be conducted in a research laboratory. Follow aseptic procedures.

BEFORE START INSTRUCTIONS

Separate all material needed for the protocol

Plate preparation

10m

- 1 For 96 well plates, the usual max volume is ~ $330\ \mu\text{L}$.
 1. Clean and disinfect a biological cabinet
 2. Place all materials inside. (e.g: *sterile tips, media, sterile 96 well plates with lids, pipettes*)
 3. Add $160\ \mu\text{L}$ TAP media per well (Or another media) on the choosen plate. (For *fluorescent experiments, black plates are recommeded. Clear bottom allows to simultaneously check absorbance*)

10m

Colony picking

15m

- 2
 1. Use sterile tooth picks to collect individual colonies.
 2. Place tooth pick into the well
 3. Proceed to the next colony
 4. After collecting the desired number of colonies, mix the tooth pick in the media by spining it
 5. Remove and discard the used tooth picks
 6. Visually inspect the presence of green material inside the wells.

15m

Sterile toothpicks

Colony picking with toothpick

Plate wrapping

2m

- 3
 1. Add the lid to the plate
 2. Wrap it with a porous tape (Microporous), 3 laps around the plate, taping together the lid and plate.
 - 3.

2m

Growth

5d

1. Add plates to a microplate shaker
2. Set the shaker to continuous mode, 900 rpm
3. Illuminate with 60-80 $\mu\text{mol photons/m}^2\text{s}$, at 25 °C
4. Grow the cells for 120:00:00 (5 days) (*Important to let cells grow enough to make the reads, 5-7 days have been tested*)

Reading

20m

- 5 1. Centrifuge the plates 2000 x g, 25°C, 00:03:00 to remove any condensation to the lid
2. Add the plates to the microplate shaker
3. Set the shaker to 00:10:00 , 900 RPM
4. Place plates in the Plate reader with the desired reading settings.
5. Analyse the results

20m