VERSION 1

APR 23, 2021

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bsevnbe6

External link:
<https://doi.org/10.7759/cureus.24763>

Protocol Citation: Tomohisa Yasuda, Toru Miwa 2021. Foulage test. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bsevnbe6>

MANUSCRIPT CITATION:
 T. Yasuda, N. Etoh, Y. Araki, J. Yamada, R. Ide, T. Ohkawara, T. Kunihiro, A Dynamic Equilibrium Examination on Stabilometry (Foulage test), Equilib. Res. 71 (2012) 61–70. doi:10.3757/jser.71.61.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

🌐 Foulage test V.1

Toru

Tomohisa Yasuda¹, Miwa^{2,3}

¹Yasuda ENT Clinic, Tokyo, Japan;

²Tazuke Kofukai Medical Research Institute, Kitano hospital;

³Kyoto University

Tazuke Kofukai Medical Research Institute, Kitano hospital

Toru Miwa

Tazuke Kofukai Medical Research Institute, Kitano hospital, ...

ABSTRACT

The Foulage test is a test in which the stepping is recorded with the eyes open and closed by the ground contact method for 60 seconds in accordance with the metronome BPM120 by counting the sway of the center of gravity. The enlargement of the area of the still-standing object on a stabilometer means unstable directly. However, the dynamic test must divide the unstable movement to the active. The stepping of Foulage test is regulated that 120bpm tempo, upright standing with both arms set on the body-side, foot closed, toes keep touching on the plate constantly, so object can change only height of heels to rise up alternatively. Foulage is a French word for "grape trampling" in which grapes are crushed barefoot in the process of winemaking. This is because the ground contact method is different from the conventional stepping method of a march that raises the sole of the foot. In a stable stepping of a healthy person, it is considered that the left and right folds and the front-back shift are constant to some extent, and 120 steps and 60 reciprocations are added to the same locus, and an inverted V-shaped locus with a certain width. Here, the inverted V-shaped locus is close to a rectangle bent at the center, and its long side can be considered as the average moving distance ($L / 120$) of one step. The outer peripheral area of the inverted V-shaped locus, that is, the area A of the bent rectangle can be estimated by multiplying the length of the long side and the short side (front-back width). We defined front-back width as FT value. Our results in normal subjects showed that the height of heels enlarges the total locus length. Pearson correlation analysis revealed that FT value of healthy object plateaued under stable condition ($400 < L < 1400$). Therefore, the test should be performed within the range of heel height where the footsteps are stable in healthy subjects. In addition, we proposed the new parameters; the variance of steps is more direct parameter of stability. The plotting figure of steps and the time-graph can show the movement of a subject visually.

Created: Feb 15, 2021

ATTACHMENTS

[20200812miwa.jpg](#)

Last Modified: Apr 23, 2021

PROTOCOL integer ID:
47285

GUIDELINES

Not applicant.

Keywords: Foulage test.,
Stepping test, Stabilometer,
Dynamic equilibrium test

MATERIALS

Stabilometer (GS-31, Anima, Tokyo, Japan), Foulage test software (Anima, Tokyo, Japan)

SAFETY WARNINGS

Falling from the platform should be prevented by the careful observation.

BEFORE START INSTRUCTIONS

Patients should be told to perform foot closed, lifting up only heels with 400 - 1400cm of heel height alternately at the tempo of BPM 120 with the metronome.

- 1 A step on the center of stabilometer (GS-31, Anima, Tokyo, Japna) with foot closed, lifting up only heels alternately at the tempo of BPM (beat per minute) 120 with the metronome, during 60 seconds with eyes open.
- 2 A step on the center of stabilometer (GS-31, Anima, Tokyo, Japna) with foot closed, lifting up only heels alternately at the tempo of BPM 120 with the metronome, during 60 seconds with eyes closed.