OCT 15, 2019

GFP-RFP plate reader assay

Sebastiaan Kuiper¹

¹Wageningen University

iGEM Wageningen 2019

Sebastiaan Kuiper
Wageningen University

ABSTRACT

Fluorescence measurements of Green fluorescent protein (GFP) and/or Red fluorescent protein (RFP) in *Escherichia coli*.

MATERIALS

MATERIALS

96-well plate, flat bottom, tissue culture treated, black wall with clear bottom **Fisher Scientific Catalog #3904**

Microplate Reader Synergy Mx **Contributed by users**

50 ml Falcon tube **Contributed by users**

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.784hryw

Protocol Citation: Sebastiaan Kuiper 2019. GFP-RFP plate reader assay. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.784hryw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Oct 15, 2019

Last Modified: Oct 15, 2019

PROTOCOL integer ID:
28668

Keywords: Green fluorescent protein, GFP, Red fluorescent protein, RFP, Plate reader, dCas, dCpf1, Inhibition, *Escherichia coli*, *E. coli*

media preparations:

1 Prepare the following media and autoclave them according to standard procedures.

M9TG Media:

Reagent	Amount to add for 1L	Final concentration (1×)
M9 salts	1x	1x
Tryptone	10 g	10 g/L
Glycerol	5 g	5 g/L

PBS buffer:

Reagent	Amount to add for 1L	Final concentration (1×)
NaCl	8 g	137 mM
KCl	0.2 g	2.7 mM
Na ₂ HPO ₄	1.44 g	10 mM
KH ₂ PO ₄	0.24 g	1.8 mM

Plate reader protocol:

2 Set the following variables for the BioTek Synergy MX Microplate Reader:

- Set Temperature: 37°C

Preheat before moving to next step

- Shake: Medium, 0:30 (MM:SS)

- Read OD600:
Absorbance Endpoint of full plate
Wavelengths:600
Read Speed: Normal,Delay: 100 msec, Measurements/Data Point: 8
- Read GFP:
Fluorescence Endpoint of full plate
 - Filter Set 1:
Excitation: 485/9,0,Emission: 512/9,0
Optics: Bottom,Gain: 50
 - Filter Set 2:
Excitation: 485/9,0,Emission: 512/9,0
Optics: Bottom,Gain: 75
 - Filter Set 3:
Excitation: 485/9,0,Emission: 512/9,0
Optics: Bottom,Gain: 100
 Read Speed: Normal, Delay: 100 msec, Measurements/Data Point: 10
- Read RFP:
Fluorescence Endpoint of full plate
 - Filter Set 1:
Excitation: 555/9,0,Emission: 585/9,0
Optics: Bottom,Gain: 100
 - Filter Set 2:
Excitation: 555/9,0,Emission: 585/9,0
Optics: Bottom,Gain: 150
 Read Speed: Normal, Delay: 100 msec, Measurements/Data Point: 10

Fluorescence analysis:

- 3 Grow desired bacteria overnight in a 50 ml falcon tub, containing 10 mL M9TG media and grown cultures @ 37 °C , 250 RPM overnight.

Note

Luria-Bertani media is also possible to use, but cells must be washed with PBS buffer prior to plate reader experiment because of the high amount of fluorescence from the yeast extract:

After overnight incubation, spin cells down at 4700 x G, for 5 min, discard the supernatant and resuspend in 10 ml PBS.

Repeat prior steps and resuspension again in PBS.

- 4 Load 160 µL PBS buffer in the wells of the 96-well plate (with clear bottom and black sides) and add 40 µL of overnight culture

Note

When cell have been grown overnight in LB media, washing steps are require before continuing! (see step 3)

Note

Add as s control a sample with 160 µL PBS buffer and 40 µL of non-fluorescent bacteria .

- 5 Start plate reader protocol go to step #2

Analyses

- 6 Retrieve the data from the computer, correct for the measurements for the OD600 and subtract the autofluorescence from the control sample of all the other samples.