APR 24, 2021

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bscpnavn

External link:
<https://doi.org/10.1371/journal.pone.0250516>

Document Citation: Anonymous 2021. High throughput Zika virus reporter virus particle microneutralization assay: Laboratory Procedure. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bscpnavn>

MANUSCRIPT CITATION:
 A high throughput reporter virus particle microneutralization assay for quantitation of Zika virus neutralizing antibodies in multiple species. Bohning et, al.

License: This is an open access document distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Created: Feb 11, 2021

Last Modified: Apr 24, 2021

High throughput Zika virus reporter virus particle microneutralization assay: Laboratory Procedure

Anonymous¹

¹Takeda Vaccines, Inc., Cambridge, MA, USA

Anonymous

ABSTRACT

Zika virus is a Flavivirus, transmitted via *Aedes* mosquitoes, that causes a range of symptoms including Zika congenital syndrome. Zika has posed a challenging situation for health, public and economic sectors of affected countries. To quantitate Zika virus neutralizing antibody titers in serum samples, we developed a high throughput plate-based Zika virus reporter virus particle (RVP) assay that uses an infective, non-replicating particle encoding Zika virus surface proteins and capsid (CprME) and a reporter gene (*Renilla* luciferase). This is the first characterization of a Zika virus RVP assay in 384-well format using a Dengue replicon *Renilla* reporter construct.

Serially diluted test sera were incubated with RVPs, followed by incubation with Vero cells. RVPs that have not been neutralized by antibodies in the test sera entered the cells and expressed *Renilla* luciferase. Quantitative measurements of neutralizing activity were determined using a plate-based assay and commercially available substrate. The principle of limiting the infection to a single round increases the precision of the assay measurements.

RVP EC₅₀ titers correlated closely with titers determined using a plaque reduction neutralization test (PRNT) ($R^2 > 95\%$). The plate-based Zika virus RVP assay also demonstrated high levels of precision, reproducibility and throughput. The assay employs identical reagents for human, rhesus macaque and mouse serum matrices. Spiking studies indicated that the assay performs equally well in different species, producing comparable titers irrespective of the serum species. The assay is conducted in 384-well plates and can be automated to simultaneously achieve high throughput and high reproducibility.

ATTACHMENTS

[Lab Protocol for Zika RVP Microneutralization Assay Bohning et al 2021.docx](#)

DOCUMENT integer ID:
47215

Keywords: Zika, Reporter
Virus Particle,
microneutralization, high
throughput, flavivirus, PRNT,
RVP

ATTACHMENTS

Lab Protocol
for Zika RVP
Microneutraliz
ation Assay
Bohning et al
2021.docx

Title: High throughput Zika virus reporter virus particle microneutralization assay: Laboratory Procedure.

Authors: Kelly Bohning 1, Stephanie Sonnberg1, #a, Hui-Ling Chen 1, #b, Melissa Zahralban-Steele 1, Timothy Powell 1, Greg Hather1, #c, Hetal K. Patel 1*, and Hansi J. Dean 1.

1 Takeda Vaccines, Inc., Cambridge, MA, USA

#a current address: Landkreis Giessen, Postfach 11 07 60, Giessen, Germany.

#b current address: PTC Therapeutics Inc, Bridgewater Township, NJ, USA

#c current address: Takeda California Inc., San Diego, CA, USA

INTRODUCTION

Neutralization of viral infectivity is one of the methods for determining the presence of functional antibodies in serum samples. Reporter virus particles (RVP) were purchased from Integral Molecular (Philadelphia, PA) Zika reporter virus particles (Z-RVPs) contain CprM/E from ZIKV strain SPH2015 packaged within the particle a Dengue-based replicon with the Renilla luciferase gene. The Z-RVPs express a Renilla luciferase reporter gene upon infection of permissive cells. The half maximal effective concentration EC50 titer of neutralizing antibodies is determined using a bioluminescent reaction which generates a glow-type luminescent signal by the interaction of the Renilla luciferase and coelenterazine substrate. The luminescent signal is measured using a luminescence enabled plate reader. Reduction in luminescent signal in the presence of serum indicates neutralization. This laboratory procedure (LP) reflects the experimental set up for an experiment with maximal six 384-well plates. The volumes listed throughout this LP can be changed based on the number of samples in the run.

Equipment, Materials and reagents

Table 1 Equipment

Equipment Description	Manufacturer/Vendor *
Sorvall Legend XTR Centrifuge	Thermo Scientific

Freezer ≤-65	Thermo Scientific
Freezer -20°C + 10°C	Thermo Scientific
Refrigerator 5°C + 4°C	Thermo Scientific
Water bath, 37°C + 2°C and 56+2°C	Fisher Scientific
Biological safety cabinet, Class II, certified	Baker
Cell Culture Incubator 37+2°C, 5+1% CO ₂	Thermo Scientific
Single channel pipettes, variable volumes, calibrated (e.g. P2, P20, P100, P200, P1000)	Thermo Scientific
8 or 12-channel pipettes, variable volumes, calibrated	Thermo Scientific
E1-Clip-tip electronic 8 channel pipettes, variable volumes, calibrated	Thermo Scientific
E1 - Clip-tip Equalizer 384 pipette 12 or 8 channel 30 µl volume, calibrated	Thermo Scientific
E1-Clip-tip Equalizer 384 pipette 12 or 8 channel 125µL volume, calibrated	Thermo Scientific
Pipet-Aid	Drummond
Countess II	Life Technologies
Enspire Reader	Perkin Elmer
Microcentrifuge	Eppendorf

**Equivalent manufacturers/vendors may be used.*

Table 2. Materials

Material Description	Manufacturer/Vendor*	Product Number*
Clip-tip 384 filter sterile tips (30 and 125µL volumes)	Thermo Scientific	94420103/94420153

Sterile 96-well polypropylene plates(untreated) U-bottom or V-bottom PCR plate	Greiner Bio-One/Thermo Scientific	650101/AB-1400
Adhesive foil PCR foil sealing sheets	Thermo Scientific	AB-0626
Clear plate seals	EXCEL Scientific	100SEALPLT
Sterile 15mL/50mL conical tubes	Falcon	352097/352070
125mL Sterile bottles	Nalgene	2019-0125
Tissue-culture treated 384-well plates white opaque bottom	Corning	3570
Sterile media reservoirs (25ml divided , 25mL or 50mL)	Vistalab Technologies	3054-1005/3054-1003/3054-1006
Sterile serological pipettes, assorted volumes (e.g. 5, 10, 25, and 50mL volumes)	Fisherbrand	13-678-11D/13-678-11E/13-678-11/13-678-11F
Tissue-Culture treated T175cm ² cell culture flask	Corning	10-126-61
Clip tip sterile filter pipet tips, assorted volumes (e.g. 20, 50, 200, 300, and 1000µL)	Thermo Scientific	94420213/94420253/ 94420313/94420513/ 94420713

*Equivalent manufacturers/vendors and/or product numbers may be used.

Table 3. Reagents

Material Description	Manufacturer/Vendor*	Product Number*
1x Opti-MEM reduced serum media	Gibco	11058-021
1X DMEM	Corning	15-017-CV
0.05% Trypsin	Gibco	25300-054

1X DPBS	Gibco	14040141
L-Glutamine(200mM)	Gibco	25030-081
100x Penicillin (10,000 IU/mL)/ Streptomycin (10,000 µg/mL)	Gibco	15140-122
Fetal Bovine Serum, Heat Inactivated	Sigma Life Science	F4135-100ML
Negative control serum	Innovative Research	IGRS-SER
Positive control serum	Takeda	Anti-Dakar Rabbit serum
Wet Ice	N/A	N/A
Zika RVP (Dengue backbone)	Integral Molecular	variable
<i>Renilla</i> -Glo Luciferase Assay System	Promega	E2750

*Equivalent manufacturers/vendors and/or product numbers may be used.

Preparation of Media and Reagents

Assay Media

1. Prepare assay media (1X Opti-MEM): In a biosafety cabinet, aseptically add the following ingredients to the 500mL bottle of 1X Opti-MEM.
2. Remove 55 mL of 1X Opti-MEM.
3. Add 50 mLs of Heat Inactivated (HI) FBS.
4. Add 5 mLs of 100X Penicillin/Streptomycin.
5. Mix by inverting several times.
6. Label with ELN reference number, initials, preparation date, and expiration date (1 month from preparation date).
7. Mix by inverting several times, cover bottle in foil before storing at 4±5°C.

Vero BREL Growth Media

1. Prepare cell growth media for cells (1X DMEM): In a biosafety cabinet, aseptically add the following ingredients to the 1X DMEM.
2. Remove 55 mL of 1X DMEM
3. Add 50 mLs of Heat Inactivated FBS
4. Add 5 mLs of L-Glutamine
5. Label with initials, preparation date, and expiration date (1 month from preparation date).
6. Mix by inverting several times, cover bottle in foil before storing at 4±5°C.

Note: Serum serial dilutions and Transfer Dilution Series to daughter plate can be performed manually or by the Tecan Fluent Automated system. Additional information is available upon request.

Prepare Serum

1. Thaw samples on wet ice or in refrigerator and spin down using microcentrifuge (quick spin). Label 0.5-1.5 mL microcentrifuge tubes with appropriate sample IDs, and positive/negative controls.
2. Aliquot amount needed to heat inactivation (HI). Typically, 20 μ L per sample and positive control, and 120 μ L for negative control. (The control amounts and all the steps listed in this procedure are for a 6-plate assay; Sample number and assay volumes will vary per experiment).
3. HI aliquots in 56 $\pm$ 2°C water bath for 30 $\pm$ 2 minutes. Serum samples prior to HI may go through 4 freeze-thaw cycles.
4. Spin down HI serum using microcentrifuge (quick spin).
5. Store HI serum on wet ice.
6. Using excel document RVP Neut calc-Dilution Plate set up tab see Appendix 2.1 for reference. Obtain a 96-well dilution plate. Label dilution plates according to sample ID's and positive and negative controls.
7. For the positive control prepare a 1:10 dilution by adding 10 μ L of HI serum and 90 μ L of assay media in a 0.5-1.5 mL microcentrifuge tube. Vortex and spin down using microcentrifuge (quick spin).
8. Then prepare a 1:20 dilution in appropriate dilution plate well by adding 95 μ L of assay media and 5 μ L of the HI 1:10 diluted serum to column 1 of the 96-well dilution plate (final dilution of 1:200).
9. For samples 1-72 and negative control, prepare 1:8 dilution by adding 56 μ L of assay media and 8 μ L of HI serum in column 1 of the 96-well dilution plate.
10. Add 36 μ L of assay media to columns 2-11 of the 96-well dilution plates. Refer to Figure 1 and Appendix 2.1 for dilution plate set up.
11. Perform 3-fold serial dilutions. Serial dilutions can be performed using a manual pipette or an electronic pipette:
 - **Manual pipette dilutions:** Using a manual 8 channel pipette perform 3-fold serial dilutions. Pipet mix column 1, 5 times. Transfer 18 μ L from column 1 to column 2, repeat across the plate to column 11. Pipette mix 5 times after each dilution. After dilutions are complete, discard last 18 μ L from column 11.
 - **Electronic pipette dilutions:** Using 8 channel electronic pipette perform 3-fold serial dilutions by the program below: Program E1 - Clip-tip Equalizer 96-well pipette 8 channel, 125 μ L volume: Mix volume 18 μ L; 0 μ L volume overfill; Mix cycle 5 times; Speed 10; Fill 18.0 μ L, speed 10; Dispense 18 μ L; 1 time; speed 10. Change tips after each mixing cycle. After dilutions are complete, discard last 18 μ L from column 11.
12. Cover 96-well dilution plate with adhesive seal. Spin dilution plates down at 300rpm for 1 minute.

Figure 1. Dilution Plate set-up

	1	2	3	4	5	6	7	8	9	10	11	Sample
A	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	1
B	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	2
C	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	3
D	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	4
E	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	5
F	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	6
G	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	7
H	56 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	36 μ L assay media	8

Transfer Dilution Series to Assay plate

1. Use the excel document RVP Neut Calc sheet -Assay Plate tab for the layout of the 384-well assay plate. See Figure 2 for 384-well assay plate layout and Appendix 2.2.
2. Obtain a white 384-well tissue culture plate.
3. To transfer the serially diluted serum in the 96-well dilution plate to the 384-well daughter plate:
4. Program E1 - Clip-tip Equalizer 384 pipette 12 channel, 30 µl volume:
Aspirate of 22.5 µl volume; 2.5 µl volume overfill; aspirate speed 3
Multi-dispense of 7.5 µl volume; dispense 3 times; speed 3.
Purge speed 10.
5. In the 384-well daughter plate the positive control and samples 1-6, 13-18, 25-30, 37-42, 49-54 and 61-66, column 12 will contain the most concentrated dilution; column 2 will be the least concentrated dilution. This can be done by turning the 96-well dilution plate upside down (A1 corner is your bottom right) prior to adding the samples to the 384-well daughter plate. (Note sample numbers will vary according to number of samples, adjust as needed).
6. In the 384-well daughter plate the negative control and samples 7-12,19-24,31-36,43-48,55-60,67-72, column 13 will be the most concentrated dilution; column 23 will contain the least concentrated dilution. Position the 96-well dilution plate in the upright position (with the A1 corner in the upper left-hand corner) prior to adding the samples to the 384-well daughter plate. Refer to Figure 2 and Appendix 2.2 for example of the 384- well plate set up. (Note sample numbers will vary according to number of samples, adjust as needed).
7. Transfer serum dilution series from 96-well plate to 384-well daughter plate.
8. Aspirate the serum serial dilution from the 96-well plate. Dispense first 7.5 µL back into dilution plate. Multi-dispense 7.5 µL per sample into the 384-well daughter plate. Purge into waste and discard tips between samples. Repeat for all samples.
9. Spin 384-well daughter plates down at 300rpm for 2 minutes
10. Proceed to **RVP Addition and Dilution** or to **Assay Plate Seal & Storage**.

Figure 2. 384-well Daughter Plate set up

Daughter Plate set-up: 384 well assay plate																									
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
A																									
B		Positive Control											Sample 7												
C																									
D		Sample 1											Sample 8												
E																									
F		Sample 2											Sample 9												
G																									
H		Sample 3											Sample 10												
I																									
J		Sample 4											Sample 11												
K																									
L		Sample 5											Sample 12												
M																									
N		Sample 6											Negative Control												
O																									
P																									

Assay Plate Seal & Storage

1. Seal the 384-well daughter plates with adhesive foil seal in biosafety cabinet.
2. Label plates with initials and date. Store assay plates at -20°C for no more than 4 days.
3. Proceed to RVP dilution and addition.

RVP Dilution and Addition

1. Use the excel document RVP Neut Calc- RVP tab to calculate RVP dilution (See Table 4 for reference and Appendix

1.1 for table formulas). Enter gray areas only. In OPTIMIZED RVP $\mu\text{L}/\text{WELL}$ column add 2.5 μL for RVP lot 229A. The RVP input amount should be assessed per RVP lot. In #WELLS/PLATES add the number of wells being utilized in one plate. In #PLATES enter 6 (total number plates ran per assay, will vary). In VOLUME DILUTED RVP/WELL (μL) enter 7.5. Volume Diluted RVP required (μL) will populate. Enter in the value from VOLUME DILUTED RVP REQUIRED (μL) to VOLUME PREP. VOLUME RVP (μL) and VOLUME ASSAY MEDIA (μL) will populate (Table 4).

Table 4. RVP Dilution Set up¹

RVP#	Lot#	dilution	Optimized RVP $\mu\text{L}/\text{well}$	# wells/plates	# plates	Total # wells	Volume diluted RVP/well (μL)	Volume diluted RVP required (μL)	Volume Prep(μL)	Volume RVP(μL)	Volume Assay Media(μL)
SHP2015 Zika RVP	229A	12.0	2.5	322	6	1932	7.5	18113	18113	6038	12075

¹See Appendix 1.1 to calculate formulas

1. If the 384-well daughter plates were stored in -20°C , equilibrate plates for 10-20 minutes at room temperature in the BSC.
2. Thaw 384-well daughter plates in a single layer. Do not stack plates.
3. Spin 384-well daughter plates down at 300rpm for 2 minutes.
4. In BSC, carefully remove foil from top of plate, then proceed to **RVP dilution**.

RVP Dilution

1. Collect required number of vials of RVP from VOLUME RVP (μL) from Table 4.
2. Quick thaw required number of RVP tubes in $37\pm 2^{\circ}\text{C}$ water bath for 2-3 minutes (5mL vials of RVP will take a bit longer to thaw). Immediately store on wet ice after thawing.
3. Pool all RVP single aliquot tubes into one size appropriate conical tube, mix gently by inversion 2-3 times.
4. Add ASSAY MEDIA VOLUME μL from Table 4 volume assay media in new sterile conical tube. Add RVP VOLUME μL from Table 1 volume pooled RVP to conical tube containing assay media.
5. Mix gently by inverting 3-4 times, store on wet ice. Use immediately.

RVP Addition

1. Transfer diluted RVP to sterile media reservoir, keep reservoir on wet ice.
2. Program E1-Clip-tip Equalizer 384 pipette 12 channel 30 μL volume:
 - Add 11 pipette tips to the pipettor.
 - Aspirate of 22.5 μL volume; 2.5 μL volume overfill; aspirate speed 3;
 - multi-dispense 7.5 μL dispense 3 times; speed 3;
 - Purge speed 10.
1. Aspirate diluted RVP, multi-dispense the first 7.5 μL in reservoir, then multi-dispense 7.5 μL to the appropriate sample in the 384 well plate. Purge into waste and discard tips between samples. Continue with the rest of the plate adding the RVP one sample at a time
2. Spin down the 384-well daughter plate at 300rpm for 2 minutes.
3. Immediately transfer plates to humidified $37\pm 2^{\circ}\text{C}$, 5% CO_2 incubator Do not stack plates. Orient plates long edge parallel to door.
4. Incubate 384-well daughter plates for 60 ± 2 minutes. Record start and stop time along with temperature in and out of the incubator.

5. Immediately proceed to cell culture harvest, cells will be added once the 60 ± 2 incubation time is complete.

Assay Incubation

Cell Harvest

1. Place Vero BREL growth media and 0.05% Trypsin in $37 \pm 1^\circ\text{C}$ water bath 10 minutes prior to splitting cells.
2. Obtain T-175 cell culture flasks from clean cell culture incubator.
3. Record cell passage number, harvest date and date last passaged in ELN
4. Wash flasks with 10 mL 1X DPBS twice.
5. Add 5 mL of 0.05% Trypsin to cell culture flask, place in $(37 \pm 2^\circ\text{C}$, 5% CO_2 incubator) for 5 minutes. Check to be sure cells have detached from flask under microscope.
6. Add 5 mL of Vero BREL Growth media to trypsinized cells.
7. Transfer cells to 50 mL conical tube, centrifuge for 5 minutes, $130 \times g$, at room temperature.
8. Aspirate supernatant from pelleted cells. Resuspend cells in 10 mL of assay media, per flask split.

Cell Count

1. Add 20 μL of cells to 20 μL of trypan blue to cell countess. Record the lot number of Trypan blue.
2. Add 10 μL of each of mixture to both sides of a hemocytometer. Record the lot number of Cell countess chamber.
3. Count cell using Countess II
4. Record viability and cell count in RVP Neut Calc sheet -cells tab (Table 5 and Appendix 1.2 for formulas).

Cell Dilution

1. Use the excel document RVP Reagent Calc - Cells tab to calculate number of cells needed for the assay. Enter values in grayed areas only. Non-grayed areas will automatically populate. Refer to Table 5 for reference only.
2. Enter 4625 in cells well column.
3. Enter correct number of wells/plate used (example in this portion is 322 wells per 1-384 well plate).
4. Enter in number of plates used in # plates column.
5. Total wells, Volume required will automatically populate.
6. Round volume up if needed and enter value in Volume Prepared.
7. Cells required will automatically populate.
8. Average number of cells (mL) will automatically calculate once the two cell counts (cells/mL live) are recorded in the worksheet. The Assay media volume (mL) will also auto populate.
9. Keep concentrated cell stock on ice until 5 minutes are left of neutralization incubation.

Table 5. Cell Dilution Tab¹

Vero Brel	N/A	N/A
Optimized Cells well	4625	N/A
cells / ml	306333	N/A
# wells/plate	122	N/A
# plates	6	N/A
Total wells	1932	N/A
Volume req.	38	mls
Volume Prepared	40	mls
Cells Req.	1233333	cells

Harvest Date	
Passage Number	
Date of Passage	
Flask Size	1175
Flask Media Volume	40

Set	Cells/ml total	Cells/ ml Bue	% live
count 1			
count 2			
Average (including additional dilution factor*)	NA	#DIV/0!	#DIV/0!

Amount of cells (mL)		Assay Media volume(mL)	
#DIV/0!	mL	#DIV/0!	mL

¹See Appendix 1.2 for formulas to calculate

Cell Addition

1. Add the amount of cells (mL) and assay media (mL) to the conical tube.
2. Resuspend cells by gently inverting 2-3 times.
3. Transfer diluted cells to sterile media reservoir in batches. Swirl conical tube before adding cells to the reservoir.
4. Program E1-Clip Tip Equalizer 384 pipette 12 channel:
 - Aspirate of 105 μ L volume; 10 μ L volume overfill; speed 10
 - Multi-dispense 15 μ L volume; dispense 7 times; speed 3
 - Purge speed 10.
1. Aspirate cell dilution, dispense first 15 μ L in reservoir, multi-dispense to the next 6 wells in to the 384-well daughter plates. Purge in the waste. Discard between aspiration cycles.
2. Repeat step 5 until all wells containing samples have cells added.
3. Program E1-Clip Tip Equalizer 384 pipette 12 channel 125 μ L volume:
 - Aspirate of 120 μ L volume; 12 μ L volume overfill; speed 10;
 - Multi-dispense 30 μ L volume; dispense 4 times; speed 10;
 - Purge speed 10.
1. Add 30 μ L assay media to the empty wells. Add 30 μ L of assay media to row A 1-24; P 1-24 and columns B1-B0 and B24-O24 and column 1 B-O and column 24 B-O.
2. Spin plates down at plates 300rpm for 2 minutes.
3. Immediately transfer to humidified 37 $\pm$ 2 $^{\circ}$ C, 5% CO₂ incubator. Do not stack plates. Be sure that the plates are positioned in a consistent orientation, long side parallel to back of the incubator.
4. Incubate plates for 72+ 2hrs. Record incubation times.

Renilla-Glo Detection

Storage conditions for Renilla-Glo luciferase assay buffer: -20°C. 4 °C for up to 1 year; room temperature, 22°C + 4°C for up to 3 months. Prepare Renilla-Glo substrate-buffer immediately before use, reagent will lose 10% activity after 2 hours at room temperature.

1. Allow *Renilla*-Glo buffer to equilibrate to room temperature. Do not thaw above 25°C. *Renilla*-Glo buffer can be stored at room temperature for up to 3 months once removed from -20°C. Record the date of thawing and date of expiration on the buffer bottle, ELN, and worksheet.
2. Record time and date when plates are taken out of incubator.
3. In biological safety cabinet, allow plates to equilibrate to room temperature for 20±5 minutes. No stacking of plates, spread them in one layer.
4. Use the excel document RVP Reagent Calc-*Renilla*-Glo Reagent tab to calculate *Renilla*-Glo buffer and substrate amounts. Enter in grayed areas only in Table 6. Enter in number of wells in #wells, and number of plates in assay in #plates. Buffer (mL), and Substate (mL) will auto populate (6 plates were used for example).
5. Quickly obtain *Renilla*-Glo substrate from -20°C freezer. In a 125 mL sterile bottle combine Buffer (mL) and Substrate (mL). Invert 3-4 times to mix.

Table 6. *Renilla*-Glo Buffer Preparation¹

Renilla Glo Buffer	
#wells	322
#plates	6
Buffer (mL)	72.5
Substrate(mL)	0.725

¹See Appendix 1.3 to calculate values

***Renilla*-Glo Reagent Addition**

1. In a BSC, transfer *Renilla*-Glo substrate- buffer solution to sterile media reservoir.
2. Program E1-Clip-tip Equalizer 384 pipette 12 channel 125 µL volume:
 - Aspirate 120 µL volume; 12 µL volume overfill; speed 5
 - Multi-dispense 30 µL; dispense 4 times; speed 7
 - Purge speed 10.
1. Aspirate *Renilla*-Glo Reagent, dispense first 30 µL back into reservoir, multi-dispense 30 µL to wells B2-O23 of the 384-well assay plates. Discard tips after each aspiration cycle.
2. Incubate plates in dark at room temperature for 15±2 minutes. Start timer once *Renilla*-Glo Reagent has been added to all plates. Record the order of addition *Renilla*-Glo into the plates in the ELN.
3. Read Plates on Enspire
4. If detecting multiple plates, load assay plates in cell stacker in order of *Renilla*-Glo reagent addition
5. Run program.
6. Save raw data files.
7. Proceed to Zika RVP Microneutralization Assay for data regression and analysis.
8. Note: LP for analysis is available upon request.

Appendix

Appendix 1. Zika RVP Microneutralization Assay Calculation Formulas:

1.1 RVP Preparation:

- $\text{Dilution} = (\text{Total Volume}) \div (\text{Optimized RVP (uL/well)})$
- Optimized RVP amount per well
- $\text{\#wells/plate} = \text{number of wells occupied with sample per assay plate}$
- $\text{\# plates} = \text{number of plates utilized in experiment}$
- $\text{Total number of wells} = (\text{\#wells/plate}) \times \text{\# plates}$
- $\text{Volume diluted RVP/well (uL)} = 7.5 \text{ volume of RVP to add to each well}$
- $\text{Volume diluted RVP required (uL)} = (\text{Total \# wells} \times \text{volume diluted RVP/well (uL)}) \times 1.25$
- $\text{Volume Prepared (uL)} = \text{Diluted RVP required rounded to nearest whole number}$
- $\text{Volume RVP} = (\text{Volume Prepared} \div \text{Volume diluted RVP/well (uL)}) \times \text{Optimized RVP (uL/well)}$
- $\text{Volume Assay Media (uL)} = \text{Volume Prepared (uL)} - \text{Volume RVP (uL)}$

1.2 Cell Preparation

- $\text{Optimized cells/well} = 4625$
- $\text{Cells/mL} = ((\text{Optimized cells/well}) / (\text{Volume of cells added/well})) \times 1000$
- $\text{Total number of wells} = \text{Number of plates} \times \text{Number of wells/plate}$
- $\text{Volume Required} = (\text{Total wells} \times 15 \times 1.3) / (1000)$
- $\text{Volume Prepared} = \text{Round Volume Required up}$

1.3 Renilla-Glo Preparation

- $\text{\#wells} = \text{numbers of wells per assay plate}$
- $\text{\#plates} = \text{number of 384 plates used in the assay}$
- $\text{Buffer (mL)} = (\text{\#wells}) \times (\text{\#plates}) \times (0.03 \text{ mL}) \times (1.25 \text{ overage})$
- $\text{Substrate (mL)} = \text{Buffer (mL)} \div 100$

Appendix 2. RVP Neut Calc Excel Workbook

2.1 RVP Neut Calc - Dilution Plate set up Tab

Serum sample pre-dilution worksheet:				
Dilution Plate #	Sample ID	Barcode	Subject ID	Pre-dilution
Dilution Plate #1	Positive Control	N/A		
	1			
	2			
	3			
	4			
	5			
	6			
Dilution Plate #2	7			
	8			
	9			
	10			
	11			
	12			
	Negative Control	N/A		
Dilution Plate #3	Positive Control	N/A		
	13			
	14			
	15			
	16			
	17			
	18			
Dilution Plate #4	19			
	20			
	21			
	22			
	23			
	24			
	Negative Control	N/A		
Dilution Plate #5	Positive Control	N/A		
	25			
	26			
	27			
	28			
	29			
	30			
Dilution Plate #6	31			
	32			
	33			
	34			
	35			
	36			
	Negative Control	N/A		
Dilution Plate #7	Positive Control	N/A		
	37			
	38			
	39			
	40			
	41			
	42			
Dilution Plate #8	43			
	44			
	45			
	46			
	47			
	48			
	Negative Control	N/A		
Dilution Plate #9	Positive Control	N/A		
	49			
	50			
	51			
	52			
	53			
	54			
Dilution Plate #10	55			
	56			
	57			
	58			
	59			
	60			
	Negative Control	N/A		
Dilution Plate #11	Positive Control	N/A		
	61			
	62			
	63			
	64			
	65			
	66			
Dilution Plate #12	67			
	68			
	69			
	70			
	71			
	72			
	73			
Dilution Plate #13	74			
	75			
	76			
	77			
	78			
	79			
	80			
Dilution Plate #14	81			
	82			
	83			
	84			
	85			
	86			
	87			
Dilution Plate #15	88			
	89			
	90			
	91			
	92			
	93			
	94			
Dilution Plate #16	95			
	96			
	97			
	98			
	99			
	100			
	101			
Dilution Plate #17	102			
	103			
	104			
	105			
	106			
	107			
	108			
Dilution Plate #18	109			
	110			
	111			
	112			
	113			
	114			
	115			
Dilution Plate #19	116			
	117			
	118			
	119			
	120			
	121			
	122			
Dilution Plate #20	123			
	124			
	125			
	126			
	127			
	128			
	129			
Dilution Plate #21	130			
	131			
	132			
	133			
	134			
	135			
	136			
Dilution Plate #22	137			
	138			
	139			
	140			
	141			
	142			
	143			
Dilution Plate #23	144			
	145			
	146			
	147			
	148			
	149			
	150			
Dilution Plate #24	151			
	152			
	153			
	154			
	155			
	156			
	157			
Dilution Plate #25	158			
	159			
	160			
	161			
	162			
	163			
	164			
Dilution Plate #26	165			
	166			
	167			
	168			
	169			
	170			
	171			
Dilution Plate #27	172			
	173			
	174			
	175			
	176			
	177			
	178			
Dilution Plate #28	179			
	180			
	181			
	182			
	183			
	184			
	185			
Dilution Plate #29	186			
	187			
	188			
	189			
	190			
	191			
	192			
Dilution Plate #30	193			
	194			
	195			
	196			
	197			
	198			
	199			
Dilution Plate #31	200			
	201			
	202			
	203			
	204			
	205			
	206			
Dilution Plate #32	207			
	208			
	209			
	210			
	211			
	212			
	213			
Dilution Plate #33	214			
	215			
	216			
	217			
	218			
	219			
	220			
Dilution Plate #34	221			
	222			
	223			
	224			
	225			
	226			
	227			
Dilution Plate #35	228			
	229			
	230			
	231			
	232			
	233			
	234			
Dilution Plate #36	235			
	236			
	237			
	238			
	239			
	240			
	241			
Dilution Plate #37	242			
	243			
	244			
	245			
	246			
	247			
	248			
Dilution Plate #38	249			
	250			
	251			
	252			
	253			
	254			
	255			
Dilution Plate #39	256			
	257			
	258			
	259			
	260			
	261			
	262			
Dilution Plate #40	263			
	264			
	265			
	266			
	267			
	268			
	269			
Dilution Plate #41	270			
	271			
	272			
	273			
	274			
	275			
	276			
Dilution Plate #42	277			
	278			
	279			
	280			
	281			
	282			
	283			
Dilution Plate #43	284			
	285			
	286			
	287			
	288			
	289			
	290			
Dilution Plate #44	291			
	292			
	293			
	294			
	295			
	296			
	297			
Dilution Plate #45	298			
	299			
	300			
	301			
	302			
	303			
	304			
Dilution Plate #46	305			
	306			
	307			
	308			
	309			
	310			
	311			
Dilution Plate #47	312			
	313			
	314			
	315			
	316			
	317			
	318			
Dilution Plate #48	319			
	320			
	321			
	322			
	323			
	324			
	325			
Dilution Plate #49	326			
	327			
	328			
	329			
	330			
	331			
	332			
Dilution Plate #50	333			
	334			
	335			
	336			
	337			
	338			
	339			
Dilution Plate #51	340			
	341			
	342			
	343			
	344			
	345			
	346			
Dilution Plate #52	347			
	348			
	349			
	350			
	351			
	352			
	353			
Dilution Plate #53	354			
	355			
	356			
	357			
	358			
	359			
	360			
Dilution Plate #54	361			
	362			
	363			
	364			
	365			
	366			
	367			
Dilution Plate #55	368			
	369			
	370			
	371			
	372			
	373			
	374			
Dilution Plate #56	375			
	376			
	377			
	378			
	379			
	380			
	381			
Dilution Plate #57	382			
	383			
	384			
	385			
	386			
	387			
	388			
Dilution Plate #58	389			
	390			
	391			
	392			
	393			
	394			
	395			
Dilution Plate #59	396			
	397			
	398			
	399			
	400			
	401			
	402			
Dilution Plate #60	403			
	404			
	405			
	406			
	407			
	408			
	409			
Dilution Plate #61	410			
	411			
	412			
	413			
	414			
	415			
	416			
Dilution Plate #62	417			
	418			
	419			
	420			
	421			
	422			
	423			
Dilution Plate #63	424			
	425			
	426			
	427			
	428			
	429			
	430			
Dilution Plate #64	431			
	432			
	433			
	434			
	435			
	436			
	437			
Dilution Plate #65	438			
	439			
	440			
	441			
	442			
	443			
	444			
Dilution Plate #66	445			
	446			
	447			
	448			
	449			
	450			
	451			
Dilution Plate #67	452			
	453			
	454			
	455			
	456			
	457			
	458			
Dilution Plate #68	459			
	460			
	461			
	462			
	463			
	464			

Daughter Plate set-up: 384 well assay plate																								
P1	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																								
B																								
C																								
D																								
E																								
F																								
G																								
H																								
I																								
J																								
K																								
L																								
M																								
N																								
O																								
P																								
P2	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																								
B																								
C																								
D																								
E																								
F																								
G																								
H																								
I																								
J																								
K																								
L																								
M																								
N																								
O																								
P																								
P3	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																								
B																								
C																								
D																								
E																								
F																								
G																								
H																								
I																								
J																								
K																								
L																								
M																								
N																								
O																								
P																								

P4		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																									
B																									
C																									
D																									
E																									
F																									
G																									
H																									
I																									
J																									
K																									
L																									
M																									
N																									
O																									
P																									

P5		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																									
B																									
C																									
D																									
E																									
F																									
G																									
H																									
I																									
J																									
K																									
L																									
M																									
N																									
O																									
P																									

P6		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24
A																									
B																									
C																									
D																									
E																									
F																									
G																									
H																									
I																									
J																									
K																									
L																									
M																									
N																									
O																									
P																									

2.3 RVP Neut Calc- Neut Assay Conditions Tab

Cell Type	Neut Time (min)	Assay Time (hrs)
vero brel	60	72hrs

2.4 RVP Neut Calc- RVP Tab

Zika RVP											
RVP#	Lot#	dilution	Optimized RVP ul/well	# wells/plates	# plates	Total # wells	Volume diluted RVP/well (ul)	Volume diluted RVP required (ul)	Volume Prep(ul)	Volume RVP(ul)	Volume Assay Media(ul)
SHP2015 Zika RVP	229A	12.0	2.5	322	6	1932	7.5	18113	18113	6038	12075

2.5 RVP Neut Calc v03- Cells Tab

Vero Brell	N/A	N/A			
Optimized Cells well	4625	N/A			
cells / ml	308333	N/A			
# wells/plate	322	N/A			
# plates	6	N/A			
Total wells	1932	N/A			
Volume req.	38	mls			
Volume Prepared	40	mls			
Cells Req.	12333333	cells			
Harvest Date					
Passage Number					
Date of Passage					
Flask Size	t175				
Flask Media Volume	40				
Set	Cells/mL total	Cells/ mL live	% live		
count 1					
count 2					
Average (including additional dilution factor*)	NA	#DIV/0!	#DIV/0!		
	Amount of cells (mL)			Assay Media volume(mL)	
	#DIV/0!	mL	in	#DIV/0!	mL

2.6 RVP Neut Calc- Renilla-Glo Reagent Tab

Renilla Glo Buffer	
#wells	322
#plates	6
Buffer (mL)	72.5
Substrate(mL)	0.725