AUG 15, 2019

HLA genotyping using SS-SBT methods

PLOS One

Ryosuke Tashiro¹, Hidetoshi Inoko², Kuniyasu Niizuma¹,
Teiji
Tominaga¹

¹Tohoku University; ²Genitive Pharma Inc.

Ryosuke Tashiro
Tohoku University

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.3rxgm7n](https://doi.org/10.17504/protocols.io.3rxgm7n)

External link:
<https://doi.org/10.1371/journal.pone.0220858>

Protocol Citation: Ryosuke Tashiro, Hidetoshi Inoko, Kuniyasu Niizuma, Teiji Tominaga 2019. HLA genotyping using SS-SBT methods. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.3rxgm7n>

MANUSCRIPT CITATION:

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Created: Jun 05, 2019

Last Modified: Aug 15, 2019

ABSTRACT

Genomic DNA was obtained from the patients' 2ml of whole blood using the QIAamp DNA Mini Kit for genomic DNA purification (Qiagen GmbH, Hilden, Germany), and 400 ng of purified genomic DNA was used for polymerase chain reaction (PCR) amplification. DNA was preserved in 4°C freezer. The basic cycling parameters were as follows: (i) first denaturation at 94°C for 2 min, followed by 30 cycles of denaturation at 98°C for 10 s and 60°C for 20 s and extension at 68°C for 5 min (*HLA-A*, *HLA-B*, and *HLA-C*); (ii) first denaturation at 94°C for 2 min, followed by 30 cycles of denaturation at 98°C for 10 s and annealing at 70°C for 5 min (*HLA-DRB1* and *HLA-DPB1*); and (iii) first denaturation at 94°C for 2 min, followed by 30 cycles of denaturation at 98°C for 10 s and annealing at 70°C for 9 min (*HLA-DQB1*). Long-range PCR reactions were performed using the thermal cycler Gene Amp PCR System 9700 (Life Technologies, Carlsbad, CA, USA). The PCR products obtained were purified with Agencourt AMPure XP (Beckman Coulter, CA, USA) and quantified by the Quant-iT Picogreen dsDNA Assay Kit (Thermo Fisher Scientific, MA, USA). Next, the PCR products were clonally amplified and barcoded using the Ion Plus Fragment Library Kit (Life Technologies), and the barcoded library was sequenced using the Ion Torrent Personal Genome Machine DNA sequencing system (Life Technologies). The NGS read data were analyzed by Sequence Alignment Based Assigning Software (SeaBass), and finally, the *HLA* alleles were determined.

MATERIALS

Ion Plus Fragment Library kit (Thermo Fisher Scientific)
Ion PGM Hi-Q View Sequencing kit (Thermo Fisher Scientific)
Ion Chip kits, Ion 318, Chip v2 BC, Ion 316 Chip v2 BC or Ion 314 Chip v2 BC (Thermo Fisher Scientific)

PROTOCOL integer ID:
24087

BEFORE START INSTRUCTIONS

Keywords: HLA genotyping,
SS-SBT, NGS

Details of experimental procedure are shown in the following paper: Shiina T, Suzuki S, Kuslki JK, Inoko H. Super high resolution for single molecule-sequence-based typing of classical HLA loci using Ion Torrent PGM. *Methods Mol Biol* 2018;1802:115-133.

1 DNA extraction

2 Long-ranged PCR

CITATION

Shiina T, Suzuki S, Kuslki JK, Inoko H (2018). Super high resolution for single molecule-sequence-based typing of classical HLA loci using Ion Torrent PGM. *Methods Mol Biol*.

LINK

[10.1007/978-1-4939-8546-3_8](https://doi.org/10.1007/978-1-4939-8546-3_8)

2.1 Firsts denature 94 °C
00:02:00

2.2 Denature 98 °C 00:00:10

2.3 Annealing (30 cycles)

60 °C HLA-A, B,C	70 °C HLA-DRB1	70 °C HLA-DQB1	00:09:00 HLA-DQB1
00:02:00 HLA-A, B, C	00:05:00 HLA-DRB1		

2.4

Extension (30cycles)

68 °C HLA-
A,B,C

00:05:00 HLA-
A,B,C

3

Construction of barcoded library

CITATION

Shiina T, Suzuki S, Kulski JK, Inoko H (2019). Super high resolution for single molecule-sequencebased typing of classical HLA loci using Ion Torrent PGM. Methods Mol Biol.

LINK

[10.1007/978-1-4939-8546-3_8](https://doi.org/10.1007/978-1-4939-8546-3_8)

4

Preparation of the Enriched Template-Positive Ion Sphere particles (ISPs)

CITATION

Shiina T, Suzuki S, Kulski JK, Inoko H (2019). Super high resolution for single molecule-sequence-based typing of classical HLA loci using Ion Torrent PGM. Methods Mol Biol.

LINK

[10.1007/978-1-4939-8546-3_8](https://doi.org/10.1007/978-1-4939-8546-3_8)

5

Sequencing

CITATION

Shiina T, Suzuki S, Kulski JK, Inoko H (2018). Super high resolution for single molecule-sequence-based typing of classical HLA loci using Ion Torrent PGM. Mol Methods Biol.

LINK

[10.1007/978-1-4939-8546-3_8](https://doi.org/10.1007/978-1-4939-8546-3_8)

6

Data analysis

CITATION

Shiina T, Suzuki S, Kulski JK, Inoko H (2018). Super high resolution for single molecule-sequence-based typing of classical HLA loci using Ion Torrent PGM. *Methods Mol Biol.*

LINK

[10.1007/978-1-4939-8546-3_8](https://doi.org/10.1007/978-1-4939-8546-3_8)

- 6.1** Output of NGS read data
- 6.2** Homology search using Blat
- 6.3** Selection of allele candidates
- 6.4** Mapping of reads and candidate allele sequence
- 6.5** Calculation of coverage
- 6.6** Final confirmation