VERSION 2

NOV 05, 2018

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocol.s.io.vbye2pw

External link:

<https://www.hobodataloggers.com.au/hoboware-free-download>

Protocol Citation: Sophie Martin 2018. HOB0 data loggers. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.vbye2pw>

MANUSCRIPT CITATION:

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Nov 05, 2018

HOB0 data loggers V.2

Sophie Martin¹

¹Station Biologique de Roscoff

EDYCO

Sophie Martin

Station Biologique de Roscoff

ABSTRACT

This protocol is intended to help you use HOB0 data loggers and HOB0ware software.

ATTACHMENTS

[HOB0ware-User's-Guide.pdf](#)

[HOB0ware-Getting-Started-Guide.pdf](#)

MATERIALS

- HOB0 Logger
- USB base station/coupler to connect to the computer
- HOB0ware software installed on the computer: HOB0 ware is a FREE Download at the following address: <https://www.hobodataloggers.com.au/hoboware-free-download>

Last Modified: Nov 05, 2018

PROTOCOL integer ID: 17496

Keywords: HOB0, Field survey, Light, Temperature

Connection

- 1 Connect the HOB0 logger to the USB Base Station connected to the computer.

Safety information

Be careful to insert correctly the HOB0 data logger in the Base Station.
Push it thoroughly in the Base Station, making sure that the HOB0 data logger fits into the corresponding slot on the Base Station.

HOB0ware

- 2 Open the HOB0ware software (FREE download)

Software

HOB0ware	NAME
----------	------

HOBOWare Basic functions

3 4 icons at the top left:

1. First icon: **Launching** (Beware the launch of the HoBo probe overwrites the data stored there)
2. Second icon: **Reading out** (retrieve saved data)
3. Third icon: **Logger status** (allows to check the level of charge of the battery and the used memory)
4. Fourth icon: **Stop** the device

HOBO data logger Programing

4 Click the **Launching** icon on the toolbar (**Icon 1**)

The Plot Setup window contains the following information:

- **Description:** Indicate the file name
- **Parameters:** Select Temperature and Light Intensity
- **Recording interval:** "5 seconds" for example (this gives below the possible recording time depending on the memory)

To **Start** recording:

- Click on "Start" (Immediate) or "Delayed start" (Postponed, indicate date / time)

HOBO data logger Reading and Data export

5

Reading out the logger and plotting data

- Click the **Read out** icon on the toolbar (**Icon 2**)
- Choose the location to save the datafile.
- Choose the **units** of measurement (Temperature in °C and Light Intensity in Lux, leave the rest checked)
- Click the **Plot** button

To save the file in excel format: click on the Excel icon (icon n°12)

A window appears.
Leave everything checked.

Open the "Preference / Settings" icon (third icon)

Data Export:

- Export file type: **Excel**
- Separate date and time (No: it will keep date and time grouped, facilitating graphics over several days)
- Time format: 24 h
- Do not check "**Export** automatically ...", do it **manually**.
- Click on "OK" then "Export" and "Save" the file.