

Sep 01, 2024

Version 1

ImageJ Analyses of Mitochondria (HSP60) in TH+ cells V.1

DOI

<https://dx.doi.org/10.17504/protocols.io.6qpvr8y83lmk/v1>

Chiara Pavan¹, Clare Parish¹

¹Florey Institute of Neuroscience and Mental Health



courtney.wright Wright

University of Sydney

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.6qpvr8y83lmk/v1>

Protocol Citation: Chiara Pavan, Clare Parish 2024. ImageJ Analyses of Mitochondria (HSP60) in TH+ cells. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.6qpvr8y83lmk/v1>

License: This is an open access protocol distributed under the terms of the **[Creative Commons Attribution License](#)**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development

We are still developing and optimizing this protocol

Created: August 30, 2024



Last Modified: September 01, 2024

Protocol Integer ID: 106734

Keywords: ASAPCRN, mitochondria, analysis, iPSC, image analysis, HSP60, imagej analyses of mitochondria, analyse mitochondria, dopaminergic neurons in culture, mitochondria, fiji to analyse mitochondria, derived dopaminergic neuron, hsp60, cell, imagej analysis, parameters such as network branching

Funders Acknowledgements:

Michael J Fox Foundation

Grant ID: ASAP-000497

Abstract

This protocol explains how to use Fiji to analyse mitochondria in iPSC derived dopaminergic neurons in culture for parameters such as network branching, sphericity and mean total branch length.

Image Attribution

The authors would like to acknowledge the inputs of Dr. Leo Lam and Associate Professor Michael Lazarou in development of this protocol.

Before start

Requirements prior to analysis

- Install Fiji
- Locate images to analyse (HSP60+TH staining)
- Install Mitochondria Analyzer Plugin (see references)

Image acquisition


- 1 Images analysed with this protocol were taken with Leica Stellaris 8 confocal (for mitochondrial morphology and ATG13-HSP60 co-localization, with FLIM 40x/1.30 oil immersion objective, HC PL APO, CS2, using Leica Power HyD detector S and HyD detector X).

All mitochondria morphology images were acquired in 3D by optical sectioning under *Lightning*[™] mode with a z-stack of 1.73µm and maximum voxel size of 94.7 nm laterally (x, y) and 350 nm axially (z). For each experimental sample 10 cells per sample were analyzed.


Note: Deconvolved images were used for the mitochondrial morphology analysis.

Analysis pipeline

- 2 1. Open the image




2. In the TH channel (use the bar under the figure to go to the TH channel, here channel 2), use the Freehand Selection tool to mark the cell of interest



If the TH signal is not bright enough, go to Image>Adjust > color balance (only in the TH channel, don't modify the other channels)

3. Save the region of interest selected by going to Analyze > Tool > ROI manager > Add [t]
4. Duplicate the HSP60 channel by clicking outside the selected ROI, then going to Image > Duplicate; select Channel 1 and tick hyperstack
5. Clear the staining outside the ROI by selecting the ROI (press on the ROI name on the ROI manager panel), then going to Edit > Clear Outside > Yes
6. Convert the image to 8 bit by going to Image>Type>8bit
7. To optimize the Threshold, go to Plugins > Mitochondria Analyzer > 3D Threshold Optimize (for more info, see below)

NOTE: This may take some time



Looking back at the original picture, you can use this Threshold Optimize Test picture to decide the values of Block Size, and c values that better identify the true signal.

8. Once decided the optimal values, go back to your picture (the one with only HSP60, duplicated, cleared), and apply the chosen parameters by going to Plugins > Mitochondria Analyzer > 3D Threshold and inserting the desired parameters.

NOTE: This may take some time

Protocol references

<https://github.com/AhsenChaudhry/Mitochondria-Analyzer/blob/master/README.md>