JUL 27, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.5qyg5xn](https://doi.org/10.17504/protocols.io.5qyg5xn)

External link:
<https://doi.org/10.1371/journal.pone.0223636>

Protocol Citation: Seth H. Frisbie, Erika J. Mitchell, Stéphane Roudeau, Florelle Domart, Asuncion Carmona, Richard Orterga 2019. Infant formula sample preparation for element analysis using PIXE/RBS. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.5qyg5xn>

MANUSCRIPT CITATION: Frisbie SH, Mitchell EJ, Roudeau S, Domart F, Carmona A, Orterga R (2019) Manganese levels in infant formula and young child nutritional beverages in the United States and France: Comparison to breast milk and regulations. PLoS ONE 14(11): e0223636. doi: [10.1371/journal.pone.0223636](https://doi.org/10.1371/journal.pone.0223636)

Infant formula sample preparation for element analysis using PIXE/RBS

PLOS One

Seth H. Frisbie¹, Erika J. Mitchell², Stéphane Roudeau³, Florelle Domart³, Asuncion Carmona³, Richard Orterga³

¹Department of Chemistry and Biochemistry, Norwich University, Northfield, VT, USA;

²Better Life Laboratories, Inc., East Calais, VT, USA;

³University of Bordeaux, Centre d'Etudes Nucléaires de Bordeaux Gradignan (CENBG), CNRS, Gradignan, France

Richard Orterga

ABSTRACT

This protocol was developed to quantify chemical elements from infant formula samples by PIXE/RBS (Particle Induced X-ray Emission / Rutherford Backscattering Spectrometry). This protocol aims to analyze powdered samples. Liquid samples can also be prepared beginning from step 6.

MATERIALS

MATERIALS

Water Optima™ LC/MS Grade Fisher Scientific **Contributed by users Catalog**
#10728098

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Jul 23, 2019

Last Modified: Jul 27, 2019

PROTOCOL integer ID:
26103

Keywords: Infant formula,
trace element, manganese,
PIXE, RBS

- 1 Work under a clean atmosphere to avoid dust contamination.
- 2 Use a disposable plastic spatula to deliver between 1.0000 grams (g) and 1.0099 g of powdered product to a laboratory-grade plastic centrifuge tube.
- 3 Add 10.00 milliliters (mL) of ultra-trace elemental analysis grade water (Water Optima™).
- 4 Vortex until the complete dissolution or uniform suspension of powder.
- 5 Store the solution or suspension at 4°C for a maximum of 24 hours.

- 6 Immediately before using, vortex the solution again.
 - 7 Deliver 1.0 microliter (μL) of the solution or suspension with an air displacement pipet onto a labeled PIXE/RBS sample holder.
 - 8 Repeat step 7 until a total of three different drops from the same sample are placed in three different locations of the same sample holder.
- Note**
- Use a different pipette tip to deliver each drop. Always vortex immediately before filling the pipette tip.
- 9 After delivering the three 1.0 μL drops onto the three different locations on the sample holder, store the sample solution or suspension at 4 °C.
 - 10 DDry the drops at room temperature in a laminar flow hood for 24h.
 - 11 Vortex the solution.
 - 12 Superimpose onto each dried drop another 1.0 μL of solution or suspension to increase the quantity of matter in order to be more sensitive for analysis by PIXE/RBS.

Note

Use a different pipette tip to deliver each drop. Always vortex immediately before filling the pipette tip.

- 13** Dry samples at room temperature in a laminar flow hood for 24h and store the solution or suspension at 4°C.
- 14** Repeat steps 11, 12 and 13 until each spot on the sample holder is comprised of three superimposed drops of 1.0 µL.
- 15** Store the sample holders with the dried spots in a desiccator and protect from dust until PIXE/RBS analyses.