DEC 03, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.3eggjbw](https://doi.org/10.17504/protocols.io.3eggjbw)

External link:
<https://doi.org/10.1371/journal.pone.0225821>

Protocol Citation: Yodi Soebadi, Marko Bakula, Lukman Hakim, Robert Puers, Dirk De Ridder 2019. KU Leuven Exp Urology - Urodynamics in Female Minipigs. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.3eggjbw>

MANUSCRIPT CITATION: Soebadi MA, Bakula M, Hakim L, Puers R, Ridder DD (2019) Wireless intravesical device for real-time bladder pressure measurement: Study of consecutive voiding in awake minipigs. PLoS ONE 14(12): e0225821. doi: [10.1371/journal.pone.0225821](https://doi.org/10.1371/journal.pone.0225821)

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

🌐 KU Leuven Exp Urology - Urodynamics in Female Minipigs

📖 PLOS One

Yodi

Soebadi^{1,2}, Marko Bakula¹, Lukman Hakim²,

Dirk De

Robert Puers¹, Ridder¹

¹Katholieke Universiteit Leuven; ²Universitas Airlangga

Yodi Soebadi

ABSTRACT

Summary:

This protocol provides information for the invasive urodynamic measurement of bladder pressure, by urethral catheter as well as wireless device (bladder pill).

MATERIALS

Equipment

1. Aquarius TT urodynamics machine – calibrated
2. Calibration tube 30 cm
3. Scale 2 kg
4. Luggage scale
5. Speculum, forceps, needle holder/scissor combo
6. Guidewire
7. Amplatz renal dilator set (antenna, dilator, sheath)
8. Cystoscope, grasping forceps, light cable
9. Telepack/endoscopy camera

Consumable

1. Ketamine 3 vials
2. Xylazine 6 ml
3. Catheter single channel & multi channel (3 new each)
4. NaCl 0.9% 3 L x3
5. Catheter gel
6. Suture polyglactin (vicryl) 3-0
7. Duct tape

Protocol status: Working

Created: May 27, 2019

Last Modified: Dec 03, 2019

PROTOCOL integer ID:
23720

- 1 Preparation: Bring machine and all consumables to room
- 2 Fill calibration tube with saline
- 3 Check calibration with tube
- 4 Open charge switch AND flush balloon
- 5 Measurement: Isolate minipig for study, bring in transport cart & administer sedation: Ketamine 15 mg/kg (Nimatek 100 mg/ml) Xylazine 2 mg/kg (Xyl-M 2%); (at 60-80 kg, 6-8 ml of Nimatek, 1-1.5 ml of xylazine>>
- 6 Wait 10-15 minutes outside of cage until minipig sedated
- 7 Place minipig prone at caudal end,.

- 8 Prepare urethral silicone cath, gel, tape.
- 9 Insert single and triple channel catheter in urethra, insert rectal catheter
- 10 Put belt or frame around midsection
- 11 Set zero on UDS machine then charge catheter.
- 12 Put tape around charge switch
- 13 Tape plastic around posterior and tape funnel
- 14 Wait for animal to wake up (start eating) and start filling at 10% of voided volume (at volume more than 500 ml fill 50 ml/minute)
- 15 Repeat up to 5 voids

16 At end of procedure, anaesthetise and extract pill.