AUG 05, 2020

LC-MS/MS analysis

avinash.kale¹

¹domnic colvin

UM-DAE Centre for Excellence in Basic Sciences

avinash.kale

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bjdvki66

Document Citation: avinash.kale 2020. LC-MS/MS analysis. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bjdvki66>

License: This is an open access document distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Created: Aug 05, 2020

Last Modified: Aug 05, 2020

DOCUMENT integer ID:
40085

LC-MS/MS analysis

Nano-LCMS using 6540 UHD Accurate Mass QTOF (CAMS, Venture Center, NCL Pune, India) was used to carry out all our mass spectrometric analysis. The peptides were analyzed using reversed-phase nanoscale liquid chromatography coupled to tandem mass spectrometry. The system consisted of a column; initially, the sample would be subjected to Agilent Binary LC1290 using column Polaris, high-performance chip, 360 nl enrichment chip (150mm X 75µm), and a separation column G4240-62030 with a thermostat temperature of 4°C. The mobile phase was Water (0.1% Formic Acid) and Acetonitrile: Water (90:10 v/v). The chromatography was carried out in a gradient manner (3% to 90% Acetonitrile: Water). The mass spec analysis was carried out on Agilent 6540 UHD QTOF MS (Gas temp: 250, Gas flow (l/minute): 8 minutes,) MS analysis was carried out in a data-dependent manner with

survey scan resolution. Mass range: 250-1700(m/z), MS/MS: 50-1700(m/z), Data type: Centroid and software: Masshunter workstation software. Obtained data were searched against the UniProt protein databases of all the respective four *Bacillus* sp. using default parameters.