VERSION 2

APR 29, 2020

LC-MS/MS Label-Free Proteomic Data Analysis Parameters V.2

Danielle Gutierrez¹, Jamie Allen¹, Zach Jenkins¹, Jeff Spraggins¹

¹Vanderbilt University

VU Biomolecular Multimodal Imaging Center / Spraggins Research Group

Human BioMolecular Atlas Program (HuBMAP) Method Development

Jamie Allen
Vanderbilt University

ABSTRACT

List of parameters and settings for searching label free proteomic data in MaxQuant version 1.6.7.

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bfshjnb6

Protocol Citation: Danielle Gutierrez, Jamie Allen, Zach Jenkins, Jeff Spraggins 2020. LC-MS/MS Label-Free Proteomic Data Analysis Parameters. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bfshjnb6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Apr 29, 2020

Last Modified: Apr 08, 2023

PROTOCOL integer ID:
36393

Keywords: HuBMAP, BIOMIC, MSRC, Vanderbilt, Proteomics, MaxQuant, Data Analysis

- 1 Label-free proteomic samples were searched using MaxQuant version 1.6.7.
- 2 *Group Specific Parameter settings included:*
Standard
Multiplicity = 1
Variable modifications: Oxidation (M); Acetylation (Protein N-term), Carbamidomethyl (C)
Fixed modifications: Carbamidomethyl (C)
Max number of modifications per peptide: 5
Instrument settings: Orbitrap (default settings)
Digestion: Specific, Trypsin /P
Max missed cleavages: 2
LFQ: None
- 3 *Global Parameters included:*
Database: UniProt Human Proteome (UP000005640, 9606) Reviewed, downloaded July 30, 2019
Min peptide length: 7
Max peptide mass: 4600 Da
MS/MS analyzer: Default settings
Identification: Default settings