VERSION 3

MAR 05, 2021

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bsu5ney6

Protocol Citation: Danielle Gutierrez¹, Jamie Allen¹, Zach Jenkins, Jeff Spraggins 2021. LC-MS/MS Label-Free Proteomic Data Analysis Parameters. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bsu5ney6> Version created by Danielle Gutierrez

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Feb 28, 2021

Last Modified: Apr 08, 2023

PROTOCOL integer ID:
47741

Keywords: HuBMAP, BIOMIC, MSRC, Vanderbilt, Proteomics, MaxQuant, Data Analysis

LC-MS/MS Label-Free Proteomic Data Analysis Parameters V.3

Danielle Gutierrez¹, Jamie Allen¹, Zach Jenkins¹, Jeff Spraggins¹

¹Vanderbilt University

VU Biomolecular Multimodal Imaging Center / Spraggins Research Group

Human BioMolecular Atlas Program (HuBMAP) Method Development

Jamie Allen
Vanderbilt University

DISCLAIMER

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

List of parameters and settings for searching label free proteomic data in MaxQuant version 1.6.17.

- 1 Label-free proteomic samples were searched using MaxQuant version 1.6.17.
- 2 *Group Specific Parameter settings included:*
Standard
Multiplicity = 1
Variable modifications: Oxidation (M); Acetylation (Protein N-term), Carbamidomethyl (C)
Fixed modifications: Carbamidomethyl (C)
Max number of modifications per peptide: 5
Instrument settings: Orbitrap (default settings)
Digestion: Specific, Trypsin /P
Max missed cleavages: 2
LFQ: None
- 3 *Global Parameters included:*
Database: UniProt Human Proteome (UP000005640, 9606) Reviewed, downloaded July 30, 2019
Min peptide length: 7
Max peptide mass: 4600 Da
MS/MS analyzer: Default settings
Identification: Default settings