SEP 22, 2020

Magnetic bead cleaning of PCR products

Forked from a private protocol

Alexander Eiler¹

¹University of Oslo

AQUA at UiO

Alexander Eiler

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.q2udyew

Protocol Citation: Alexander Eiler 2020. Magnetic bead cleaning of PCR products. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.q2udyew>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Jun 15, 2018

Last Modified: Sep 22, 2020

PROTOCOL integer ID:
 13108

DISCLAIMER

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to [protocols.io](#) is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with [protocols.io](#), can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

GUIDELINES

Work at post-PCR lab.

You can use PCR tubes/strips or PCR-plates but for easiest work with magnetic beads flat-bottom 96-well microplates are recommended.

- 1 Calculate the volume needed for the beads:
 $0.8 * \text{sample volume} =$

- 2 Shake the magnetic bead buffer bottle to fully resuspend magnetic particles.
- 3 Add Sample.
- 4 Add beads.
- 5 Pipette mix 10 times and Incubate at room temperature for 5 minutes.
00:05:00
- 6 Place the reaction plate onto a Magnet Plate for 2 minutes to separate beads from solution. (Keep on the magnet plate until step 11).
00:02:00
- 7 Aspirate the supernatant from the reaction plate and discard.
- 8 Dispense 200 μ L of 70% ethanol and incubate at room temperature for at least 30 seconds. Aspirate out the ethanol and discard. You can discard the ethanol by inverting the plate keeping the magnetic plate on the bottom of it.
200 μ L
00:00:30
- 9 Repeat the washing step as above: dispense 200 μ L of 70% ethanol and incubate at room temperature for at least 30 seconds. Aspirate out the ethanol and discard. You can discard the ethanol by inverting the plate keeping the magnetic plate on the bottom of it.
200 μ L

00:00:30

10 Remove the remaining ethanol by pipetting.

11 Take the plate off the magnet

12 Dry off the remaining ethanol. To speed up the process, you can place the plate on a warm (30°C) heat block.

Drying takes at least 30 minutes. Check that drying was complete before continuing.

00:30:00

13 Add 40 µL of elution buffer, pipette mix 10 times. Make sure the buffer reaches up to the beads!

40 µL

14 Incubate at room temperature for 2 minutes.

00:02:00

15 Place the reaction plate onto a Magnet Plate for 1 minute to separate beads from solution.

00:01:00

16 Transfer purified product to a new PCR plate or strips.