VERSION 2

NOV 15, 2018

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.vm9e496

Protocol Citation: Andrea L. Lawrence, Maureen Laroche, Jana Batovska, Cameron E. Webb, Stacey E. Lynch, Mark J. Blacket, Jan Šlapeta, Philippe Parola 2018. MALDI-TOF MS preparation for identification of mosquitoes. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.vm9e496>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Nov 15, 2018

Last Modified: Nov 15, 2018

PROTOCOL integer ID:
17825

🌐 MALDI-TOF MS preparation for identification of mosquitoes V.2

Andrea L. Lawrence^{1,2}, Maureen Laroche³, Jana Batovska⁴, Cameron E. Webb^{1,2}, Stacey E. Lynch⁴, Mark J. Blacket⁴, Jan Šlapeta⁵, Philippe Parola³

¹Marie Bashir Institute of Infectious Diseases and Biosecurity, University of Sydney, Sydney, New South Wales 2006, Australia;

²Medical Entomology, NSW Health Pathology, ICPMR, Westmead Hospital, Westmead, New South Wales 2145, Australia;

³Aix Marseille University, IRD, AP-HM, SSA, VITROME, IHU-Méditerranée Infection, 19-21 Boulevard Jean Moulin 13005 Marseille, France;

⁴Agriculture Victoria, AgriBio, Centre for AgriBioscience, Bundoora, Victoria 3083, Australia;

⁵Sydney School of Veterinary Science, Faculty of Science, University of Sydney, Sydney, New South Wales 2006, Australia

Andrea Lawrence

ABSTRACT

MALDI-TOF MS specimen preparation for rapid identification of Australian mosquitoes

GUIDELINES

There are 3 main steps involved with preparing and launching a mosquito sample for MALDI-TOF analysis:

- Preparing the matrix

The matrix solution ionises the proteins in the sample and is applied to the plate after the sample. The matrix must be very acidic.

- Dissecting and crushing the sample
- Preparing and launching the plate

Keywords: protein profiling, identification, mosquito, Australia, taxonomy, MALDI-TOF MS, biosecurity

MATERIALS

MATERIALS

- ✘ 1.5 mL Eppendorf tubes **Contributed by users**
- ✘ Acetonitrile **Bio Basic Inc. Catalog #AC1400.SIZE.1L**
- ✘ Trifluoroacetic acid (TFA) **Bio Basic Inc. Catalog #TC8960.SIZE.100mL**
- ✘ Water, uHPLC grade **Contributed by users**
- ✘ α -cyano-4-hydroxycinnamic acid **Sigma Aldrich Catalog #C8982 SIGMA**
- ✘ Formic acid, 70% **Contributed by users**
- ✘ Glass beads, acid washed $\leq 106\mu\text{m}$ **Sigma Aldrich Catalog #G4649-500G**

- 1 Preparing the matrix
- 2 1. Using a small spatula, put approx. 50 μL worth of matrix powder (α -cyano-4-hydroxycinnamic acid) in the Eppendorf tube and return the stock powder to the fridge immediately.
- 3 2. Add 500 μL of 100% ACN (acetonitrile), 475 μL of HPLC grade water and 25 μL of 100% trifluoroacetic acid to the Eppendorf tube.
- 4 3. Put Eppendorf tube with solution in an ultrasonic bath for 10 minutes to homogenise the solution (yellow matrix powder will still be visible); then centrifuge the solution for 10 minutes at 13 000 rpm/ 20 784 x g.
- 5 4. Retrieve the supernatant using a pipette, taking care to avoid taking up any powder. Discard the powder. The matrix can be kept at 4 °C in the dark for up to 2-3 days when creating high quality spectra for reference database. It can be kept for up to a week when using it for simple identification when high quality spectra are not vital.
- 6 Preparing the sample
- 7 1. If using live mosquitoes, put the mosquitoes in the freezer (-20) for at least 1 minute prior to

dissection.

- 8 2. Using two sets of forceps, carefully remove the legs from the mosquito, using one pair to steady the specimen and the other to remove the legs. Place the legs in a labelled and sterile Eppendorf tube. The legs will stick to the sides of the tube due to static so try to put the legs at the bottom of the tube.
- 9 3. If more than one species are being processed, decontaminate the forceps with 70% ethanol between species.
- 10 4. Spin down the legs for 1 minute at 13 000 rpm/ 20 784 x g.
- 11 5. Add 15µl of 50% CAN and 15µl of 70% formic acid to each tube then add glass beads (<106µm, acid washed). Estimate the volume, making the amount of glass beads just less than half of the volume of liquid.
- 12 6. Place tubes in the TissueLyser II (Qiagen) canisters, making sure each side is balanced. NB: Both the tubes and the canisters need to be balanced. Make sure the top and bottom lids correlate and that the heavy sides are both either facing inward or both outward.
- 13 7. Crush the legs using a cycle of 3 x 1 minute at 30 Hz, letting the sample rest for approx. 15 seconds between each minute. NB: the legs will not be fully crushed in the tube; too much protein will inhibit the reaction as in PCR.
- 14 Preparing the plate
- 15 1. Load 1µl of each sample on the plate in quadruplicate keeping within the spots.

- 16** 2. Allow the sample to evaporate then load 1 µl of matrix solution over the sample, allow to dry completely.
- 17** 3. Place the plate in a cool, dark (photosensitive matrix) place until it's time to launch it. NB: the prepared plates can be left for max. 2-3 days in a dark place at room temperature before affecting the spectra. If using for simple identification – it can be kept for a longer period.
- 18** 4. Plate is ready to be launched in MALDI-TOF MS machine.