

Sep 24, 2020

# McGill Nanopore Native Barcoding LibPrep Protocol, 10 ng NB

 Forked from [nCoV-2019 McGill Nanopore LibPrep Protocol, 10 ng NB](#)



DOI

[dx.doi.org/10.17504/protocols.io.bmijk4cn](https://dx.doi.org/10.17504/protocols.io.bmijk4cn)

Sarah J Reiling<sup>1</sup>, Anne-Marie Roy<sup>1</sup>, Shu-Huang Chen<sup>1</sup>, Ioannis Ragoussis<sup>1</sup>

<sup>1</sup>McGill University

McGill Genome Centre



Sarah J Reiling

McGill University

OPEN  ACCESS



DOI: [dx.doi.org/10.17504/protocols.io.bmijk4cn](https://dx.doi.org/10.17504/protocols.io.bmijk4cn)

**Protocol Citation:** Sarah J Reiling, Anne-Marie Roy, Shu-Huang Chen, Ioannis Ragoussis 2020. McGill Nanopore Native Barcoding LibPrep Protocol, 10 ng NB . **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bmijk4cn>

**License:** This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working

**We use this protocol and it's working**

**Created:** September 19, 2020

**Last Modified:** September 24, 2020

**Protocol Integer ID:** 42283

**Keywords:** nanopore native barcoding,

## Abstract

This protocol works for 12 native barcodes, 24 native barcodes, and 96 native barcodes.



## Before you start

1

### Note

For native barcoding library preparations, it is highly recommended to always add a libprep negative control. Because we have two subsequent ligation reactions, any leftover native barcodes from the first ligation step may bind to end-prepped DNA strands in the second ligation step.

To reduce the presence of leftover native barcodes, we perform two bead cleanup steps after the first ligation reaction. If in doubt, run a gel or Tapestation to verify that no barcodes are present before continuing with the second ligation step.

## Native barcoding

2






Barcode the samples using native barcodes. ONT recommends to load 100 - 200 fmol total concentration. We recommend to start with 100-fold this concentration at the beginning of the library preparation due to losses and normalization steps.

### Note

This is a 'one-pot ligation' protocol for native barcoded ligation libraries, which means that we do not clean up the end-prepped product; instead, it is added as is to the native barcoding step. We have seen no reduction in performance compared to standard libraries, and is made faster by using the Ultra II® ligation module which is compatible with the Ultra II® end repair/dA-tailing module removing a clean-up step.

3

Set up the following reaction for each sample:

Component	Volume
DNA amplicons	 5 µL
Nuclease-free water	 7.5 µL
Ultra II End Prep Reaction Buffer	 1.75 µL
Ultra II End Prep Enzyme Mix	 0.75 µL
<b>Total</b>	 15 µL

4



Incubate at room temperature for 00:10:00

Incubate at 65 °C for 00:05:00

Incubate on ice for 00:01:00

5

Add the following directly to the previous reactions:

Component	Volume
NBXX barcode	2.5 µL
Ultra II Ligation Master Mix	10 µL
Ligation Enhancer	0.3 µL
<b>Total</b>	17 µL 20 ul total with 3 ul from step 3

6

Incubate at room temperature for 00:15:00

Incubate at 70 °C for 00:10:00

Incubate on ice for 00:01:00

Note

The 70°C incubation is to inactivate the DNA ligase to prevent barcode cross-ligation when reactions are pooled in the next step.

7

Clean-up the native barcoded samples using the following protocol:

Add 0.8X of SPRI beads (16 ul) to the sample tube and mix gently by pipetting.

Incubate for 5 min at room temperature.

Pellet on magnet for 5 min. Remove supernatant.

Add 200 ul of 80% ethanol to the pellet and wash twice.

Elute in **20 ul** elution buffer.

8

Quantify the barcodes using a fluorimetric dsDNA assay.

9

Normalize the barcodes to **10 ng** each and pool (except neg. ctrl - which will be equal volume).

10





Clean-up the barcode pool using the following protocol.

Add 0.8X of SPRI beads to the sample tube and mix gently by pipetting.

Incubate for 5 min at room temperature.

Pellet on magnet for 5 min. Remove supernatant.






Add 200 ul of 80% ethanol to the pellet and wash twice.

Elute in **30 ul** elution buffer.

11 **Optional:** Run a tapestation to verify that no unligated barcodes are present in the elute.

12

Set up the following AMII adapter ligation reaction:

Component	Volume
Barcoded amplicon pools	 30 µL
NEBNext Quick Ligation Reaction Buffer (5X)	 10 µL
AMII adapter mix	 5 µL
Quick T4 DNA Ligase	 5 µL
<b>Total</b>	 50 µL

13

Incubate at room temperature for  00:15:00

14

Clean-up the native barcodes using the following protocol:

Add an equal volume (1:1) of SPRI beads to the sample tube and mix gently by pipetting.

Incubate for 5 min at room temperature.

Pellet on magnet for 5 min. Remove supernatant.

Add 200 ul of **SFB or LFB (depending on your desired fragment size)** to the pellet and **resuspend beads completely by pipette mixing**.

Pellet on beads, remove supernatant, and repeat the wash step with another 200 ul of **SFB or LFB**.

Elute in 15 ul **EB (provided in the ONT kit)**.

Incubate at room temperature for  00:02:00

Place on magnetic rack.

Transfer final library to a new 1.5 mL Eppendorf tube.

**Note**

SFB will remove excess adapter without damaging the adapter-protein complexes. Do not use 70% ethanol as in early clean-ups.

- 15 Quantify the final library using a fluorimetric dsDNA assay.

**Note**

Final library can now be stored in 10 mM Tris pH 8 at 4°C for up to a week if needed otherwise proceed directly to MinION sequencing.

- 16 Prime the flowcell and load 100 - 200 fmol sequencing library onto the flowcell. Dilute library in EB if required.

**Note**

The original ONT protocol says to load 20 ng for 400 bp amplicons, but this leads to only ~50% pore occupancy. Loading 40 ng leads to ~70% pore occupancy but the flow cell needs to be refueled after 24 hrs. This guideline can be used for both MinION and PromethION fc, however, it needs to be adjusted for Flongle fc.

## 16.1

Thaw the following reagents at room temperature before placing on ice:


Sequencing buffer (SQB)

Loading beads (LB)

Flush buffer (FLB)

Flush tether (FLT)

## 16.2

Add  30 µL FLT to the FLB tube (1.16 mL) and mix well by vortexing.

## 16.3


If required place a new MinION flowcell onto the MinION by flipping open the lip and pushing one end of the flowcell under the clip and pushing down gently.

## 16.4

Rotate the inlet port cover clockwise by 90° so that the priming port is visible.




16.5

Take a P1000 pipette and tip and set the volume to  800  $\mu\text{L}$  . Place the tip in the inlet port and holding perpendicularly to the plane of the flowcell remove any air from the inlet port by turning the volume dial anti-clockwise.

**Note**

Be careful not to remove so much volume that air is introduced onto the rectangular array via the outlet.

16.6

Load  800  $\mu\text{L}$  of FLB (plus FLT) into the flow cell via the inlet port, dispense slowly and smoothly trying to avoid the introduction of any air bubbles.


16.7

Wait for  00:05:00

16.8

Gently lift the SpotON cover to open the SpotON port.

16.9


Load another  200  $\mu\text{L}$  of FLB (plus FLT) into the flow cell via the inlet port, this will initiate a siphon at the SpotON port to allow you to load the library dilution.

16.10

In a new tube prepare the library dilution for sequencing:

**Component****MinION fc****PromethION fc**

SQB

 37.5  $\mu\text{L}$  75  $\mu\text{L}$ 

LB

 25.5  $\mu\text{L}$  51  $\mu\text{L}$ 

Final library (adjust with EB)

 12  $\mu\text{L}$  24  $\mu\text{L}$ **Total** 75  $\mu\text{L}$  150  $\mu\text{L}$ 

16.11

Mix the prepared library gently by pipetting up and down just prior to loading.

16.12



- 16.13 Add the sequencing library to the flow cell via the SpotON sample port in a dropwise fashion. Ensure each drop siphons into the port before adding the next.  
Gently replace the SpotON sample port cover, making sure the bung enters the SpotON port, close the inlet port and close the lid.
- 17 Start the sequencing run using MinKNOW.
- 17.1 Monitor the progress of the run using the MinKNOW interface.