VERSION 2

AUG 23, 2018

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.suceesw

Protocol Citation: Mitchell Finger, Michael Lyon, Judy A Northill, Ian M Mackay 2018. Measles Vaccine Virus Taqman-MGB. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.suceesw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it is working

Created: Aug 23, 2018

Last Modified: Aug 23, 2018

PROTOCOL integer ID:
14948

Measles Vaccine Virus Taqman-MGB V.2

Mitchell Finger¹, Michael Lyon¹, Judy A Northill¹,
Ian M Mackay¹

¹Public Health Virology, Forensic and Scientific Services

Public Health Virology, Forensic and Scientific Services

Ian M Mackay

Public Health Virology, Forensic and Scientific Services

ABSTRACT

This previously unpublished protocol aims to amplify genotype A measles virus (MeV) strains but not non-measles viruses.

Mitchell Finger and Michael Lyon developed this in-house test in 2010.

The assay targets the intergenic region between the M (matrix) and F (fusion) genes, designed as a qualitative test for investigating measles vaccine virus (MVV) strains.

Numbering indicates the oligonucleotide location on the sequence with MeV strain Edmonston (Moraten vaccine), complete genome, GenBank accession number AF266287.

MATERIALS

STEP MATERIALS

SuperScript™ III Platinum™ One-Step qRT-PCR Kit **Life Technologies Catalog**
#11732088

SuperScript™ III Platinum™ One-Step qRT-PCR Kit **Life Technologies Catalog**
#11732088

BEFORE START INSTRUCTIONS

If using a different brand or model of real-time thermocycler, check the concentration of ROX is adequate.

Method assumes the user is familiar with the thermocycler and software used to run the protocol and with PCR in general.

Oligonucleotide sequences

1	Name	Sequence 5'-3'
	Measles F 4729 Vac	AAACCCCCAGCAATTGGAA
	Measles R 4795 Vac	GGTCACCTCGGTCGCTTGT
	Measles Probe 4757	FAM - CCCTCTTCCTCAACACA - MGBNFQ

Reagents

2 SuperScript™ III Platinum™ One-Step qRT-PCR Kit Life Technologies Catalog #11732088

Reaction set-up

3 The assay has been used on both a Rotor-Gene 6000 and a Rotor-Gene Q real-time thermocycler

Prepare sufficient mix for the number of reactions.

Include a suitable 'dead volume' as necessary if using a robotic dispenser.

Reagent	Volume (µl) x1	Final reaction concentration
Nuclease-free water	4.42	N/A
Measles F 4729 Vac 150pmol/µl	0.04	300nM
Measles R 4795 Vac 150pmol/µl	0.04	300nM
Measles Probe FAM 100pmol/µl	0.06	300nM
2X Reaction Mix ¹	10	1X
SuperScript® III/Platinum® <i>Taq</i> Mix ¹	0.4	1X
ROX Reference Dye (25µM)	0.04	0.05µM
Template	5	N/A
TOTAL	20	

1

Superscript

TM

III Platinum

TM

One-step qRT-PCR kit

Dispense 15µL to each reaction well.
Add 5µL of template (extracted RNA, controls or NTC [nuclease-free water]).
Total reaction volume is 20µL

Amplification

4

	50°C	5min	1X
	95°C	2min	1X
	95°C	3sec	40X
	60°C	30sec ¹	

1

Fluorescence acquisition step

Result Analysis

5

The definition used for a satisfactory positive result from a real-time fluorogenic PCR should include each of the following:

A **sigmoidal curve** – the trace travels horizontally, curves upward, continues in an exponential rise and followed by a curve towards a horizontal plateau phase

A **suitable level of fluorescence** intensity as measured in comparison to a positive control (y-

axis)

A **defined threshold (C_T) value** which the fluorescent curve has clearly exceeded (Fig.1 arrow), which sits early in the log-linear phase and is <40 cycles

A flat or non-sigmoidal curve or a curve that crosses the threshold with a C_T >40 cycles is considered a negative result.

NTCs should not produce a curve

Figure 1. Examples of satisfactory sigmoidal amplification curve shape when considering an assay's fluorescent signal output. The crossing point or threshold cycle (C_T) is indicated (yellow arrow); it is the value at which fluorescence levels surpass a predefined (usually set during validation, or arbitrary) threshold level as shown in this normalized linear scale depiction. LP-log-linear phase of signal generated during the exponential part of the PCR amplification; TP-a slowing of the amplification and accompanying fluorescence signal marks the transition phase; PP-the plateau phase is reached when there is little or no increase in fluorescent signal despite continued cycling.