VERSION 2

SEP 30, 2019

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.7tfhnjn

External link:

<https://www.diacomp.org/shared/document.aspx?id=44&docType=Protocol>

Protocol Citation: E. Dale Abel 2019. Measurement of Left Ventricular Performance in Langendorff Perfused Mouse Hearts. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.7tfhnjn>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Sep 30, 2019

Last Modified: Sep 30, 2019

Measurement of Left Ventricular Performance in Langendorff Perfused Mouse Hearts V.2

E. Dale Abel¹

¹University of Iowa

Diabetic Complications Consortium

Tech. support email: rmcindoe@augusta.edu

Lili Liang

ABSTRACT

Summary:

This protocol describes the procedure used by the DiaComp for measuring left ventricular performance in isolated retrogradely perfused mouse hearts.

Diabetic Complication:

Cardiovascular

Keywords: Left Ventricular
Performance, cardiovascular

Langendorff Heart Perfusion Protocol

- 1 Hearts are isolated and the aorta is cannulated using a 20g steel cannula. Hearts are perfused at a constant pressure of 60 mmHg by an aortic cannula delivering warm (37°C) Krebs buffer containing (in mM) 118 NaCl, 4.7 KCl, 25 NaHCO₃, 1.2 MgSO₄, 1.2 KH₂PO₄, 2 CaCl₂ gassed with 95% O₂, 5% CO₂. Hearts are perfused with glucose 11mM as sole substrate or in combination with 1 or 1.2 mM palmitate. The pulmonary artery is transected to facilitate coronary venous drainage. A left ventricular polyethylene apical drain is inserted through a left atrial incision to allow thebesian venous drainage. Left ventricular pressure is monitored from a water-filled balloon placed through the left atrial appendage and connected to a Millar transducer. The volume of the balloon is adjusted to obtain a left ventricular diastolic pressure of 7 mmHg. Heart rates are adjusted to 360 beats/min by pacing at 6 Hz at the level of the atria.

Inotropic stress protocol

- 2 After 30 min stabilization, data are acquired under baseline conditions (buffer calcium concentration =2 mM). The hearts are then switched to a second buffer containing 4 mM CaCl₂. Contractile parameters are again measured after 20 min of stabilization. The langendorff protocols yield the following parameters:
 - (1) Left ventricle systolic pressure (LVSP): Units mmHg.
 - (2) Left ventricle developed pressure (LVDP or LVDevP), which is LVSP – LV Diastolic pressure: Units mmHg.
 - (3) Heart Rate (HR): Units beats per minute
 - (4) Rate Pressure Product (RPP), which is LVDP x HR. Units mmHg/sec
 - (5) dP/dtmin and dP/dtmax, which are the maximal rates of LV pressure decay and LV, pressure development respectively. Units mmHg/msec.
 - (6) Coronary Flow (ml/min). This is determined by measuring the coronary effluent from the perfused heart.

Myocardial oxygen consumption (MVO₂).

- 3 Coronary effluent is sampled from the pulmonary artery using a capillary tube. Oxygen content

was measured using an Oxygen foxy probe (OceanOptics) and calculated using the formula:

$$MVO_2 = \% O_2_{\text{perfusate}} - \% O_2_{\text{pulmonary artery}} \times \text{Coronary Flow} \times \text{Atmospheric Pressure} / 760 \times O_2 \text{ Solubility} \times O_2 \text{ Density}$$

Where O_2 solubility and O_2 density are 23.9 $\mu\text{l/ml}$ and 0.03933 $\mu\text{mol}/\mu\text{l}$ respectively in a solution at 37°C respectively.

Cardiac efficiency is calculated as the ratio of RPP/ MVO_2 expressed as a percentage.