

Sep 14, 2019 Version 1

Microplate reader operating procedure V.1

DOI

dx.doi.org/10.17504/protocols.io.7cnhive

宏亮 董¹

¹Northeast Forest University

2019 iGEM NEFU_China

Tech. support email: shengyiyanwork@gmail.com



宏亮 董

OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.7cnhive

Protocol Citation: 宏亮 董 2019. Microplate reader operating procedure. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.7cnhive>

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development

We are still developing and optimizing this protocol

Created: September 14, 2019

Last Modified: September 14, 2019

Protocol Integer ID: 27758

Before start

1. Click on "instrument" on the left, click on "temperature", and check "temperature control" to control the real-time temperature.
2. In the "Action" bar of the left column, you can drag it to the process list under the 96-well plate, drag up and down to sort, click the trash can to delete the option. You can design your own processes that meet your needs.



10m

- 1 Turn on the power and warm up for 10 minutes.
- 2 Open the program and select "New" to create a new process.
- 3 Select "Absorbance" in the left column. There is a schematic of a 96-well plate in the middle of the screen.
- 4 You can select the hole to be used according to the spotting hole, and the selected hole will turn blue.
- 5 Put in the 96-well plate, set the parameters, click "start" to measure the absorbance, and then remove the 96-well plate.
- 6 Wait a moment, it will generate an Excel file with results and various parameters. Save and copy the file to the u disk and process the data.