VERSION 2

SEP 18, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.7fkhjkw

Protocol Citation: 宏亮 董
 2019. Microplate reader
 operating procedure.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.7fkhjkw>

License: This is an open
 access protocol distributed
 under the terms of
 the [Creative Commons
 Attribution License](#), which
 permits unrestricted use,
 distribution, and reproduction
 in any medium, provided the
 original author and source are
 credited

Protocol status: Working
 We use this protocol and it's
 working

Created: Sep 18, 2019

Last Modified: Sep 18, 2019

PROTOCOL integer ID:
 27852

Keywords: Microplate reader,
 Absorbance, Fluorescence

🌐 Microplate reader operating procedure V.2

宏亮 董¹

¹Northeast Forest University

2019 iGEM NEFU_China

Tech. support email: shengyiyanwork@gmail.com

宏亮 董

ABSTRACT

We use Tecan Spark[®] multimode microplate reader to measure absorbance.

MATERIALS

Tecan Spark[®] multimode microplate reader
 96-well plate

SAFETY WARNINGS

Please read the instructions carefully before use.
 Please turn off the power after use.

BEFORE START INSTRUCTIONS

1. Click on "instrument" on the left, click on "temperature", and check "temperature control" to control the real-time temperature.
2. In the "Action" bar of the left column, you can drag it to the process list under the 96-well plate, drag up and down to sort, click the trash can to delete the option. You can design your own processes that meet your needs.

- 1 Turn on the power and warm up for 10 minutes.
- 2 Open the program and select "New" to create a new process.
- 3 Select "Absorbance" in the left column. There is a schematic of a 96-well plate in the middle of the screen.
- 4 You can select the hole to be used according to the spotting hole, and the selected hole will turn blue.
- 5 Put in the 96-well plate, set the parameters, click "start" to measure the absorbance, and then remove the 96-well plate.
- 6 Wait a moment, it will generate an Excel file with results and various parameters. Save and copy the file to the u disk and process the data.