

Oct 16, 2019 Version 1

Microplate Reader Workflow V.1

DOI

dx.doi.org/10.17504/protocols.io.8bnhsme

NUS iGEM¹

¹National University of Singapore



OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.8bnhsme

Protocol Citation: NUS iGEM 2019. Microplate Reader Workflow. [protocols.io](#)

<https://dx.doi.org/10.17504/protocols.io.8bnhsme>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: October 16, 2019

Last Modified: September 11, 2023

Protocol Integer ID: 28750

- 1 Refresh overnight cell culture in LB media at  37 °C

Protocol



NAME

Preparation of LB Media

CREATED BY
NUS iGEM[PREVIEW](#)

- 1.1 Weigh  25 g of Luria Broth Base powder.

 Luria Broth Base (Miller's LB Broth Base)™, powder Thermo Fisher Catalog #12795027

- 1.2 Add the powder into  1 L of water.

 Water refers to sterilized deionized water

- 1.3 Autoclave entire bottle of LB media.

- 2 Measure OD₆₀₀ of cell culture for desired OD value

- 3 Load  200 µL of cell samples into three wells of a 96-well plate (triplicates)

- 4 Induce appropriate volumes of chemical inducers (not exceeding  6 µL) in each well

- 5 Include LB media as blanks

6 Load the plate (with lid) into microplate reader

| Equipment | |
|---|-------|
| Cytation 5 | NAME |
| Automated microscope; Multi-mode plate reader | TYPE |
| BioTek | BRAND |
| N.A. | SKU |
| https://www.bioteck.com/products/cytation.html | LINK |
|   | |

7 Start protocol to run continuous measurement overnight