VERSION 2

SEP 07, 2018

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.tb2eiqe

Protocol Citation: Asar AK Khan, Sumaira Shams, Saima Khan, Muhammad Iftikhar Khan, Sardar Khan, Abid Ali 2018. Modified ZN Staining Protocol . **protocols.io** <https://dx.doi.org/10.17504/protocols.io.tb2eiqe>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Sep 07, 2018

Last Modified: Sep 07, 2018

PROTOCOL integer ID:
 15450

Modified ZN Staining Protocol V.2

Version 1 is forked from [mZN Staining Protocol](#)

Asar AK Khan¹, Sumaira Shams¹, Saima Khan¹, Muhammad Iftikhar Khan², Sardar Khan³, Abid Ali¹

¹Department of Zoology Abdul Wali Khan University Mardan, KPK Pakistan;

²Centre for Biotechnology and Microbiology University of Swat, Pakistan;

³Department of Environmental Science, University of Peshawar

Asar AK Khan

Department of Zoology Abdul Wali Khan University Mardan 2320...

ABSTRACT

The Modified Ziehl-Neelsen stain (mZN stain) is a type of differential bacteriological stain used to identify acid-fast organisms, mainly *Mycobacteria*. Acid fast organisms are those which are capable of retaining the primary stain when treated with an acid (*fast=holding capacity*). Members of the Actinomycetes, genus *Nocardia* (*N. brasiliensis* and *N. asteroides* are opportunistic pathogens) are partially acid-fast. Oocysts of coccidian parasites, such as *Cryptosporidium* and *Isospora*, are also acid-fast. Hence they can also be detected and identified through mZN staining procedure.

MATERIALS

MATERIALS

- ☒ Carbol-Fuchsin **Contributed by users**
- ☒ Distilled Water **Contributed by users**
- ☒ Methanol **Sigma Aldrich Catalog #M3641**
- ☒ Disposable Latex Gloves, Medium, 100/Box **Bio Basic Inc. Catalog #GL002M.SIZE.1PK**
- ☒ Methylene Blue **Gold Biotechnology Catalog #M-680**
- ☒ Microscope slides **Contributed by users**
- ☒ Compound Microscope **Contributed by users**
- ☒ ethanol **BBI Biotech**
- ☒ Acid Alcohol **Contributed by users**

STEP MATERIALS

- ☒ Carbol-Fuchsin **Contributed by users**
- ☒ Acid Alcohol **Contributed by users**
- ☒ Methylene Blue **Gold Biotechnology Catalog #M-680**

- 1 The stool sample was Spread evenly on the middle of the slide with constant rotational movement.

⌚ 00:10:00 (5 to 10 minutes) for rotational movement

⚖ 3 mg (Amount of stool sample)

- 2 The slides were then placed on dryer with smeared surface upwards to air-dried them.

🔥 60 °C

⌚ 00:10:00 minutes

- 3 The dried smear was fixed with absolute methanol.

⌚ 00:05:00 or (3-5 minutes)

- 4 Now, the Carbol-fuchsin solution was added to the slide to cover the whole smear.

☒ Carbol-Fuchsin **Contributed by users**

🕒 00:20:00 minutes

- 5 The slides were washed gently with tap water with the help of a dropper.

Safety information

Do not expose the slides to the high pressure of tap water directly, rather it will be better to use a dropper for washing the slides.

- 6 After washing the slide, decolorizer (Acid Alcohol) was added to the smear and the slide washed again with tap water.

☒ Acid Alcohol **Contributed by users**

🧴 3 mL or 4-6 drops

- 7 Then the counter stain (Methylene Blue) was added and left for 5 minutes and then washed the slide with clean water.

☒ Methylene Blue **Gold Biotechnology Catalog #M-680**

🕒 00:05:00 minutes wait for methylene blue

- 8 The back side of the slides were cleaned with a tissue paper and put in the draining rack to air-dry.

🕒 00:05:00 minutes, wait for slide to dry

- 9 The smear was examined with the help of a compound microscope with 40x and 100x (immersion oil lens) objective and scanned thoroughly for

parasite identification.

Equipment	
	NAME
Olympus	BRAND
CH20i	SKU
Biological microscope , Anti-fungus treated optics , Built to last- Superior build quality	SPECIFICATIONS