MojoSort™ Isolation Kits Protocol - 1 V.2

Sam Li¹

¹BioLegend

Sam Li
BioLegend

ABSTRACT

DOI:

[dx.doi.org/10.17504/protocols.io.7x6hpre](https://doi.org/10.17504/protocols.io.7x6hpre)

External link:

<https://www.biolegend.com/protocols/mojosort-isolation-kits-protocol-1/4599/>

Protocol Citation: Sam Li . MojoSort™ Isolation Kits Protocol - 1. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.7x6hpre>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Created: Oct 03, 2019

Last Modified: Oct 03, 2019

PROTOCOL integer ID: 28382

Keywords: mojosort, isolation, magnet, selection

Product description and procedure summary:

Target cells are depleted by incubating the sample with the biotin antibody cocktail followed by incubation with magnetic Streptavidin Nanobeads (Cat. No. [480015/480016](#)). The magnetically labeled fraction is retained by the use of a magnetic separator. The untouched cells are collected. These are the cells of interest; do not discard the liquid. Some of the downstream applications include functional assays, gene expression, phenotypic characterization, etc.

Note: This protocol has been optimized to remove washing steps after antibody cocktail and nanobeads incubations, resulting in a shorter and more convenient protocol. This procedure is optimized for the isolation of 10^7 to 2×10^8 cells per tube. If working with fewer than 10^7 cells, keep volumes as indicated for 10^7 cells. For best results, optimize the conditions to your specific cell number and tissue. Prepare fresh MojoSort™ Buffer solution by diluting the 5X concentrate with sterile distilled water.

Scale up volumes if using 14 mL tubes and Magnet, and place the tube in the magnet for 10 minutes.

GUIDELINES

Important Note

MojoSort™ magnetic particles can be used with other commercially available magnetic separators, both free standing magnets and column-based systems. Because MojoSort™ protocols are optimized for the MojoSort™ separator, the protocols may need to be adjusted for other systems. Please contact BioLegend Technical Service (tech@biolegend.com) for more information and guidance. We do not recommend using MojoSort™ particles for BD's IMag™ or Life Technologies' DynaMag™.

Application notes: To use this product in magnetic separation columns, a titration of the Nanobeads should be performed. Optimal concentration for magnetic separation columns is lot-specific. Please contact BioLegend Technical Service (tech@biolegend.com) for further assistance on how to use MojoSort™ Nanobeads in magnetic separation columns.

MATERIALS

MATERIALS

✕ MojoSort™ Buffer **BioLegend Catalog #480017**

✕ MojoSort™ Magnet **BioLegend Catalog #480019**

✕ MojoSort™ Mouse CD3 T Cell Isolation Kit **BioLegend Catalog #480023, 480024, 480031**

✕ MojoSort™ Mouse CD4 T Cell Isolation Kit **BioLegend Catalog #480005, 480006, 480033**

✕ MojoSort™ Mouse CD8 T Cell Isolation Kit **BioLegend Catalog #480007, 480008, 480035**

✕ MojoSort™ Human CD4 T Cell Isolation Kit **BioLegend Catalog #480009, 480010**

✕ MojoSort™ Human CD3 T Cell Isolation Kit **BioLegend Catalog #480021, 480022**

✕ MojoSort™ Mouse CD4 Naïve T Cell Isolation Kit **BioLegend Catalog #480039, 480040**

✕ MojoSort™ Human CD4 Naïve T Cell Isolation Kit **BioLegend Catalog #480041, 480042**

✕ MojoSort™ Mouse CD8 Naïve T Cell Isolation Kit **BioLegend Catalog #480043, 480044**

✕ MojoSort™ Human CD8 Naïve T Cell Isolation Kit **BioLegend Catalog #480045, 480046**

✕ MojoSort™ Mouse Pan B Cell Isolation Kit **BioLegend Catalog #480051, 480052**

✕ MojoSort™ Human B Cell (CD43-) Isolation Kit **BioLegend Catalog #480061, 480062**

✕ MojoSort™ Human CD4 Memory T Cell Isolation Kit **BioLegend Catalog #480063, 480064**

✕ MojoSort™ Human Naïve B Cell Isolation Kit **BioLegend Catalog #480067, 480068**

✕ MojoSort™ Human Pan B Cell Isolation Kit **BioLegend Catalog #480081, 480082**

✕ MojoSort™ Mouse Pan B Cell Isolation Kit II **BioLegend Catalog #480087, 480088**

- Adjustable pipettes
- 70µm filters (one per sample)
- 5mL (12 x 75mm) or 14mL (17 x 100mm) polypropylene tubes
- Reagents for sample preparation
- Reagents and instruments (Flow cytometer) to determine yield and purity

- 1 Prepare cells from your tissue of interest or blood without lysing erythrocytes. Kits for human samples have been optimized for PBMCs, please prepare the cells using a suitable method.

- 2 In the final wash of your sample preparation, resuspend the cells in MojoSort™ Buffer by adding up to 4mL in a 5 mL (12 x 75 mm) polypropylene tube.
Note: Keep MojoSort™ Buffer on ice throughout the procedure.
- 3 Filter the cells with a 70µm cell strainer, centrifuge at 300xg for 5 minutes, and resuspend in an appropriate volume of MojoSort™ Buffer. Count and adjust the cell concentration to 1×10^8 cells/mL.
- 4 Aliquot 100µL of cell suspension (10^7 cells) into a new tube. **Add 10µL of the Biotin-Antibody Cocktail.** Mix well and **incubate on ice for 15 minutes.** Scale up the volume accordingly if separating more cells. For example, add 100µL of biotinylated antibody cocktail for separating 1×10^8 cells in 1 ml of MojoSort™ Buffer. When working with less than 10^7 cells, use indicated volumes for 10^7 cells.
Optional: Take an aliquot before adding the cocktail to monitor purity and yield.
- 5 Resuspend the beads by vortexing, maximum speed, 5 touches. **Add 10µL of Streptavidin Nanobeads.** Mix well and **incubate on ice for 15 minutes.** Scale up the volume accordingly if separating more cells. For example, add 100µL of Nanobeads for separating 1×10^8 cells in 1 ml of MojoSort™ Buffer. When working with less than 10^7 cells, use indicated volumes for 10^7 cells.
- 6 Add 2.5mL of MojoSort™ Buffer.
Note: If you observe aggregates, filter the suspension. To maximize yield, you can disrupt the aggregates by pipetting the solution up and down.
- 7 Place the tube in the magnet for 5 minutes.
Optional: Take a small aliquot before placing the tube in the magnet to monitor purity and yield. Keep unused cells to be used as control or other applications if needed.
- 8 Pour out and collect the liquid. These are your cells of interest; **DO NOT DISCARD.**
- 9 Repeat steps 6-8 with labeled cells once more for a total of **2 separations.** Pool the unlabeled fractions. The labeled cells may be useful as staining controls, to monitor purity/yield, or other purposes.
Note: Repeating the magnetic separation increases the yield, without a strong impact on the purity. The yield will typically increase about 8-10% with a second separation. The purity may decrease 1-2% with each separation.