

Apr 08, 2025

Version 1

Molecular Cartography/Resolve OCT Sectioning V.1

DOI

<https://dx.doi.org/10.17504/protocols.io.bp2l68bxrgqe/v1>

Joanna Bi¹

¹Stanford University

Human BioMolecular Atlas Program (HuBMAP) Method Development Community
Tech. support email: Jeff.spraggins@vanderbilt.edu



Joanna Bi

Stanford University

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.bp2l68bxrgqe/v1>

Protocol Citation: Joanna Bi 2025. Molecular Cartography/Resolve OCT Sectioning. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bp2l68bxrgqe/v1>

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited



Protocol status: Working

We use this protocol and it's working

Created: April 08, 2025

Last Modified: April 08, 2025

Protocol Integer ID: 126300

Keywords: oct samples for molecular cartography, sectioning oct sample, molecular cartography, oct sectioning protocol, resolve, resolve oct sectioning protocol

Abstract

Protocol for sectioning OCT samples for Molecular Cartography/Resolve.

- 1 Set cryostat to -20° C. Place samples inside to acclimate to temperatures for 10 minutes.
- 2 Add OCT media to the chuck and place the sample block on top. Let freeze for 5 minutes.
- 3 Insert and fasten cryostat blade and set sectioning thickness to 30-60 um.
- 4 Place chuck in specimen station and adjust sample distance and angle to blade.
- 5 Trim sample block until desired depth is reached.
- 6 Lightly score block with a razor blade so that tissue fits within the designated scan areas on the 8-sample Resolve slides.
- 7 Section at 5 um and place section within the a scan area of the slide.
- 8 Repeat as needed to fill the slide.
- 9 Store slides at -80°C directly after sectioning.