May 12, 2023 Version 1

NCEM Drop conventional (TM-014) V.1

DOI
dx.doi.org/10.17504/protocols.io.36wgqjx95vk5/v1

sandra.crameri[1]
[1]CSIRO ACDP

sandra.crameri

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN ACCESS

DOI: https://dx.doi.org/10.17504/protocols.io.36wgqjx95vk5/v1

Protocol Citation: sandra.crameri 2023. NCEM Drop conventional (TM-014). **protocols.io** https://dx.doi.org/10.17504/protocols.io.36wgqjx95vk5/v1

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: May 12, 2023

Last Modified: May 12, 2023

Protocol Integer ID: 81818

Keywords: standard routine conventional negative contrast em, ncem drop, ncem, conventional

Abstract

Standard routine conventional negative contrast EM.

Protocol materials

nano-W **Catalog** ##2018-5ML

Troubleshooting

HEADER

1 SAN:

SPEC No:

OPERATOR:

Conventional

11m

2 Adsorb 10 µL sample to grid 00:10:00 , visually inspect to ensure sample does not dry out in this time. 10m

3 Drain excess sample from grid using filter paper, leave wet.

4 Stain nano-W **Catalog #**#2018-5ML 00:01:00 1m

5 Drain & dry using filter paper