

Jun 17, 2023

Version 2

NEBNext Ultra II Ligation Module (NEB # E7595) for NEBNext Ultra II FS DNA Module (NEB # E7810) V.2

DOI

dx.doi.org/10.17504/protocols.io.n92ld3xw9g5b/v2



New England Biolabs¹, Juliet Bonnevie¹

¹New England Biolabs

New England Biolabs (NEB)

Tech. support phone: +1(800)632-7799 email: info@neb.com



Isabel Gautreau

New England Biolabs

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.n92ld3xw9g5b/v2>

External link: <https://www.neb.com/products/e7595-nebnext-ultra-ii-ligation-module#Protocols,%20Manuals%20&%20Usage>

Protocol Citation: New England Biolabs, Juliet Bonnevie 2023. NEBNext Ultra II Ligation Module (NEB # E7595) for NEBNext Ultra II FS DNA Module (NEB # E7810). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.n92ld3xw9g5b/v2> Version created by **Juliet Bonnevie**

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: June 09, 2023

Last Modified: June 22, 2023

Protocol Integer ID: 83069

Keywords: ultra ii ligation module, ultra ii fs dna module, neb, ultra ii end repair, ligation, e7810, e7546, dna, module

Abstract

This module is part of the Ultra™ II workflow, and is optimized for use with the NEBNext® Ultra II End Repair/dA-Tailing Module (NEB #**E7546**), for Illumina®-compatible library construction.

The NEBNext Ultra II Ligation Module is optimized for use with the NEBNext Ultra II End Repair/dA-Tailing Module (NEB #**E7546**) or the NEBNext Ultra II FS DNA Module (NEB #**E7810**).

Guidelines

Safe Stop Point: This is a point where you can safely stop the protocol and store the samples prior to proceeding to the next step in the protocol.

Caution: Signifies a step in the protocol that has two paths leading to the same point.

Color: A color listed before or after a reagent name indicates the cap color of the reagent to be added.

Adaptor Dilution Guidelines

The appropriate adaptor dilution for your sample input and type may need to be optimized experimentally. The dilutions provided here are a general starting point.

Table 2.1: Adaptor Dilution

| | A | B | C |
|--|----------------|---|--|
| | Input | Adaptor Dilution (Volume of adaptor: Total volume) | Working Adaptor Concentration |
| | 100 ng–500 ng | No Dilution | 15 μ M |
| | 5 ng–99 ng | 10-Fold (1:10) | 1.5 μ M |
| | less than 5 ng | 25-Fold (1:25) | 0.6 μ M |



Materials

MATERIALS

✕ NEBNext Adaptor for Illumina **New England Biolabs Catalog #E7337** in Kits E7335, E7500, E771

✕ NEBNext Ligation Enhancer **New England Biolabs Catalog #E7374**

✕ NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

✕ USER Enzyme (Multiplex Oligos for Illumina) **New England Biolabs Catalog #E7338**

STEP MATERIALS

✕ USER Enzyme (Multiplex Oligos for Illumina) **New England Biolabs Catalog #E7338**

✕ NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

✕ Ligation Enhancer **New England Biolabs Catalog #E7374** in Kits E7370 or E7445

✕ NEBNext Adaptor for Illumina **New England Biolabs**

Materials that you may need that are not provided with this kit include:

Tris-HCL Buffer (pH 7.5)

10 mM NaCl

Protocol materials

✕ NEBNext Ligation Enhancer **New England Biolabs Catalog #E7374**

✕ NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

✕ USER Enzyme (Multiplex Oligos for Illumina) **New England Biolabs Catalog #E7338**

✕ USER Enzyme (Multiplex Oligos for Illumina) **New England Biolabs Catalog #E7338**

✕ NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

✕ Ligation Enhancer **New England Biolabs Catalog #E7374** in Kits E7370 or E7445

✕ NEBNext Adaptor for Illumina **New England Biolabs**

✕ NEBNext Adaptor for Illumina **New England Biolabs Catalog #E7337** in Kits E7335, E7500, E771

✕ NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

✕ Ligation Enhancer **New England Biolabs Catalog #E7374** in Kits E7370 or E7445

✕ NEBNext Adaptor for Illumina **New England Biolabs**

✕ USER Enzyme (Multiplex Oligos for Illumina) **New England Biolabs Catalog #E7338**

Troubleshooting



Before start

Starting Material: 100 pg–500 ng fragmented, end repaired and dA-Tailed DNA generated using the NEBNext Ultra II FS DNA Module (**NEB #E7810**)

Note

Caution: If DNA input is < 100 ng, dilute the (red) NEBNext Adaptor for Illumina in 10 mM Tris-HCl, pH 7.5-8.0 with 10 mM NaCl as indicated in Table 2.1.



DNA Ligation/End Prep


30m

- 1 Add the following components directly to the FS Reaction Mixture:


| A | B |
|---|----------------|
| Component | Volume |
| FS Reaction Mixture | 35 µl |
| (red) NEBNext Adaptor for Illumina** | 2.5 µl |
| (red) NEBNext Ultra II Ligation Master Mix* | 30 µl |
| (red) NEBNext Ligation Enhancer | 1 µl |
| Total volume | 68.5 µl |

* Mix the Ultra II Ligation Master Mix by pipetting up and down several times prior to adding to the reaction.

** The NEBNext adaptor is provided in NEBNext Oligo kit options, which can be found at www.neb.com/oligos

 NEBNext Ultra II Ligation Master Mix **New England Biolabs Catalog #E7648**

 Ligation Enhancer **New England Biolabs Catalog #E7374** in Kits E7370 or E7445

 NEBNext Adaptor for Illumina **New England Biolabs**

Note

Note: The Ligation Master Mix and Ligation Enhancer can be mixed ahead of time and is stable for at least 8 hours @ 4°C. We do not recommend premixing the Ligation Master Mix, Ligation Enhancer and adaptor prior to use in the Adaptor Ligation Step.

- 2 Set a 100 µl or 200 µl pipette to 50 µl and then pipette the entire volume up and down at least 10 times to mix thoroughly. Perform a quick spin to collect all liquid from the sides of the tube.

**Note**

Caution: The NEBNext Ultra II Ligation Master Mix is very viscous. Care should be taken to ensure adequate mixing of the ligation reaction, as incomplete mixing will result in reduced ligation efficiency. The presence of a small amount of bubbles will not interfere with performance.

- 3 Incubate at 20 °C for 00:15:00 in a thermocycler **with the heated lid off**.

15m

- 4 Add 3 µl of (red) USER Enzyme to the ligation mixture.



USER Enzyme (Multiplex Oligos for Illumina) **New England**
Biolabs Catalog #E7338

Note

Note: Steps 4 and 5 are only required for use with NEBNext Adaptors. USER enzyme can be found in most NEBNext oligo kits, all options can be found on the www.neb.com/oligos page. If you are using the indexed UMI adaptor, USER is not needed. Please see corresponding manual for use with UMI on the E7395 product page under protocols, manuals, and usage tab.

- 5 Mix well and incubate at 37 °C for 00:15:00 with the heated lid set to \geq 47 °C .

15m

- 6 DNA is now ready for size selection or cleanup.

Note

Note: Please see NEB #E7805/#E6177 manual for recommended size selection/cleanup and PCR amplification protocols.

Note

Safe Stop Point: Samples can be stored overnight at -20°C.

